



MWH-Pasadena/Boeing Project ID: Routine Outfall 002
300 North Lake Avenue, Suite 1200 Report Number: IOD1705
Pasadena, CA 91101
Attention: Bronwyn Kelly
Sampled: 04/22/05
Received: 04/22/05

ACID & BASE/NEUTRALS BY GC/MS (EPA 625)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD1705-01 (Outfall 002 - Water)									
Reporting Units: ug/l									
Bis(2-ethylhexyl)phthalate	EPA 625	5D25044	1.1	5.0	1.3	0.962	04/25/05	04/27/05	J
2,4-Dinitrotoluene	EPA 625	5D25044	0.23	9.0	ND	0.962	04/25/05	04/27/05	
N-Nitrosodimethylamine	EPA 625	5D25044	0.22	8.0	ND	0.962	04/25/05	04/27/05	
Pentachlorophenol	EPA 625	5D25044	0.78	8.0	ND	0.962	04/25/05	04/27/05	
2,4,6-Trichlorophenol	EPA 625	5D25044	0.10	6.0	ND	0.962	04/25/05	04/27/05	
Surrogate: 2-Fluorophenol (30-120%)					58 %				
Surrogate: Phenol-d6 (35-120%)					60 %				
Surrogate: 2,4,6-Tribromophenol (45-120%)					78 %				
Surrogate: Nitrobenzene-d5 (45-120%)					60 %				
Surrogate: 2-Fluorobiphenyl (45-120%)					54 %				
Surrogate: Terphenyl-d14 (45-120%)					75 %				

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing 300 North Lake Avenue, Suite 1200 Pasadena, CA 91101 Attention: Bronwyn Kelly	Project ID: Routine Outfall 002 Report Number: IOD1705	Sampled: 04/22/05 Received: 04/22/05
--	---	---

ORGANOCHLORINE PESTICIDES (EPA 608)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD1705-01 (Outfall 002 - Water) - cont.									
Reporting Units: ug/l									
alpha-BHC	EPA 608	5D26073	0.0010	0.010	ND	0.98	04/26/05	04/28/05	
<i>Surrogate: Decachlorobiphenyl (45-120%)</i>					77 %				
<i>Surrogate: Tetrachloro-m-xylene (35-115%)</i>					48 %				

Del Mar Analytical, Irvine
Michele Harper
Project Manager



Del Mar Analytical

17461 Derian Ave., Suite 100, Irvine, CA 92614 (949) 261-1022 FAX (949) 260-3297
 1014 E. Cooley Dr., Suite A, Colton, CA 92324 (909) 370-4667 FAX (909) 370-1046
 9484 Chesapeake Dr., Suite 805, San Diego, CA 92123 (858) 505-8596 FAX (858) 505-9689
 9830 South 51st St., Suite B-120, Phoenix, AZ 85044 (480) 785-0043 FAX (480) 785-0851
 2520 E. Sunset Rd. #3, Las Vegas, NV 89120 (702) 798-3620 FAX (702) 798-3621

MWH-Pasadena/Boeing
 300 North Lake Avenue, Suite 1200
 Pasadena, CA 91101
 Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD1705

Sampled: 04/22/05
 Received: 04/22/05

METALS

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD1705-01 (Outfall 002 - Water) - cont.									
Reporting Units: ug/l									
Copper	EPA 200.8	5D22097	0.49	2.0	2.8	1	04/22/05	04/24/05	
Lead	EPA 200.8	5D22097	0.13	1.0	0.33	1	04/22/05	04/24/05	J
Mercury	EPA 245.1	5D25065	0.063	0.20	ND	1	04/25/05	04/25/05	

Del Mar Analytical, Irvine
 Michele Harper
 Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from Del Mar Analytical.



MWH-Pasadena/Boeing
 300 North Lake Avenue, Suite 1200
 Pasadena, CA 91101
 Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD1705

Sampled: 04/22/05
 Received: 04/22/05

INORGANICS

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD1705-01 (Outfall 002 - Water) - cont.									
Reporting Units: mg/l									
Ammonia-N (Distilled)	EPA 350.2	5D26085	0.30	0.50	ND	1	04/26/05	04/26/05	
Biochemical Oxygen Demand	EPA 405.1	5D22071	0.59	2.0	0.86	1	04/22/05	04/27/05	J
Chloride	EPA 300.0	5D22051	2.6	5.0	49	10	04/22/05	04/22/05	
Nitrate/Nitrite-N	EPA 300.0	5D22051	0.072	0.26	ND	1	04/22/05	04/22/05	
Oil & Grease	EPA 413.1	5D25054	0.94	5.0	ND	1	04/25/05	04/25/05	
Sulfate	EPA 300.0	5D22051	1.8	5.0	400	10	04/22/05	04/22/05	
Surfactants (MBAS)	SM5540-C	5D22102	0.044	0.10	0.077	1	04/22/05	04/22/05	J
Total Dissolved Solids	SM2540C	5D25087	10	10	1000	1	04/25/05	04/25/05	
Total Suspended Solids	EPA 160.2	5D27086	10	10	ND	1	04/27/05	04/27/05	
Sample ID: IOD1705-01RE1 (Outfall 002 - Water)									
Reporting Units: mg/l									
Sulfate	EPA 300.0	5D25050	1.8	5.0	420	10	04/25/05	04/26/05	
Total Dissolved Solids	SM2540C	5D26122	10	10	1000	1	04/25/05	04/26/05	
Sample ID: IOD1705-01 (Outfall 002 - Water)									
Reporting Units: ml/l/hr									
Total Settleable Solids	EPA 160.5	5D22101	0.10	0.10	ND	1	04/22/05	04/22/05	
Sample ID: IOD1705-01 (Outfall 002 - Water)									
Reporting Units: NTU									
Turbidity	EPA 180.1	5D23050	0.040	1.0	6.4	1	04/23/05	04/23/05	
Sample ID: IOD1705-01 (Outfall 002 - Water)									
Reporting Units: ug/l									
Total Cyanide	EPA 335.2	5D25090	2.2	5.0	ND	1	04/25/05	04/26/05	
Perchlorate	EPA 314.0	5D27058	0.80	4.0	ND	1	04/27/05	04/28/05	
Sample ID: IOD1705-01 (Outfall 002 - Water)									
Reporting Units: umhos/cm									
Specific Conductance	EPA 120.1	5D25093	1.0	1.0	1400	1	04/25/05	04/25/05	

Del Mar Analytical, Irvine
 Michele Harper
 Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD1705

Sampled: 04/22/05

Received: 04/22/05

SHORT HOLD TIME DETAIL REPORT

Sample ID: Outfall 002 (IOD1705-01) - Water	Hold Time (in days)	Date/Time Sampled	Date/Time Received	Date/Time Extracted	Date/Time Analyzed
EPA 160.5	2	04/22/2005 11:00	04/22/2005 18:00	04/22/2005 20:00	04/22/2005 21:00
EPA 180.1	2	04/22/2005 11:00	04/22/2005 18:00	04/23/2005 08:00	04/23/2005 09:00
EPA 300.0	2	04/22/2005 11:00	04/22/2005 18:00	04/22/2005 19:04	04/22/2005 20:18
EPA 405.1	2	04/22/2005 11:00	04/22/2005 18:00	04/22/2005 20:31	04/27/2005 13:00
SM5540-C	2	04/22/2005 11:00	04/22/2005 18:00	04/22/2005 21:00	04/22/2005 21:43

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002
Report Number: IOD1705

Sampled: 04/22/05
Received: 04/22/05

METHOD BLANK/QC DATA

PURGEABLES BY GC/MS (EPA 624)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC Limits	RPD	RPD Limit	Data Qualifiers
---------	--------	-----------------	-----	-------	-------------	---------------	-------------	-----	-----------	-----------------

Batch: 5E03004 Extracted: 05/03/05

Blank Analyzed: 05/03/2005 (5E03004-BLK1)

Benzene	ND	2.0	0.28	ug/l						
Trichlorotrifluoroethane (Freon 113)	ND	5.0	1.2	ug/l						
Carbon tetrachloride	ND	5.0	0.28	ug/l						
Chloroform	ND	2.0	0.33	ug/l						
1,1-Dichloroethane	ND	2.0	0.27	ug/l						
1,2-Dichloroethane	ND	2.0	0.28	ug/l						
1,1-Dichloroethene	ND	3.0	0.32	ug/l						
Ethylbenzene	ND	2.0	0.25	ug/l						
Tetrachloroethene	ND	2.0	0.32	ug/l						
Toluene	ND	2.0	0.36	ug/l						
1,1,1-Trichloroethane	ND	2.0	0.30	ug/l						
1,1,2-Trichloroethane	ND	2.0	0.30	ug/l						
Trichloroethene	ND	5.0	0.26	ug/l						
Trichlorofluoromethane	ND	5.0	0.34	ug/l						
Vinyl chloride	ND	5.0	0.26	ug/l						
Xylenes, Total	ND	4.0	0.52	ug/l						
Surrogate: Dibromofluoromethane	28.4			ug/l	25.0		114	80-120		
Surrogate: Toluene-d8	27.3			ug/l	25.0		109	80-120		
Surrogate: 4-Bromofluorobenzene	27.5			ug/l	25.0		110	80-120		

LCS Analyzed: 05/03/2005 (5E03004-BS1)

Benzene	25.0	2.0	0.28	ug/l	25.0		100	70-120		
Carbon tetrachloride	31.7	5.0	0.28	ug/l	25.0		127	70-140		
Chloroform	27.3	2.0	0.33	ug/l	25.0		109	75-130		
1,1-Dichloroethane	25.0	2.0	0.27	ug/l	25.0		100	70-135		
1,2-Dichloroethane	31.0	2.0	0.28	ug/l	25.0		124	60-150		
1,1-Dichloroethene	23.7	3.0	0.32	ug/l	25.0		95	75-135		
Ethylbenzene	25.7	2.0	0.25	ug/l	25.0		103	80-120		
Tetrachloroethene	26.9	2.0	0.32	ug/l	25.0		108	75-125		
Toluene	24.0	2.0	0.36	ug/l	25.0		96	75-120		
1,1,1-Trichloroethane	23.2	2.0	0.30	ug/l	25.0		93	75-140		
1,1,2-Trichloroethane	26.6	2.0	0.30	ug/l	25.0		106	70-125		
Trichloroethene	25.1	5.0	0.26	ug/l	25.0		100	80-120		
Trichlorofluoromethane	32.2	5.0	0.34	ug/l	25.0		129	65-145		
Vinyl chloride	24.7	5.0	0.26	ug/l	25.0		99	50-130		
Surrogate: Dibromofluoromethane	27.8			ug/l	25.0		111	80-120		

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD1705

Sampled: 04/22/05
Received: 04/22/05

METHOD BLANK/QC DATA

PURGEABLES BY GC/MS (EPA 624)

Analyte Result Reporting Limit MDL Units Spike Level Source Result %REC Limits RPD Data Qualifiers

Batch: 5E03004 Extracted: 05/03/05

LCS Analyzed: 05/03/2005 (5E03004-BS1)

Surrogate: Toluene-d8 27.9 ug/l 25.0 112 80-120
Surrogate: 4-Bromofluorobenzene 28.4 ug/l 25.0 114 80-120

Matrix Spike Analyzed: 05/03/2005 (5E03004-MS1)

Source: IOD1705-01

Table with 12 columns: Analyte, Result, Reporting Limit, MDL, Units, Spike Level, Source Result, %REC Limits, RPD, Data Qualifiers. Rows include Benzene, Carbon tetrachloride, Chloroform, 1,1-Dichloroethane, 1,2-Dichloroethane, 1,1-Dichloroethane, Ethylbenzene, Tetrachloroethene, Toluene, 1,1,1-Trichloroethane, 1,1,2-Trichloroethane, Trichloroethene, Trichlorofluoromethane, Vinyl chloride, Surrogate: Dibromofluoromethane, Surrogate: Toluene-d8, Surrogate: 4-Bromofluorobenzene.

Matrix Spike Dup Analyzed: 05/03/2005 (5E03004-MSD1)

Source: IOD1705-01

Table with 12 columns: Analyte, Result, Reporting Limit, MDL, Units, Spike Level, Source Result, %REC Limits, RPD, Data Qualifiers. Rows include Benzene, Carbon tetrachloride, Chloroform, 1,1-Dichloroethane, 1,2-Dichloroethane, 1,1-Dichloroethane, Ethylbenzene, Tetrachloroethene, Toluene, 1,1,1-Trichloroethane, 1,1,2-Trichloroethane, Trichloroethene, Trichlorofluoromethane.

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD1705

Sampled: 04/22/05

Received: 04/22/05

METHOD BLANK/QC DATA

PURGEABLES BY GC/MS (EPA 624)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5E03004 Extracted: 05/03/05											
Matrix Spike Dup Analyzed: 05/03/2005 (5E03004-MSD1)						Source: IOD1705-01					
Vinyl chloride	28.8	5.0	0.26	ug/l	25.0	ND	115	40-135	0	30	
Surrogate: Dibromofluoromethane	28.6			ug/l	25.0		114	80-120			
Surrogate: Toluene-d8	27.8			ug/l	25.0		111	80-120			
Surrogate: 4-Bromofluorobenzene	28.9			ug/l	25.0		116	80-120			

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD1705

Sampled: 04/22/05
Received: 04/22/05

METHOD BLANK/QC DATA

ACID & BASE/NEUTRALS BY GC/MS (EPA 625)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC %REC Limits	RPD RPD	RPD Limit	Data Qualifiers
---------	--------	-----------------	-----	-------	-------------	---------------	------------------	---------	-----------	-----------------

Batch: 5D25044 Extracted: 04/25/05

Blank Analyzed: 04/27/2005 (5D25044-BLK1)

Bis(2-ethylhexyl)phthalate	ND	5.0	1.1	ug/l						
2,4-Dinitrotoluene	ND	9.0	0.23	ug/l						
N-Nitrosodimethylamine	ND	8.0	0.22	ug/l						
Pentachlorophenol	ND	8.0	0.78	ug/l						
2,4,6-Trichlorophenol	ND	6.0	0.10	ug/l						
Surrogate: 2-Fluorophenol	12.0			ug/l	20.0		60	30-120		
Surrogate: Phenol-d6	12.4			ug/l	20.0		62	35-120		
Surrogate: 2,4,6-Tribromophenol	16.5			ug/l	20.0		82	45-120		
Surrogate: Nitrobenzene-d5	6.40			ug/l	10.0		64	45-120		
Surrogate: 2-Fluorobiphenyl	6.70			ug/l	10.0		67	45-120		
Surrogate: Terphenyl-d14	7.92			ug/l	10.0		79	45-120		

LCS Analyzed: 04/27/2005 (5D25044-BS1)

Bis(2-ethylhexyl)phthalate	8.18	5.0	1.1	ug/l	10.0		82	60-130		
2,4-Dinitrotoluene	6.88	9.0	0.23	ug/l	10.0		69	60-120		J
N-Nitrosodimethylamine	6.98	8.0	0.22	ug/l	10.0		70	40-120		J
Pentachlorophenol	8.32	8.0	0.78	ug/l	10.0		83	50-120		
2,4,6-Trichlorophenol	8.38	6.0	0.10	ug/l	10.0		84	60-120		
Surrogate: 2-Fluorophenol	12.5			ug/l	20.0		62	30-120		
Surrogate: Phenol-d6	13.6			ug/l	20.0		68	35-120		
Surrogate: 2,4,6-Tribromophenol	17.1			ug/l	20.0		86	45-120		
Surrogate: Nitrobenzene-d5	6.96			ug/l	10.0		70	45-120		
Surrogate: 2-Fluorobiphenyl	7.34			ug/l	10.0		73	45-120		
Surrogate: Terphenyl-d14	7.86			ug/l	10.0		79	45-120		

M-NR1

LCS Dup Analyzed: 04/27/2005 (5D25044-BSD1)

Bis(2-ethylhexyl)phthalate	7.64	5.0	1.1	ug/l	10.0		76	60-130	7	20	
2,4-Dinitrotoluene	6.60	9.0	0.23	ug/l	10.0		66	60-120	4	20	J
N-Nitrosodimethylamine	7.06	8.0	0.22	ug/l	10.0		71	40-120	1	20	J
Pentachlorophenol	7.70	8.0	0.78	ug/l	10.0		77	50-120	8	25	J
2,4,6-Trichlorophenol	8.32	6.0	0.10	ug/l	10.0		83	60-120	1	20	
Surrogate: 2-Fluorophenol	12.3			ug/l	20.0		62	30-120			
Surrogate: Phenol-d6	13.3			ug/l	20.0		66	35-120			
Surrogate: 2,4,6-Tribromophenol	16.2			ug/l	20.0		81	45-120			
Surrogate: Nitrobenzene-d5	6.80			ug/l	10.0		68	45-120			
Surrogate: 2-Fluorobiphenyl	7.16			ug/l	10.0		72	45-120			

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD1705

Sampled: 04/22/05
Received: 04/22/05

METHOD BLANK/QC DATA

ACID & BASE/NEUTRALS BY GC/MS (EPA 625)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5D25044 Extracted: 04/25/05											
LCS Dup Analyzed: 04/27/2005 (5D25044-BSD1)											
Surrogate: Terphenyl-d14	7.50			ug/l	10.0		75	45-120			

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD1705

Sampled: 04/22/05
Received: 04/22/05

METHOD BLANK/QC DATA

ORGANOCHLORINE PESTICIDES (EPA 608)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5D26073 Extracted: 04/26/05											
Blank Analyzed: 04/27/2005 (5D26073-BLK1)											
alpha-BHC	ND	0.010	0.0010	ug/l							
Surrogate: Decachlorobiphenyl	0.380			ug/l	0.500		76	45-120			
Surrogate: Tetrachloro-m-xylene	0.250			ug/l	0.500		50	35-115			
LCS Analyzed: 04/27/2005 (5D26073-BS1)											
alpha-BHC	0.380	0.010	0.0010	ug/l	0.500		76	45-115			M-NR1
Surrogate: Decachlorobiphenyl	0.418			ug/l	0.500		84	45-120			
Surrogate: Tetrachloro-m-xylene	0.310			ug/l	0.500		62	35-115			
LCS Dup Analyzed: 04/27/2005 (5D26073-BSD1)											
alpha-BHC	0.399	0.010	0.0010	ug/l	0.500		80	45-115	5	30	
Surrogate: Decachlorobiphenyl	0.418			ug/l	0.500		84	45-120			
Surrogate: Tetrachloro-m-xylene	0.290			ug/l	0.500		58	35-115			

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
 300 North Lake Avenue, Suite 1200
 Pasadena, CA 91101
 Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD1705

Sampled: 04/22/05

Received: 04/22/05

METHOD BLANK/QC DATA

METALS

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5D22097 Extracted: 04/22/05											
Blank Analyzed: 04/24/2005 (5D22097-BLK1)											
Copper	ND	2.0	0.49	ug/l							
Lead	ND	1.0	0.13	ug/l							
LCS Analyzed: 04/24/2005 (5D22097-BS1)											
Copper	80.0	2.0	0.49	ug/l	80.0		100	85-115			
Lead	84.6	1.0	0.13	ug/l	80.0		106	85-115			
Matrix Spike Analyzed: 04/24/2005 (5D22097-MS1) Source: IOD1705-01											
Copper	79.5	2.0	0.49	ug/l	80.0	2.8	96	70-130			
Lead	81.2	1.0	0.13	ug/l	80.0	0.33	101	70-130			
Matrix Spike Dup Analyzed: 04/24/2005 (5D22097-MSD1) Source: IOD1705-01											
Copper	79.8	2.0	0.49	ug/l	80.0	2.8	96	70-130	0	20	
Lead	81.3	1.0	0.13	ug/l	80.0	0.33	101	70-130	0	20	
Batch: 5D25065 Extracted: 04/25/05											
Blank Analyzed: 04/25/2005 (5D25065-BLK1)											
Mercury	ND	0.20	0.063	ug/l							
LCS Analyzed: 04/25/2005 (5D25065-BS1)											
Mercury	7.84	0.20	0.063	ug/l	8.00		98	85-115			
Matrix Spike Analyzed: 04/25/2005 (5D25065-MS1) Source: IOD1505-07											
Mercury	8.20	0.20	0.063	ug/l	8.00	ND	102	70-130			
Matrix Spike Dup Analyzed: 04/25/2005 (5D25065-MSD1) Source: IOD1505-07											
Mercury	8.29	0.20	0.063	ug/l	8.00	ND	104	70-130	1	20	

Del Mar Analytical, Irvine
 Michele Harper
 Project Manager



MWH-Pasadena/Boeing 300 North Lake Avenue, Suite 1200 Pasadena, CA 91101 Attention: Bronwyn Kelly	Project ID: Routine Outfall 002 Report Number: IOD1705	Sampled: 04/22/05 Received: 04/22/05
--	---	---

METHOD BLANK/QC DATA

INORGANICS

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC %REC	RPD Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5D22051 Extracted: 04/22/05											
Blank Analyzed: 04/22/2005 (5D22051-BLK1)											
Chloride	ND	0.50	0.26	mg/l							
Nitrate/Nitrite-N	ND	0.26	0.072	mg/l							
Sulfate	ND	0.50	0.18	mg/l							
LCS Analyzed: 04/22/2005 (5D22051-BS1)											
Chloride	4.61	0.50	0.26	mg/l	5.00		92	90-110			M-3
Sulfate	9.30	0.50	0.18	mg/l	10.0		93	90-110			M-3
Batch: 5D22071 Extracted: 04/22/05											
Blank Analyzed: 04/27/2005 (5D22071-BLK1)											
Biochemical Oxygen Demand	ND	2.0	0.59	mg/l							
LCS Analyzed: 04/27/2005 (5D22071-BS1)											
Biochemical Oxygen Demand	202	100	30	mg/l	198		102	85-115			
LCS Dup Analyzed: 04/27/2005 (5D22071-BSD1)											
Biochemical Oxygen Demand	199	100	30	mg/l	198		101	85-115	1	20	
Batch: 5D22102 Extracted: 04/22/05											
Blank Analyzed: 04/22/2005 (5D22102-BLK1)											
Surfactants (MBAS)	ND	0.10	0.044	mg/l							
LCS Analyzed: 04/22/2005 (5D22102-BS1)											
Surfactants (MBAS)	0.251	0.10	0.044	mg/l	0.250		100	90-110			



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002
Report Number: IOD1705

Sampled: 04/22/05
Received: 04/22/05

METHOD BLANK/QC DATA

INORGANICS

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5D22102 Extracted: 04/22/05											
Matrix Spike Analyzed: 04/22/2005 (5D22102-MS1)						Source: IOD1705-01					
Surfactants (MBAS)	0.301	0.10	0.044	mg/l	0.250	0.077	90	50-125			
Matrix Spike Dup Analyzed: 04/22/2005 (5D22102-MSD1)						Source: IOD1705-01					
Surfactants (MBAS)	0.331	0.10	0.044	mg/l	0.250	0.077	102	50-125	9	20	
Batch: 5D23050 Extracted: 04/23/05											
Blank Analyzed: 04/23/2005 (5D23050-BLK1)											
Turbidity	ND	1.0	0.040	NTU							
Duplicate Analyzed: 04/23/2005 (5D23050-DUP1)						Source: IOD1610-01					
Turbidity	38.8	1.0	0.040	NTU		38			2	20	
Batch: 5D25050 Extracted: 04/25/05											
Blank Analyzed: 04/25/2005 (5D25050-BLK1)											
Sulfate	ND	0.50	0.18	mg/l							
LCS Analyzed: 04/25/2005 (5D25050-BS1)											
Sulfate	9.73	0.50	0.18	mg/l	10.0		97	90-110			M-3
Batch: 5D25054 Extracted: 04/25/05											
Blank Analyzed: 04/25/2005 (5D25054-BLK1)											
Oil & Grease	ND	5.0	0.94	mg/l							

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD1705

Sampled: 04/22/05
Received: 04/22/05

METHOD BLANK/QC DATA

INORGANICS

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC %REC	Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5D25054 Extracted: 04/25/05											
LCS Analyzed: 04/25/2005 (5D25054-BS1)											
Oil & Grease	22.0	5.0	0.94	mg/l	20.0		110	65-120			M-NRI
LCS Dup Analyzed: 04/25/2005 (5D25054-BSD1)											
Oil & Grease	20.0	5.0	0.94	mg/l	20.0		100	65-120	10	20	
Batch: 5D25087 Extracted: 04/25/05											
Blank Analyzed: 04/25/2005 (5D25087-BLK1)											
Total Dissolved Solids	ND	10	10	mg/l							
LCS Analyzed: 04/25/2005 (5D25087-BS1)											
Total Dissolved Solids	984	10	10	mg/l	1000		98	90-110			
Duplicate Analyzed: 04/25/2005 (5D25087-DUP1)											
Total Dissolved Solids	998	10	10	mg/l		Source: IOD1705-01			0	10	
Batch: 5D25090 Extracted: 04/25/05											
Blank Analyzed: 04/26/2005 (5D25090-BLK1)											
Total Cyanide	ND	5.0	2.2	ug/l							
LCS Analyzed: 04/26/2005 (5D25090-BS1)											
Total Cyanide	182	5.0	2.2	ug/l	200		91	90-110			
Matrix Spike Analyzed: 04/26/2005 (5D25090-MS1)											
Total Cyanide	175	5.0	2.2	ug/l	200	Source: IOD1667-01	ND	88		70-115	

Del Mar Analytical, Irvine
Michele Harper
Project Manager



Del Mar Analytical

17461 Derian Ave., Suite 100, Irvine, CA 92614 (949) 261-1022 FAX (949) 260-3297
 1014 E. Cooley Dr., Suite A, Colton, CA 92324 (909) 370-4667 FAX (909) 370-1046
 9484 Chesapeake Dr., Suite 805, San Diego, CA 92123 (858) 505-8596 FAX (858) 505-9689
 9830 South 51st St., Suite B-120, Phoenix, AZ 85044 (480) 785-0043 FAX (480) 785-0851
 2520 E. Sunset Rd. #3, Las Vegas, NV 89120 (702) 798-3620 FAX (702) 798-3621

MWH-Pasadena/Boeing 300 North Lake Avenue, Suite 1200 Pasadena, CA 91101 Attention: Bronwyn Kelly	Project ID: Routine Outfall 002 Report Number: IOD1705	Sampled: 04/22/05 Received: 04/22/05
--	---	---

METHOD BLANK/QC DATA

INORGANICS

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC %REC	RPD Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5D25090 Extracted: 04/25/05											
Matrix Spike Dup Analyzed: 04/26/2005 (5D25090-MSD1)						Source: IOD1667-01					
Total Cyanide	173	5.0	2.2	ug/l	200	ND	86	70-115	1	15	
Batch: 5D25093 Extracted: 04/25/05											
Duplicate Analyzed: 04/25/2005 (5D25093-DUP1)						Source: IOD1705-01					
Specific Conductance	1400	1.0	1.0	umhos/cm		1400			0	5	
Batch: 5D26085 Extracted: 04/26/05											
Blank Analyzed: 04/26/2005 (5D26085-BLK1)											
Ammonia-N (Distilled)	ND	0.50	0.30	mg/l							
LCS Analyzed: 04/26/2005 (5D26085-BS1)											
Ammonia-N (Distilled)	10.9	0.50	0.30	mg/l	10.0		109	80-115			
Matrix Spike Analyzed: 04/26/2005 (5D26085-MS1)						Source: IOD1488-01					
Ammonia-N (Distilled)	11.2	0.50	0.30	mg/l	10.0	2.8	84	70-120			
Matrix Spike Dup Analyzed: 04/26/2005 (5D26085-MSD1)						Source: IOD1488-01					
Ammonia-N (Distilled)	10.9	0.50	0.30	mg/l	10.0	2.8	81	70-120	3	15	
Batch: 5D26122 Extracted: 04/26/05											
Blank Analyzed: 04/26/2005 (5D26122-BLK1)											
Total Dissolved Solids	ND	10	10	mg/l							

Del Mar Analytical, Irvine
 Michele Harper
 Project Manager



Del Mar Analytical

17461 Derian Ave., Suite 100, Irvine, CA 92614 (949) 261-1022 FAX (949) 260-3297
 1014 E. Cooley Dr., Suite A, Colton, CA 92324 (909) 370-4667 FAX (909) 370-1046
 9484 Chesapeake Dr., Suite 805, San Diego, CA 92123 (858) 505-8596 FAX (858) 505-9689
 9830 South 51st St., Suite B-120, Phoenix, AZ 85044 (480) 785-0043 FAX (480) 785-0851
 2520 E. Sunset Rd. #3, Las Vegas, NV 89120 (702) 798-3620 FAX (702) 798-3621

MWH-Pasadena/Boeing
 300 North Lake Avenue, Suite 1200
 Pasadena, CA 91101
 Attention: Bronwyn Kelly

Project ID: Routine Outfall 002
 Report Number: IOD1705

Sampled: 04/22/05
 Received: 04/22/05

METHOD BLANK/QC DATA

INORGANICS

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5D26122 Extracted: 04/26/05											
LCS Analyzed: 04/26/2005 (5D26122-BS1)											
Total Dissolved Solids	956	10	10	mg/l	1000		96	90-110			
Duplicate Analyzed: 04/26/2005 (5D26122-DUP1)											
Total Dissolved Solids	1360	10	10	mg/l		1300			5	10	
Source: IOD1832-01											
Batch: 5D27058 Extracted: 04/27/05											
Blank Analyzed: 04/27/2005 (5D27058-BLK1)											
Perchlorate	ND	4.0	0.80	ug/l							
LCS Analyzed: 04/27/2005 (5D27058-BS1)											
Perchlorate	51.2	4.0	0.80	ug/l	50.0		102	85-115			
Matrix Spike Analyzed: 04/27/2005 (5D27058-MS1)											
Perchlorate	54.5	4.0	0.80	ug/l	50.0	1.5	106	80-120			
Source: IOD1820-02											
Matrix Spike Dup Analyzed: 04/27/2005 (5D27058-MSD1)											
Perchlorate	54.8	4.0	0.80	ug/l	50.0	1.5	107	80-120	1	20	
Source: IOD1820-02											
Batch: 5D27086 Extracted: 04/27/05											
Blank Analyzed: 04/27/2005 (5D27086-BLK1)											
Total Suspended Solids	ND	10	10	mg/l							
LCS Analyzed: 04/27/2005 (5D27086-BS1)											
Total Suspended Solids	987	10	10	mg/l	1000		99	85-115			

Del Mar Analytical, Irvine
 Michele Harper
 Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD1705

Sampled: 04/22/05

Received: 04/22/05

METHOD BLANK/QC DATA

INORGANICS

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5D27086 Extracted: 04/27/05											
Duplicate Analyzed: 04/27/2005 (5D27086-DUP1)						Source: IOD1644-01					
Total Suspended Solids	ND	10	10	mg/l		ND				10	

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD1705

Sampled: 04/22/05
Received: 04/22/05

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits appear in bold on this page.

LabNumber	Analysis	Analyte	Units	Result	MRL	Compliance Limit
IOD1705-01	413.1 Oil and Grease	Oil & Grease	mg/l	0.19	5.0	10.00
IOD1705-01	608-Pest Boeing 001/002 Q (LL)	alpha-BHC	ug/l	0	0.010	0.0100
IOD1705-01	624-Boeing 001/002 Q (Fr113+X)	1,1-Dichloroethene	ug/l	0	3.0	3.20
IOD1705-01	624-Boeing 001/002 Q (Fr113+X)	Trichloroethene	ug/l	0	5.0	5.00
IOD1705-01	625-Boeing 001/002 Q-LL	2,4,6-Trichlorophenol	ug/l	0	6.0	6.50
IOD1705-01	625-Boeing 001/002 Q-LL	2,4-Dinitrotoluene	ug/l	0	9.0	9.10
IOD1705-01	625-Boeing 001/002 Q-LL	Bis(2-ethylhexyl)phthalate	ug/l	1.30	5.0	4.00
IOD1705-01	625-Boeing 001/002 Q-LL	N-Nitrosodimethylamine	ug/l	0	8.0	8.10
IOD1705-01	625-Boeing 001/002 Q-LL	Pentachlorophenol	ug/l	0	8.0	8.20
IOD1705-01	BOD	Biochemical Oxygen Demand	mg/l	0.86	2.0	20
IOD1705-01	Chloride - 300.0	Chloride	mg/l	49	5.0	150
IOD1705-01	Copper-200.8	Copper	ug/l	2.80	2.0	7.10
IOD1705-01	Cyanide-335.2 5ppb	Total Cyanide	ug/l	-5	5.0	4.30
IOD1705-01	Lead-200.8	Lead	ug/l	0.33	1.0	2.60
IOD1705-01	MBAS - SM5540-C	Surfactants (MBAS)	mg/l	0.077	0.10	0.50
IOD1705-01	Mercury - 245.1	Mercury	ug/l	0.033	0.20	0.20
IOD1705-01	Nitrogen, NO3+NO2 -N	Nitrate/Nitrite-N	mg/l	0.032	0.26	8.00
IOD1705-01	Perchlorate 314.0	Perchlorate	ug/l	0	4.0	6.00
IOD1705-01	Sulfate-300.0	Sulfate	mg/l	400	5.0	300
IOD1705-01	TDS - SM 2540C	Total Dissolved Solids	mg/l	1000	10	950
IOD1705-01RE1	Sulfate-300.0	Sulfate	mg/l	420	5.0	300
IOD1705-01RE1	TDS - SM 2540C	Total Dissolved Solids	mg/l	1000	10	950
IOD1705-02	624-Boeing 001/002 Q (Fr113+X)	1,1-Dichloroethene	ug/l	0	3.0	3.20
IOD1705-02	624-Boeing 001/002 Q (Fr113+X)	Trichloroethene	ug/l	0	5.0	5.00

MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD1705

Sampled: 04/22/05
Received: 04/22/05

DATA QUALIFIERS AND DEFINITIONS

- J** Estimated value. Analyte detected at a level less than the Reporting Limit (RL) and greater than or equal to the Method Detection Limit (MDL). The user of this data should be aware that this data is of limited reliability.
- M1** The MS and/or MSD were above the acceptance limits due to sample matrix interference. See Blank Spike (LCS).
- M-3** Results exceeded the linear range in the MS/MSD and therefore are not available for reporting. The batch was accepted based on acceptable recovery in the Blank Spike (LCS).
- M-NRI** There was no MS/MSD analyzed with this batch due to insufficient sample volume. See Blank Spike/Blank Spike Duplicate.
- ND** Analyte NOT DETECTED at or above the reporting limit or MDL, if MDL is specified.
- RPD** Relative Percent Difference



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD1705

Sampled: 04/22/05
Received: 04/22/05

Certification Summary

Del Mar Analytical, Irvine

Method	Matrix	Nelac	California
EPA 120.1	Water	X	X
EPA 160.2	Water	X	X
EPA 160.5	Water	X	X
EPA 180.1	Water	X	X
EPA 200.8	Water	X	X
EPA 245.1	Water	X	X
EPA 300.0	Water	X	X
EPA 314.0	Water	N/A	X
EPA 335.2	Water	X	X
EPA 350.2	Water		X
EPA 405.1	Water	X	X
EPA 413.1	Water	X	X
EPA 608	Water	X	X
EPA 624	Water	X	X
EPA 625	Water	X	X
SM2540C	Water	X	X
SM5540-C	Water	X	X

Nevada and NELAP provide analyte specific accreditations. Analyte specific information for Del Mar Analytical may be obtained by contacting the laboratory or visiting our website at www.dmalabs.com.

Subcontracted Laboratories

Alta Analytical California Cert #1640, Nevada Cert #CA-413

1104 Windfield Way - El Dorado Hills, CA 95762

Analysis Performed: 1613-Dioxin-HR
Samples: IOD1705-01

Analysis Performed: EDD + Level 4
Samples: IOD1705-01

Del Mar Analytical, Irvine
Michele Harper
Project Manager



17461 Denian Ave., Irvine CA 92606 (949) 261-1022 FAX (949) 261-1228
1014 E. Cooley Dr., Suite A, Colton, CA 92324 (909) 370-4667 FAX (949) 370-1046
9484 Chesapeake Dr., Suite 805, San Diego, CA 92123 (858) 505-8596 FAX (858) 505-9689
9830 South 51st St., Suite B-120, Phoenix, AZ 85044 (480) 785-0043 FAX (480) 785-0851
2520 E. Sunset Rd. #3, Las Vegas, NV 89120 (702) 798-3620 FAX (702) 798-3621

June 20, 2005

MWH- Pasadena / Boeing
300 North Lake Avenue, Suite 1200
Pasadena , CA 91101

Attention: Bronwyn Kelly
Project: Routine Outfall 002
Sampled: 04/22/05
Del Mar Analytical Number: IOD1705

Dear Ms. Kelly:

Alta Analytical Laboratories performed the EPA Method 1613 for tetra-through-octa chlorinated dioxins and furans for the project referenced above. Please use the following cross-reference table when reviewing your results.

MWH ID	Del Mar ID	Alta ID
Outfall 002	IOD1705-01	26089-001

Attached is the original report from the subcontract laboratory. If you have any questions or require further assistance, please do not hesitate to contact me at (949) 261-1022, extension 215.

Sincerely yours,

DEL MAR ANALYTICAL


Michele Harper
Project Manager

Enclosure



May 17, 2005

Alta Project I.D.: 26089

Ms. Michele Harper
Del Mar Analytical, Irvine
17461 Derian Avenue, Suite 100
Irvine, CA 92614

Dear Ms. Harper,

Enclosed are the results for the one aqueous sample received at Alta Analytical Laboratory on April 26, 2005 under your Project Name "IOD1705". This sample was extracted and analyzed using EPA Method 1613 for tetra-through-octa chlorinated dioxins and furans. A standard turnaround time was provided for this work.

The following report consists of a Sample Inventory (Section I), Analytical Results (Section II) and the Appendix, which contains the chain-of-custody, a list of data qualifiers and abbreviations, Alta's current certifications, and copies of the raw data (if requested).

Alta Analytical Laboratory is committed to serving you effectively. If you require additional information, please contact me at 916-933-1640 or by email at mmaier@altalab.com. Thank you for choosing Alta as part of your analytical support team.

Sincerely,

Martha M. Maier
Director of HRMS Services



The Analytical Laboratory certifies that the report is in strict accordance with all the requirements set forth by NELAP for those organizations using the methods. This report should not be reproduced or excerpted without the written approval of ALTA.



Alta Analytical Laboratory Inc.

1104 Windfield Way
El Dorado Hills, CA 95762

FAX (916) 673-0106
(916) 933-1640

Section I: Sample Inventory Report

Date Received: 4/26/2005

Alta Lab. ID

Client Sample ID

26089-001

IOD1705-01

SECTION II



Method Blank		EPA Method 1613					
Matrix:	Aqueous	QC Batch No.:	6778	Lab Sample:	0-MB001		
Sample Size:	1.000 L	Date Extracted:	11-May-05	Date Analyzed DB-5:	13-May-05		
				Date Analyzed DB-225:	NA		
Analyte	Conc. (ug/L)	DL ^a	EMPC ^b	Labeled Standard	%R	LCL-UCL ^d	Qualifiers
2,3,7,8-TCDD	ND	0.000000583		IS 13C-2,3,7,8-TCDD	89.2	25 - 164	
1,2,3,7,8-PeCDD	ND	0.000000577		13C-1,2,3,7,8-PeCDD	89.9	25 - 181	
1,2,3,4,7,8-HxCDD	ND	0.00000108		13C-1,2,3,4,7,8-HxCDD	79.4	32 - 141	
1,2,3,6,7,8-HxCDD	ND	0.00000104		13C-1,2,3,6,7,8-HxCDD	83.1	28 - 130	
1,2,3,7,8,9-HxCDD	ND	0.00000100		13C-1,2,3,4,6,7,8-HpCDD	80.1	23 - 140	
1,2,3,4,6,7,8-HpCDD	ND	0.00000132		13C-OCDD	64.9	17 - 157	
OCDD	ND	0.00000357		13C-2,3,7,8-TCDF	82.3	24 - 169	
2,3,7,8-TCDF	ND	0.000000734		13C-1,2,3,7,8-PeCDF	82.8	24 - 185	
1,2,3,7,8-PeCDF	ND	0.00000110		13C-2,3,4,7,8-PeCDF	83.3	21 - 178	
2,3,4,7,8-PeCDF	ND	0.000000966		13C-1,2,3,4,7,8-HxCDF	87.8	26 - 152	
1,2,3,4,7,8-HxCDF	ND	0.000000314		13C-1,2,3,6,7,8-HxCDF	88.7	26 - 123	
1,2,3,6,7,8-HxCDF	ND	0.000000318		13C-2,3,4,6,7,8-HxCDF	85.2	28 - 136	
2,3,4,6,7,8-HxCDF	ND	0.000000408		13C-1,2,3,7,8,9-HxCDF	81.4	29 - 147	
1,2,3,7,8,9-HxCDF	ND	0.000000708		13C-1,2,3,4,6,7,8-HpCDF	80.7	28 - 143	
1,2,3,4,6,7,8-HpCDF	ND	0.000000597		13C-1,2,3,4,7,8,9-HpCDF	75.7	26 - 138	
1,2,3,4,7,8,9-HpCDF	ND	0.000000804		13C-OCDF	63.1	17 - 157	
OCDF	ND	0.00000171		CRS 37Cl-2,3,7,8-TCDD	105	35 - 197	
Totals				Footnotes			
Total TCDD	ND	0.000000583		a. Sample specific estimated detection limit.			
Total PeCDD	ND	0.000000577		b. Estimated maximum possible concentration.			
Total HxCDD	ND	0.00000104		c. Method detection limit.			
Total HpCDD	ND	0.00000132		d. Lower control limit - upper control limit.			
Total TCDF	ND	0.000000734					
Total PeCDF	ND	0.00000103					
Total HxCDF	ND	0.000000417					
Total HpCDF	ND	0.000000686					

Analyst: JMH

Approved By: Martha M. Maier 17-May-2005 12:04

EPA Method 1613

OPR Results		Lab Sample: 0-OPR001		Date Analyzed DB-5: 13-May-05		Date Analyzed DB-225: NA	
Matrix:	Aqueous <th>QC Batch No.:</th> <td>6778 <th>Date Analyzed DB-5:</th> <td>13-May-05 <th>Date Analyzed DB-225:</th> <td>NA </td></td></td>	QC Batch No.:	6778 <th>Date Analyzed DB-5:</th> <td>13-May-05 <th>Date Analyzed DB-225:</th> <td>NA </td></td>	Date Analyzed DB-5:	13-May-05 <th>Date Analyzed DB-225:</th> <td>NA </td>	Date Analyzed DB-225:	NA
Sample Size:	1.000 L <th>Date Extracted:</th> <td>11-May-05 <th colspan="4"></th> </td>	Date Extracted:	11-May-05 <th colspan="4"></th>				
Analyte	Spike Conc.	Conc. (ng/mL)	OPR Limits	Labeled Standard	%R	LCL-UCL	
2,3,7,8-TCDD	10.0	10.5	6.7 - 15.8	IS 13C-2,3,7,8-TCDD	81.9	25 - 164	
1,2,3,7,8-PeCDD	50.0	49.4	35 - 71	13C-1,2,3,7,8-PeCDD	84.4	25 - 181	
1,2,3,4,7,8-HxCDD	50.0	49.8	35 - 82	13C-1,2,3,4,7,8-HxCDD	82.8	32 - 141	
1,2,3,6,7,8-HxCDD	50.0	51.6	38 - 67	13C-1,2,3,6,7,8-HxCDD	79.2	28 - 130	
1,2,3,7,8,9-HxCDD	50.0	48.8	32 - 81	13C-1,2,3,4,6,7,8-HpCDD	76.9	23 - 140	
1,2,3,4,6,7,8-HpCDD	50.0	49.4	35 - 70	13C-OCDD	66.6	17 - 157	
OCDD	100	98.6	78 - 144	13C-2,3,7,8-TCDF	77.8	24 - 169	
2,3,7,8-TCDF	10.0	10.1	7.5 - 15.8	13C-1,2,3,7,8-PeCDF	78.5	24 - 185	
1,2,3,7,8-PeCDF	50.0	47.7	40 - 67	13C-2,3,4,7,8-PeCDF	78.2	21 - 178	
2,3,4,7,8-PeCDF	50.0	47.8	34 - 80	13C-1,2,3,4,7,8-HxCDF	87.6	26 - 152	
1,2,3,4,7,8-HxCDF	50.0	49.0	36 - 67	13C-1,2,3,6,7,8-HxCDF	89.8	26 - 123	
1,2,3,6,7,8-HxCDF	50.0	49.6	42 - 65	13C-2,3,4,7,8-HxCDF	87.1	28 - 136	
2,3,4,6,7,8-HxCDF	50.0	48.8	35 - 78	13C-1,2,3,7,8,9-HxCDF	80.5	29 - 147	
1,2,3,7,8,9-HxCDF	50.0	50.4	39 - 65	13C-1,2,3,4,6,7,8-HpCDF	78.9	28 - 143	
1,2,3,4,6,7,8-HpCDF	50.0	49.6	41 - 61	13C-1,2,3,4,7,8,9-HpCDF	68.9	26 - 138	
1,2,3,4,7,8,9-HpCDF	50.0	50.5	39 - 69	13C-OCDF	63.5	17 - 157	
OCDF	100	96.0	63 - 170	CRS 37Cl-2,3,7,8-TCDD	104	35 - 197	

Analyst: JMH
 Approved By: Martha M. Maier
 Date: 17-May-2005 08:27



Sample ID: IOD1705-01

EPA Method 1613

Client Data		Sample Data		Laboratory Data	
Name:	Del Mar Analytical, Irvine	Matrix:	Aqueous	Lab Sample:	26089-001
Project:	IOD1705	Sample Size:	0.996 L	QC Batch No.:	6778
Date Collected:	22-Apr-05			Date Analyzed DB-5:	16-May-05
Time Collected:	1100			Date Analyzed DB-225:	NA
Analyte	Conc. (ug/L)	DL ^a	EMPC ^b	Labeled Standard	%R LCL-UCL ^d Qualifiers
2,3,7,8-TCDD	ND	0.00000121		IS 13C-2,3,7,8-TCDD	84.9 25 - 164
1,2,3,7,8-PeCDD	ND	0.00000123		13C-1,2,3,7,8-PeCDD	82.4 25 - 181
1,2,3,4,7,8-HxCDD	ND	0.00000244		13C-1,2,3,4,7,8-HxCDD	73.2 32 - 141
1,2,3,6,7,8-HxCDD	ND	0.00000235		13C-1,2,3,6,7,8-HxCDD	83.6 28 - 130
1,2,3,7,8,9-HxCDD	ND	0.00000226		13C-1,2,3,4,6,7,8-HpCDD	76.1 23 - 140
1,2,3,4,6,7,8-HpCDD	0.00000269			13C-OCDD	63.4 17 - 157
OCDD	0.0000213			13C-2,3,7,8-TCDF	77.9 24 - 169
2,3,7,8-TCDF	ND	0.00000130		13C-1,2,3,7,8-PeCDF	79.8 24 - 185
1,2,3,7,8-PeCDF	ND	0.00000190		13C-2,3,4,7,8-PeCDF	79.1 21 - 178
2,3,4,7,8-PeCDF	ND	0.00000168		13C-1,2,3,4,7,8-HxCDF	81.3 26 - 152
1,2,3,4,7,8-HxCDF	ND		0.000000762	13C-1,2,3,6,7,8-HxCDF	89.2 26 - 123
1,2,3,6,7,8-HxCDF	ND		0.000000810	13C-2,3,4,6,7,8-HxCDF	83.2 28 - 136
2,3,4,6,7,8-HxCDF	ND		0.000000951	13C-1,2,3,7,8,9-HxCDF	80.5 29 - 147
1,2,3,7,8,9-HxCDF	ND	0.00000150		13C-1,2,3,4,6,7,8-HpCDF	82.7 28 - 143
1,2,3,4,6,7,8-HpCDF	ND	0.00000102		13C-1,2,3,4,7,8,9-HpCDF	80.2 26 - 138
1,2,3,4,7,8,9-HpCDF	ND	0.00000140		13C-OCDF	67.7 17 - 157
OCDF	ND	0.00000384		CRS 37Cl-2,3,7,8-TCDD	95.8 35 - 197
Totals					
Total TCDD	ND	0.00000121			
Total PeCDD	ND	0.00000123			
Total HxCDD	ND	0.00000234			
Total HpCDD	0.00000551				
Total TCDF	ND	0.00000130			
Total PeCDF	ND	0.00000179			
Total HxCDF	ND		0.00000252		
Total HpCDF	ND	0.00000118			
Footnotes					
a. Sample specific estimated detection limit.					
b. Estimated maximum possible concentration.					
c. Method detection limit.					
d. Lower control limit - upper control limit.					

Analyst: WJL

Approved By: Martha M. Maier 17-May-2005 12:04

APPENDIX

DATA QUALIFIERS & ABBREVIATIONS

B	This compound was also detected in the method blank.
D	The amount reported is the maximum possible concentration due to possible chlorinated diphenylether interference.
H	The signal-to-noise ratio is greater than 10:1.
I	Chemical Interference
J	The amount detected is below the Lower Calibration Limit of the instrument.
P	Homologue totals include any coplanar PCBs detected at concentrations less than the reporting limit.
*	See Cover Letter
Conc.	Concentration
DL	Sample-specific estimated detection limit
MDL	The minimum concentration of a substance that can be measured and reported with 99% confidence that the analyte concentration is greater than zero in the matrix tested.
EMPC	Estimated Maximum Possible Concentration
NA	Not applicable
RL	Reporting Limit – concentrations that correspond to low calibration point
ND	Not Detected
TEQ	Toxic Equivalency

Unless otherwise noted, solid sample results are reported in dry weight. Tissue samples are reported in wet weight.

The control limits are “interim limits only” until in-house limits are utilized.

CURRENT CERTIFICATIONS

NELAP — (Primary AA: California, Certificate No. 02102CA)

Department of the Navy

U.S. Army Corps of Engineers

U.S. EPA Region 5

Bureau of Reclamation — Mid-Pacific Region — (MP-470, Res-1.10)

Commonwealth of Kentucky — (Certificate No. 90063)

Commonwealth of Virginia — (Certificate No. 00013)

State of Alaska, Department of Environmental Conservation — (Certificate No. OS-00197)

State of Arizona — (Certificate No. AZ0639)

State of Arkansas, Department of Health — (Approval granted through CA certification)

State of Arkansas, Department of Environmental Quality

State of California — (Certificate No. 1640)

State of Colorado

State of Connecticut — (Certificate No. PH-0182)

State of Florida — (Certificate No. 87456)

State of Louisiana, Department of Health and Hospitals — (Certificate No. LA000014)

State of Louisiana, Department of Environmental Quality

State of Maine

State of Michigan (Certificate No. 81178087)

State of Mississippi — (Approval granted through CA certification)

State of Nevada — (Certificate No. CA413)

State of New Jersey — (Certificate No. CA003)

State of New York, Department of Health — (Certificate No. 11411)

State of North Carolina — (Certification No. 06700)

State of North Dakota, Department of Health — (Certificate No. R-078)

State of New Mexico

State of Oklahoma – (D9919)

State of Oregon – (Certificate No. CA413)

State of Pennsylvania — (Certificate No. 68-490)

State of South Carolina — (Certificate No. 87002001)

State of Tennessee — (Certificate No. 02996)

State of Texas — (Certificate No. TX247-1000A)

State of Utah — (Certificate No. E-201)

State of Washington – (Certification No. C091)

State of Wisconsin — (Certificate No. 998036160)

State of Wyoming — (USEPA Region 8 Ref: 8TMS-Q)



17461 Derian Ave. Suite 100, Irvine, CA 92614 Ph (949) 261-1022 Fax (949) 261-1228
 1014 E. Cooley Dr., Suite A, Colton, CA 92324 Ph (909) 370-4667 Fax (909) 370-1046
 9484 Chesapeake Drive, Suite 805, San Diego, CA 92123 Ph (619) 505-9596 Fax (619) 505-9689
 9830 South 51st Street, Suite B-120, Phoenix, AZ 85044 Ph (480) 785-0043 Fax (480) 785-0851
 2520 E. Sunset Rd., Suite #3, Las Vegas, NV 89120 Ph (702) 798-3820 Fax (702) 798-3821

SUBCONTRACT ORDER - PROJECT # IOD1705

SENDING LABORATORY:

Del Mar Analytical, Irvine
 17461 Derian Avenue, Suite 100
 Irvine, CA 92614
 Phone: (949) 261-1022
 Fax: (949) 261-1228
 Project Manager: Michele Harper

RECEIVING LABORATORY:

Alta Analytical
 1104 Windfield Way
 El Dorado Hills, CA 95762
 Phone: (916) 933-1640
 Fax: (916) 673-0106

26089
0.7°C

Standard TAT is requested unless specific due date is requested => Due Date: _____ Initials: _____

Analysis	Expiration	Comments
Sample ID: IOD1705-01 Water	Sampled: 04/22/05 11:00	Instant Notification
1613-Dioxin-HR	04/29/05 11:00	J flags, 17 congeners, no TEQ, sub=Alta, DP to AMEC
EDD + Level 4	05/20/05 11:00	Excel EDD email to pm, Include Std logs for Lvl IV

Containers Supplied:
 1 L Amber (IOD1705-01G)
 1 L Amber (IOD1705-01H)

SAMPLE INTEGRITY:

All containers intact: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Sample labels/COC agree: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Samples Received On Ice: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
Custody Seals Present: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	Samples Preserved Properly: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Samples Received at (temp): <u>0.7°C</u>

Released By: Steve Chan Date: 4-25-05 Time: 1700 Received By: Letitia J. Benedict Date: 4/26/05 Time: 0800

Released By: _____ Date: _____ Time: _____ Received By: _____ Date: _____ Time: _____

STANDARD OPERATING PROCEDURE

Attachment 10.B.1

SAMPLE LOG-IN CHECKLIST

ALTA Project No.: 26089

1. Date Samples Arrived: <u>4/26/05 0850</u> Initials: <u>BBB</u> Location: <u>WR-2</u>			
2. Time / Date logged in: <u>1100 4/26/05</u> Initials: <u>BBB</u> Location: <u>WK-2</u>			
3. Samples Arrived By: (circle) <u>FedEx</u> UPS World Courier Other:			
4. Shipping Preservation: (circle) <u>Ice</u> <u>Blue Ice</u> / Dry Ice / None Temp °C <u>0.7°C</u>			
5. Shipping Container(s) Intact? If not, describe condition in comment section.	YES	NO	NA
	✓		
6. Shipping Container(s) Custody Seals Present? Intact? If not intact, describe condition in comment section.	✓		
	✓		
7. Shipping Documentation Present? (circle) Shipping Label <u>Airbill</u> Tracking Number <u>7910 5131 7866</u>	✓		
8. Sample Custody Seal(s) Present? No. of Seals _____ or Seal No. Intact? If not intact, describe condition in comment section.		✓	✓
9. Sample Container Intact? If no, indicate sample condition in comment section.	✓		
10. Chain of Custody (COC) or other Sample Documentation Present?	✓		
11. COC/Documentation Acceptable? If no, complete COC Anomaly Form.	✓		
12. Shipping Container (circle): ALTA <u>Client</u> Retain or <u>Return</u> or Disposed			
13. Container(s) and/or Bottle(s) Requested?		✓	
14. Drinking Water Sample? (HRMS Only) If yes, Acceptable Preservation? Y or N Preservation Info From? (circle) COC or Sample Container or None Noted			✓

Comments: Sampler's initials found on sample label

ALTA Analytical Laboratory
El Dorado Hills, CA 95762


SOP# CH10B_R18, Page 6 of 12

CONTRACT COMPLIANCE SCREENING FORM FOR HARDCOPY DATA

AMEC Earth & Environmental
 550 South Wadsworth Boulevard
 Suite 500
 Lakewood, CO 80226

Package ID T711DF46
 Task Order 313150010
 SDG No. IOD1705
 No. of Analyses 1

Laboratory Alta
 Reviewer K. Shadowlight
 Analysis/Method Dioxins

Date: May 25, 2005
 Reviewer's Signature


ACTION ITEMS*	
1. Case Narrative Deficiencies	
2. Out of Scope Analyses	
3. Analyses Not Conducted	
4. Missing Hardcopy Deliverables	
5. Incorrect Hardcopy Deliverables	
6. Deviations from Analysis Protocol, e.g.,	Qualifications were assigned for the following:
Holding Times	* Detect below the lower method calibration level
GC/MS Tune/Inst. Performance	* EMPCs
Calibration	
Method blanks	
Surrogates	
Matrix Spike/Dup LCS	
Field QC	
Internal Standard Performance	
Compound Identification and Quantitation	
System Performance	
COMMENTS^b	

^a Subcontracted analytical laboratory is not meeting contract and/or method requirements.
^b Differences in protocol have been adopted by the laboratory but no action against the laboratory is required.

Data Qualifier Reference Table

Qualifier	Organics	Inorganics
U	The analyte was analyzed for, but was not detected above the reported sample quantitation limit.	The material was analyzed for, but was not detected above the level of the associated value. The associated value is either the sample quantitation limit or the sample detection limit.
J	The analyte was positively identified; the associated numerical value is the approximate concentration of the analyte in the sample.	The associated value is an estimated quantity.
N	The analysis indicates the presence of an analyte for which there is presumptive evidence to make a "tentative identification."	Not applicable.
NJ	The analysis indicates the presence of an analyte that has been "tentatively identified" and the associated numerical value represents its approximate concentration.	Not applicable.
UJ	The analyte was not deemed above the reported sample quantitation limit. However, the reported quantitation limit is approximate and may or may not represent the actual limit of quantitation necessary to accurately and precisely measure the analyte in the sample.	The material was analyzed for, but was not detected. The associated value is an estimate and may be inaccurate or imprecise.
R	The sample results are rejected due to serious deficiencies in the ability to analyze the sample and to meet quality control criteria. The presence or absence of the analyte cannot be verified.	The data are unusable. (Note: Analyte may or may not be present).

Qualification Code Reference Table

Qualifier	Organics	Inorganics
H	Holding times were exceeded.	Holding times were exceeded.
S	Surrogate recovery was outside QC limits.	The sequence or number of standards used for the calibration was incorrect
C	Calibration %RSD or %D were noncompliant.	Correlation coefficient is <0.995.
R	Calibration RRF was <0.05.	%R for calibration is not within control limits.
B	Presumed contamination from preparation (method) blank.	Presumed contamination from preparation (method) or calibration blank.
L	Laboratory Blank Spike/Blank Spike Duplicate %R was not within control limits.	Laboratory Control Sample %R was not within control limits.
Q	MS/MSD recovery was poor or RPD high.	MS recovery was poor.
E	Not applicable.	Duplicates showed poor agreement.
I	Internal standard performance was unsatisfactory.	ICP ICS results were unsatisfactory.
A	Not applicable.	ICP Serial Dilution %D were not within control limits.
M	Tuning (BFB or DFTPP) was noncompliant.	Not applicable.
T	Presumed contamination from trip blank.	Not applicable.
+	False positive - reported compound was not present. Not applicable.	Not applicable.
-	False negative - compound was present but not reported.	Not applicable.
F	Presumed contamination from FB, or ER.	Presumed contamination from FB or ER.
\$	Reported result or other information was incorrect.	Reported result or other information was incorrect.
?	TIC identity or reported retention time has been changed.	Not applicable.
D	The analysis with this flag should not be used because another more technically sound analysis is available.	The analysis with this flag should not be used because another more technically sound analysis is available.
P	Instrument performance for pesticides was poor.	Post Digestion Spike recovery was not within control limits.
DNQ	The compound was detected between the MDL and the RL and, by definition, is considered an estimated value.	The compound was detected between the MDL and the RL and, by definition, is considered an estimated value.
#	Unusual problems found with the data that have been described in Section 2.#, "Data Validation Findings." The number following the asterisk () will indicate the subsection where a description of the problem can be found (eg. *1 would indicate a sample was not within temperature limits).	Unusual problems found with the data that have been described in Section 2.#, "Data Validation Findings." The number following the asterisk (*) will indicate the subsection where a description of the problem can be found (eg. *1 would indicate a sample was not within temperature limits).



DATA VALIDATION REPORT

NPDES Monitoring

ANALYSIS: DIOXINS/FURANS

SAMPLE DELIVERY GROUPS: IOD1705

Prepared by

AMEC—Denver Operations
550 South Wadsworth Boulevard, Suite 500
Lakewood, Colorado 80226

1. INTRODUCTION

Task Order Title: NPDES Monitoring
Contract Task Order #: 313150010
Sample Delivery Group #: IOD1705
Project Manager: B. McIlvaine
Matrix: Water
Analysis: Dioxins/Furans
QC Level: Level IV
No. of Samples: 1
No. of Reanalyses/Dilutions: 0
Reviewer: K. Shadowlight
Date of Review: May 25, 2005

The samples listed in Table 1 were validated based on the guidelines outlined in the *AMEC Data Validation Procedure for Dioxins and Furans (DVP-19, Rev. 1)*, *EPA Method 1613*, and the *National Functional Guidelines For Chlorinated Dioxin/Furan Data Review (8/02)*. Any deviations from these procedures and guidelines are documented herein. Qualifiers were applied in cases where the data did not meet the required QC criteria or where special consideration by the data user is required. Data qualifiers were placed on Form Is with the associated qualification codes. Analytes that were rejected for any reason are denoted on the Form I as having only the "R" data qualifier and associated qualification code(s) denoting the reason for rejection. Any additional problems with the data that may have resulted in an estimated value were not denoted by a qualification code since the data had already been rejected.

Table 1. Sample Identification

Client ID	Laboratory ID (Del Mar)	Laboratory ID (Alta)	Matrix	COC Method
Outfall 002	IOD1705-01	26089-001	water	1613

2. DATA VALIDATION FINDINGS

2.1 SAMPLE MANAGEMENT

Following are findings associated with sample management:

2.1.1 Sample Preservation, Handling, and Transport

The sample in this SDG was received at Del Mar Analytical within the temperature limits of 4°C ±2°C. The sample was shipped to Alta for dioxin/furan analysis and was received below the temperature limits of 4°C ±2°C at 0.7°C; however, as the sample was not noted to have been frozen or damaged, no qualifications were required. According to the laboratory login sheet, the sample was received intact and in good condition at both laboratories. No qualifications were required.

2.1.2 Chain of Custody

The COC and transfer COC were legible and signed by the appropriate field and laboratory personnel, and accounted for the analysis presented in this SDG. As the sample was couriered directly to Del Mar Analytical, custody seals were not required. The cooler received by Alta had custody seals present and intact; however, custody seals were not present on the sample container. The EPA ID was added to the sample result summary by the reviewer. No qualifications were required.

2.1.3 Holding Times

The sample was extracted and analyzed within a year of collection. No qualifications were required.

2.2 INSTRUMENT PERFORMANCE

Following are findings associated with instrument performance:

2.2.1 GC Column Performance

A Windows Defining Mix (WDM) containing the first and last eluting congeners of each descriptor and isomer specificity compounds was not analyzed prior to the initial calibration sequence or at the beginning of each analytical sequence; however, the first and last eluting congeners and isomer specificity compounds were added to the midpoint of the initial calibration and to the continuing calibration standards (see section 2.3.2). The GC column performance in the calibrations was acceptable, with the height of the valley between the closely eluting isomers and 2,3,7,8-TCDD reported as less than 25%. No qualifications were required.

2.2.2 Mass Spectrometer Performance

The mass spectrometer performance was acceptable with the static resolving power greater than 10,000. No qualifications were required.

2.3 CALIBRATION

2.3.1 Initial Calibration

The initial calibration was analyzed 04/23/05. The calibration consisted of six concentration level standards (CS0 through CS5) analyzed to verify instrument linearity. The initial calibration was acceptable with %RSDs $\leq 20\%$ for the 16 native compounds (calibration by isotope dilution) and $\leq 35\%$ for the one native and all labeled compounds (calibration by internal standard). The relative retention times and ion abundance ratios were within the QC limits listed in Method 1613 for all standards. A representative number of %RSDs were verified from the raw data, and no calculation or transcription errors were noted. No qualifications were required.

2.3.2 Continuing Calibration

Calibration verification (VER) consisted of a mid-level standard (CS3) analyzed at the beginning of each analytical sequence. The VER was acceptable with the concentrations within the acceptance criteria listed in Table 6 of EPA Method 1613. The ion abundance ratios and relative retention times were within the method QC limits. A representative number of %Ds were verified from the raw data, and no calculation or transcription errors were noted. No qualifications were required.

WDM and isomer specificity compounds were added to the VER standard instead of being analyzed separately, as noted in section 2.2.1 of this report. No adverse effect was observed with this practice.

2.4 BLANKS

One method blank (6778-MB001) was extracted and analyzed with the sample in this SDG. There were no target compound detects reported in the method blank. A review of the method blank raw data and chromatograms indicated no false negatives. No qualifications were required.

2.5 BLANK SPIKES AND LABORATORY CONTROL SAMPLES

One Ongoing Precision Recovery (OPR) sample (6778-OPR001) was extracted and analyzed with the sample in this SDG. All recoveries were within the acceptance criteria listed in Table 6 of Method 1613. No qualifications were required.

2.6 MATRIX SPIKE/MATRIX SPIKE DUPLICATE

MS/MSD analyses were not performed in this SDG. Evaluation of method accuracy was based on the OPR results. No qualifications were required.

2.7 FIELD QC SAMPLES

Following are findings associated with field QC:

2.7.1 Field Blanks and Equipment Rinsates

The sample in this SDG had no associated field QC samples. No qualifications were required.

2.7.2 Field Duplicates

No field duplicate samples were identified for this SDG.

2.8 INTERNAL STANDARDS

The labeled standard recoveries were within the acceptance criteria listed in Table 7 of Method 1613. No qualifications were required.

2.9 COMPOUND IDENTIFICATION

The laboratory analyzed for polychlorinated dioxins/furans by EPA Method 1613. The compound identifications were verified from the raw data and no false negatives or positives were noted. No qualifications were required.

2.10 COMPOUND QUANTIFICATION AND REPORTED DETECTION LIMITS

Compound quantitation was verified from the raw data. The laboratory calculated and reported compound-specific detection limits. Any detects below the lower method calibration level (MCL) were qualified as estimated, "J." If the concentration of any component of the total was below the lower method calibration level (MCL), the total detect was qualified as estimated, "J." Any reported EMPC was qualified as an estimated nondetect, "UJ." The results and reporting limits were reported in ug/L. No further qualifications were required.



Sample ID: IOD1705-01 *Outfall 002*

EPA Method 1613

Client Data
 Name: Del Mar Analytical, Irvine
 Project: IOD1705
 Date Collected: 22-Apr-05
 Time Collected: 1100

Sample Data
 Matrix: Aqueous
 Sample Size: 0.996 L

Laboratory Data
 Lab Sample: 26089-001
 QC Batch No.: 6778
 Date Analyzed DB-5: 16-May-05
 Date Analyzed DB-22.5: NA

Date Received: 26-Apr-05
 Date Extracted: 11-May-05

Analyte	Conc. (ug/L)	DL ^a	EMPC ^b	Qualifiers	Labeled Standard	%R	LCL-UCL ^d	Qualifiers
2,3,7,8-TCDD	ND	0.00000121			13C-2,3,7,8-TCDD	84.9	25 - 164	
1,2,3,7,8-PeCDD	ND	0.00000123			13C-1,2,3,7,8-PeCDD	82.4	25 - 181	
1,2,3,4,7,8-HxCDD	ND	0.00000244			13C-1,2,3,4,7,8-HxCDD	73.2	32 - 141	
1,2,3,6,7,8-HxCDD	ND	0.00000235			13C-1,2,3,6,7,8-HxCDD	83.6	28 - 130	
1,2,3,7,8,9-HxCDD	ND	0.00000226			13C-1,2,3,4,6,7,8-HpCDD	76.1	23 - 140	
1,2,3,4,6,7,8-HpCDD	0.00000269			J	13C-OCDD	63.4	17 - 157	
OCDD	0.0000213			J	13C-2,3,7,8-TCDF	77.9	24 - 169	
2,3,7,8-TCDF	ND	0.00000130			13C-1,2,3,7,8-PeCDF	79.8	24 - 185	
1,2,3,7,8-PeCDF	ND	0.00000190			13C-2,3,4,7,8-PeCDF	79.1	21 - 178	
2,3,4,7,8-PeCDF	ND	0.00000168			13C-1,2,3,4,7,8-HxCDF	81.3	26 - 152	
1,2,3,4,7,8-HxCDF	0.000000762				13C-1,2,3,6,7,8-HxCDF	89.2	26 - 123	
1,2,3,6,7,8-HxCDF	0.000000810				13C-2,3,4,6,7,8-HxCDF	83.2	28 - 136	
2,3,4,6,7,8-HxCDF	0.000000951				13C-1,2,3,7,8,9-HxCDF	80.5	29 - 147	
1,2,3,7,8,9-HxCDF	0.00000150				13C-1,2,3,4,6,7,8-HpCDF	82.7	28 - 143	
1,2,3,4,6,7,8-HpCDF	0.00000102				13C-1,2,3,4,7,8,9-HpCDF	80.2	26 - 138	
1,2,3,4,7,8,9-HpCDF	0.00000140				13C-OCDF	67.7	17 - 157	
OCDF	0.00000384				CRS 37Cl-2,3,7,8-TCDD	95.8	35 - 197	

Totals

Total TCDD	ND	0.00000121		
Total PeCDD	ND	0.00000123		
Total HxCDD	ND	0.00000234		
Total HpCDD	0.00000551			
Total TCDF	ND	0.00000130		
Total PeCDF	ND	0.00000179		
Total HxCDF	ND		0.00000252	
Total HpCDF	ND	0.00000118		

Footnotes
 a. Sample specific estimated detection limit.
 b. Estimated maximum possible concentration.
 c. Method detection limit.
 d. Lower control limit - upper control limit.

Analyst: WJL

Approved By: Martha M. Maier 17-May-2005 12:04

AMEC VALIDATED

Project 26089

LEVEL IV

CONTRACT COMPLIANCE SCREENING FORM FOR HARDCOPY DATA

AMEC Earth & Environmental
 550 South Wadsworth Boulevard
 Suite 500
 Lakewood, CO 80226


Package ID T711VO106
 Task Order 313150010
 SDG No. IOD1705

No. of Analyses 2

Laboratory Del Mar

Reviewer M. Pokorny

Analysis/Method Volatiles

Date: June 10, 2005
 Reviewer's Signature


ACTION ITEMS^a	
1. Case Narrative	
Deficiencies	
2. Out of Scope	
Analyses	
3. Analyses Not Conducted	
4. Missing Hardcopy	
Deliverables	
5. Incorrect Hardcopy	
Deliverables	
6. Deviations from Analysis	Qualification required calibration outlier.
Protocol, e.g.,	
Holding Times	
GC/MS Tune/Inst. Perform	
Calibrations	
Blanks	
Surrogates	
Matrix Spike/Dup LCS	
Field QC	
Internal Standard Performance	
Compound Identification and	
Quantitation	
System Performance	
COMMENTS^b	
^a Subcontracted analytical laboratory is not meeting contract and/or method requirements. ^b Differences in protocol have been adopted by the laboratory but no action against the laboratory is required.	



DATA VALIDATION REPORT

NPDES Monitoring

ANALYSIS: VOLATILES

SAMPLE DELIVERY GROUP: IOD1705

Prepared by

AMEC Denver Operations
550 South Wadsworth Boulevard, Suite 500
Lakewood, Colorado 80226

1. INTRODUCTION

Task Order Title: NPDES Monitoring
Contract Task Order #: 313150010
SDG#: IOD1705
Project Manager: B. McIlvaine
Matrix: Water
Analysis: Volatiles
QC Level: Level IV
No. of Samples: 2
No. of Reanalyses/Dilutions: 0
Reviewer: M. Pokorny
Date of Review: June 10, 2005

The samples listed in Table 1 were validated based on the guidelines outlined in the *AMEC Data Validation Procedure for Levels C and D Volatile Organics (DVP-2, Rev. 2)*, *EPA Method 624, SW846 Method 8260B*, and the *National Functional Guidelines For Organic Data Review (2/94)*. Any deviations from these procedures are documented herein. Qualifiers were applied in cases where the data did not meet the required QC criteria or where special consideration by the data user is required. Data qualifiers were placed on Form Is with the associated qualification codes. Analytes that were rejected for any reason are denoted on the summary forms as having only the "R" data qualifier and associated qualification code(s) denoting the reason for rejection. Any additional problems with the data that may have resulted in an estimated value were not denoted by a qualification code since the data had already been rejected.

Table 1. Sample identification

Client ID	EPA ID	Lab No.	Matrix	Method
Outfall 002	Outfall 002	IOD1705-01	water	624
Trip Blank	Trip Blank	IOD1705-02	water	624

2. DATA VALIDATION FINDINGS

2.1 SAMPLE MANAGEMENT

The following are findings associated with sample management:

2.1.1 Sample Preservation, Handling, and Transport

The samples in this SDG were received at the laboratory within the temperature limits of 4°C \pm 2°C. The samples were properly preserved. The COC noted that the samples were received intact; however, information regarding absence of headspace was not provided. No qualifications were required.

2.1.2 Chain of Custody

The COC was signed and dated by both field and laboratory personnel. The COC accounted for the analyses presented in this SDG. As the samples were couriered directly to the laboratory, custody seals were not required. No qualifications were required.

2.1.3 Holding Times

The samples were analyzed within 14 days of collection. No qualifications were required.

2.2 GC/MS TUNING

The ion abundance windows shown on the quantitation reports were consistent with those specified in EPA Method 624, and all ion abundances were within the established windows. The samples and associated QC were analyzed within 12 hours of the BFB injection time. The BFB summary report was verified from the raw data and no discrepancies between the summary report and the raw data were noted. No qualifications were required.

2.3 CALIBRATION

One initial calibration dated 04/18/05 was associated with this SDG. The average RRFs were \geq 0.05 for the target compounds listed on the sample result summaries. The %RSDs were \leq 35% for all applicable target compounds. One continuing calibration dated 05/03/05 was associated with the sample analyses in this SDG. The %Ds for all target compounds were \leq 20% in the continuing calibration except for the %Ds for trichlorofluoromethane and carbon tetrachloride. Trichlorofluoromethane and carbon tetrachloride were qualified as estimated nondetects, "UJ," in the site sample of this SDG. The trip blank required no qualification. The RRFs were \geq 0.05 for the target compounds listed on the sample result summaries. A representative number of %RSDs and average RRFs from the initial calibration, and %Ds and RRFs from the continuing calibration were recalculated from the raw data, and no calculation or transcription errors were found. No further qualifications were required.

2.4 BLANKS

One water method blank (5E03004-BLK1) was associated with the sample analyses. There were no detects above the MDLs for the target compounds listed on the sample result summaries. The method blank raw data showed no evidence of false negatives. No qualifications were required.

2.5 BLANK SPIKES AND LABORATORY CONTROL SAMPLES

One water blank spike (5E03004-BS1) was associated with the sample analyses. All recoveries were within the laboratory-established QC limits. A representative number of recoveries were recalculated from the raw data and no calculation or transcription errors were found. No qualifications were required.

2.6 SURROGATE RECOVERY

The surrogates were recovered within the QC limits of 80-120% in the samples and associated QC. A representative number of surrogate recoveries were recalculated from the raw data and no calculation or transcription errors were found. No qualifications were required.

2.7 MATRIX SPIKE/MATRIX SPIKE DUPLICATE

Sample Outfall 002 was the MS/MSD analyzed with this SDG. All recoveries and RPDs were within the laboratory-established QC limits, except for the recovery of trichlorofluoromethane above the QC limits in the MS only. A representative number of recoveries and RPDs were recalculated from the raw data and no calculation or transcription errors were found. No qualifications were required.

2.8 FIELD QC SAMPLES

Field QC samples were evaluated, and if necessary, qualified based on method blanks and other laboratory QC results affecting the usability of the field QC data. Any remaining detects were used to evaluate the associated site sample. Following are findings associated with field QC samples:

2.8.1 Trip Blanks

Sample Trip Blank was the trip blank associated with this SDG. There were no target compounds detected above the MDLs in the trip blank. No qualifications were required.

2.8.2 Field Blanks and Equipment Rinsates

There were no field QC samples associated with this SDG. No qualifications were required.

2.8.3 Field Duplicates

There were no field duplicate samples associated with this SDG.

2.9 INTERNAL STANDARDS PERFORMANCE

Internal standard area counts and retention times for the samples in this SDG were within the control limits established by the continuing calibration standards: +100%/-50% for internal standard areas and ± 0.50 minutes for retention times. A representative number of internal standard areas and retention times were verified from the raw data, and no calculation or transcription errors were noted. No qualifications were required.

2.10 COMPOUND IDENTIFICATION

Target compound identification was verified at a Level IV data validation. The laboratory analyzed for 15 volatile target compounds by EPA Method 624. Chromatograms, retention times, and spectra for the samples and QC were examined and no target compound identification problems were noted. No qualifications were required.

2.11 COMPOUND QUANTIFICATION AND REPORTED DETECTION LIMITS

Compound quantification is verified at a Level IV data validation. The reporting limits were supported by the lowest concentrations of the initial calibration standard and by the MDL study. As there were no sample detects in this SDG, compound quantitation was verified by recalculating a representative number of blank spike and surrogate recoveries from the raw data. Results were reported in $\mu\text{g/L}$ (ppb). No calculation or transcription errors were noted. No qualifications were required.

2.12 TENTATIVELY IDENTIFIED COMPOUNDS

The laboratory did not provide TICs for this SDG. No qualifications were required.

2.13 SYSTEM PERFORMANCE

A review of the chromatograms and other raw data showed no identifiable problems with system performance. No qualifications were required.

Data Qualifier Reference Table

Qualifier	Organics	Inorganics
U	The analyte was analyzed for, but was not detected above the reported sample quantitation limit.	The material was analyzed for, but was not detected above the level of the associated value. The associated value is either the sample quantitation limit or the sample detection limit.
J	The analyte was positively identified; the associated numerical value is the approximate concentration of the analyte in the sample.	The associated value is an estimated quantity.
N	The analysis indicates the presence of an analyte for which there is presumptive evidence to make a "tentative identification."	Not applicable.
NJ	The analysis indicates the presence of an analyte that has been "tentatively identified" and the associated numerical value represents its approximate concentration.	Not applicable.
UJ	The analyte was not detected above the reported sample quantitation limit. However, the reported quantitation limit is approximate and may or may not represent the actual limit of quantitation necessary to accurately and precisely measure the analyte in the sample.	The material was analyzed for, but was not detected. The associated value is an estimate and may be inaccurate or imprecise.
R	The sample results are rejected due to serious deficiencies in the ability to analyze the sample and to meet quality control criteria. The presence or absence of the analyte cannot be verified.	The data are unusable. (Note: Analyte may or may not be present).

Qualification Code Reference Table

Qualifier	Organics	Inorganics
H	Holding times were exceeded.	Holding times were exceeded.
S	Surrogate recovery was outside QC limits.	The sequence or number of standards used for the calibration was incorrect
C	Calibration %RSD or %D were noncompliant.	Correlation coefficient is <0.995.
R	Calibration RRF was <0.05.	%R for calibration is not within control limits.
B	Presumed contamination from preparation (method) blank.	Presumed contamination from preparation (method) or calibration blank.
L	Laboratory Blank Spike/Blank Spike Duplicate %R was not within control limits.	Laboratory Control Sample %R was not within control limits.
Q	MS/MSD recovery was poor or RPD high.	MS recovery was poor.
E	Not applicable.	Duplicates showed poor agreement.
I	Internal standard performance was unsatisfactory.	ICP ICS results were unsatisfactory.
A	Not applicable.	ICP Serial Dilution %D were not within control limits.
M	Tuning (BFB or DFTPP) was noncompliant.	Not applicable.
T	Presumed contamination from trip blank.	Not applicable.
+	False positive – reported compound was not present. Not applicable.	Not applicable.
-	False negative – compound was present but not reported.	Not applicable.
F	Presumed contamination from FB, or ER.	Presumed contamination from FB or ER.
S	Reported result or other information was incorrect.	Reported result or other information was incorrect.
?	TIC identity or reported retention time has been changed.	Not applicable.
D	The analysis with this flag should not be used because another more technically sound analysis is available.	The analysis with this flag should not be used because another more technically sound analysis is available.
P	Instrument performance for pesticides was poor.	Post Digestion Spike recovery was not within control limits.
DNQ	The compound was detected between the MDL and the RL and, by definition, is considered an estimated value.	The compound was detected between the MDL and the RL and, by definition, is considered an estimated value.
#	Unusual problems found with the data that have been described in Section 2.#, "Data Validation Findings." The number following the asterisk () will indicate the subsection where a description of the problem can be found (eg. *1 would indicate a sample was not within temperature limits).	Unusual problems found with the data that have been described in Section 2.#, "Data Validation Findings." The number following the asterisk (*) will indicate the subsection where a description of the problem can be found (eg. *1 would indicate a sample was not within temperature limits).



DATA VALIDATION REPORT

NPDES Monitoring

ANALYSIS: GENERAL MINERALS

SAMPLE DELIVERY GROUP: IOD1705

Prepared by

AMEC—Denver Operations
550 South Wadsworth Boulevard, Suite 500
Lakewood, Colorado 80226

I. INTRODUCTION

Task Order Title: NPDES Monitoring
Contract Task Order #: 313150010
Sample Delivery Group #: IOD1705
Project Manager: B. McIlvaine
Matrix: Water
Analysis: General Minerals
QC Level: Level IV
No. of Samples: 1
Reviewer: L. Jarusewic
Date of Review: June 2, 2005

The sample listed in Table 1 was validated based on the guidelines outlined in the AMEC *Data Validation Procedures SOP DVP-6, Rev. 2, USEPA Methods for Chemical Analysis of Water and Wastes Method 300.0, 350.2, 120.1, and 180.1, Standard Methods for the Examination of Water and Wastewater Method SM2540C*, and validation guidelines outlined in the *USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (2/94)*. Any deviations from these procedures and guidelines are documented herein. Qualifiers were applied in cases where the data did not meet the required QC criteria or where special consideration by the data user is required. Data qualifiers were placed on Form Is with the associated qualification codes. Analytes that were rejected for any reason are denoted on the Form I as having only the "R" data qualifier and associated qualification code(s) denoting the reason for rejection. Any additional problems with the data that may have resulted in an estimated value were not denoted by a qualification code since the data had already been rejected.

Table 1. Sample identification

Client ID	EPA ID	Laboratory ID	Matrix	COC Method
Outfall 002	Outfall 002	IOD1705-01	Water	General Minerals
Outfall 002RE1	Outfall 002RE1	IOD1705-01RE1	Water	EPA 300.0 and SM2540C

2. DATA VALIDATION FINDINGS

2.1 SAMPLE MANAGEMENT

Following are findings associated with sample management:

2.1.1 Sample Preservation, Handling, and Transport

The sample in this SDG was received at the laboratory within the temperature limits of $4^{\circ}\text{C} \pm 2^{\circ}\text{C}$. No preservation problems were noted by the laboratory. No qualifications were required.

2.1.2 Chain of Custody

The COC was signed and dated by field and laboratory personnel. The COC accounted for all analyses presented in this SDG. The laboratory did not append the sample ID with the RE1 suffix for the sulfate and total dissolved solids reanalyses. The reviewer edited the Form I to reflect this information. No sample qualifications were required.

2.1.3 Holding Times

The holding times were assessed by comparing the date of collection with the dates of analyses. The 28-day analytical holding time for ammonia, sulfate, and conductivity, the seven day holding time for total dissolved solids, and the 48-hour holding time for turbidity were met. No qualifications were required.

2.2 CALIBRATION

For the applicable analyses, the initial calibration correlation coefficients were ≥ 0.995 . Initial and continuing calibration information was acceptable with recoveries within the control limits of 90-110%. For ammonia, no information regarding the standardization of the titrant was provided; however, the LCS recovery was within the CCV control limits. Calibration is not applicable to the total dissolved solids analysis. No qualifications were required.

2.3 BLANKS

The method blank and CCB results reported on the summary forms and in the raw data for blank analyses associated with the sample were nondetects at the reporting limit. No qualifications were required.

2.4 BLANK SPIKES AND LABORATORY CONTROL SAMPLES

The laboratory control sample recoveries were within the laboratory-established control limits. The LCS is not applicable to turbidity or conductivity. No qualifications were required.

2.5 SURROGATES RECOVERY

Surrogate recovery is not applicable to the analyses presented in this SDG.

2.6 LABORATORY DUPLICATES

Laboratory duplicate analyses were performed on sample Outfall 002 for total dissolved solids and conductivity. The RPDs were within the laboratory-established control limits and no qualifications were required.

2.7 MATRIX SPIKE/MATRIX SPIKE DUPLICATE

There were no MS/MSD analyses performed in association with the sample in this SDG; therefore, no assessment was made with respect to this criterion. Method accuracy was based on LCS results. No qualifications were required.

2.8 FURNACE ATOMIC ABSORPTION QC

Furnace atomic absorption was not utilized for the analyses of this sample; therefore, furnace atomic absorption QC is not applicable.

2.9 ICP SERIAL DILUTION

ICP serial dilution is not applicable to the analyses presented in this data validation report.

2.10 SAMPLE RESULT VERIFICATION

A Level IV review was performed for the sample in this data package. Calculations were verified, and the sample results reported on the Form I were verified against the raw data. No transcription errors or calculation errors were noted. Outfall 002 was reanalyzed for sulfate and total dissolved solids. As the Outfall 002RE1 sulfate and total dissolved solids results were similar to the Outfall 002 results, the Outfall 002RE1 sulfate and total dissolved solids results were rejected, "R," in favor of the original analyses. No further qualifications were required.

2.11 FIELD QC SAMPLES

Field QC samples are evaluated, and if necessary, qualified based only on laboratory blanks. Any remaining detects are used to evaluate the associated sample. The following are findings associated with field QC samples:

2.11.1 Field Blanks and Equipment Rinsates

The sample in this SDG had no associated field QC samples. No qualifications were required.

2.11.2 Field Duplicates

There were no field duplicate pairs associated with this SDG.



Del Mar Analytical

17471 Colton Ave., Suite 100, Irvine, CA 92614 (949) 261-1021 FAX (949) 261-1027
 1014 E. Conley Dr., Suite A, Colton, CA 92324 (909) 376-4667 FAX (909) 370-1046
 9484 Chesapeake Dr., Suite 805, San Diego, CA 92123 (619) 507-8596 FAX (619) 505-9689
 9830 South 51st St., Suite B-120, Phoenix, AZ 85044 (480) 785-0043 FAX (480) 785-0851
 2520 E. Sunset Rd. #3, Las Vegas, NV 89120 (702) 798-3620 FAX (702) 798-3621

MWH-Pasadena/Boeing
 300 North Lake Avenue, Suite 1200
 Pasadena, CA 91101
 Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD1705

Sampled: 04/22/05
 Received: 04/22/05

DRAFT: INORGANICS

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data	Qualifiers
Sample ID: IOD1705-01 (DRAFT: Outfall 002 - Water)										
Reporting Units: mg/l										
Ammonia-N (Distilled)	EPA 350.2	5D26085	0.30	0.50	ND	1	04/26/05	04/26/05	U	
Sulfate	EPA 300.0	5D22051	1.8	5.0	400	10	04/22/05	04/22/05		
Total Dissolved Solids	SM2540C	5D25087	10	10	1000	1	04/25/05	04/25/05		
Sample ID: IOD1705-01RE1 (DRAFT: Outfall 002 RE1 - Water)										
Reporting Units: mg/l										
Sulfate	EPA 300.0	5D25050	1.8	5.0	420	10	04/25/05	04/26/05	R	D
Total Dissolved Solids	SM2540C	5D26122	10	10	1000	1	04/25/05	04/26/05	↓	↓
Sample ID: IOD1705-01 (DRAFT: Outfall 002 - Water)										
Reporting Units: NTU										
Turbidity	EPA 180.1	5D23050	0.040	1.0	6.4	1	04/23/05	04/23/05		
Sample ID: IOD1705-01 (DRAFT: Outfall 002 - Water)										
Reporting Units: umhos/cm										
Specific Conductance	EPA 120.1	5D25093	1.0	1.0	1400	1	04/25/05	04/25/05		


J 6/2/05

AMEC VALIDATED

LEVEL IV

DRAFT REPORT
 DRAFT REPORT
 DATA SUBJECT TO CHANGE

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from Del Mar Analytical.





LABORATORY REPORT

Prepared For: MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project: Routine Outfall 002

Sampled: 04/28/05
Received: 04/28/05
Issued: 06/20/05 16:55

NELAP #01108CA California ELAP#1197 CSDLAC #10117

The results listed within this Laboratory Report pertain only to the samples tested in the laboratory. The analyses contained in this report were performed in accordance with the applicable certifications as noted. All soil samples are reported on a wet weight basis unless otherwise noted in the report. This Laboratory Report is confidential and is intended for the sole use of Del Mar Analytical and its client. This report shall not be reproduced, except in full, without written permission from Del Mar Analytical. The Chain of Custody, 1 page, is included and is an integral part of this report.

This entire report was reviewed and approved for release.

SAMPLE CROSS REFERENCE

SUBCONTRACTED: Refer to the last page for specific subcontract laboratory information included in this report.

LABORATORY ID	CLIENT ID	MATRIX
IOD2044-01	Outfall 002	Water
IOD2044-02	Trip Blank	Water

Reviewed By:

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002
Report Number: IOD2044

Sampled: 04/28/05
Received: 04/28/05

PURGEABLES BY GC/MS (EPA 624)

Table with columns: Analyte, Method, Batch, MDL Limit, Reporting Limit, Sample Result, Dilution Factor, Date Extracted, Date Analyzed, Data Qualifiers. Includes sections for Sample ID: IOD2044-01 (Outfall 002 - Water) and Sample ID: IOD2044-02 (Trip Blank - Water).

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD2044

Sampled: 04/28/05

Received: 04/28/05

ACID & BASE/NEUTRALS BY GC/MS (EPA 625)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD2044-01 (Outfall 002 - Water)									
Reporting Units: ug/l									
Bis(2-ethylhexyl)phthalate	EPA 625	5E01024	1.1	5.0	ND	0.98	05/01/05	05/04/05	
2,4-Dinitrotoluene	EPA 625	5E01024	0.23	9.0	ND	0.98	05/01/05	05/04/05	
N-Nitrosodimethylamine	EPA 625	5E01024	0.22	8.0	ND	0.98	05/01/05	05/04/05	
Pentachlorophenol	EPA 625	5E01024	0.78	8.0	ND	0.98	05/01/05	05/04/05	
2,4,6-Trichlorophenol	EPA 625	5E01024	0.10	6.0	ND	0.98	05/01/05	05/04/05	
Surrogate: 2-Fluorophenol (30-120%)					59 %				
Surrogate: Phenol-d6 (35-120%)					60 %				
Surrogate: 2,4,6-Tribromophenol (45-120%)					74 %				
Surrogate: Nitrobenzene-d5 (45-120%)					63 %				
Surrogate: 2-Fluorobiphenyl (45-120%)					66 %				
Surrogate: Terphenyl-d14 (45-120%)					70 %				

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD2044

Sampled: 04/28/05

Received: 04/28/05

ORGANOCHLORINE PESTICIDES (EPA 608)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD2044-01 (Outfall 002 - Water) - cont.									
Reporting Units: ug/l									
alpha-BHC	EPA 608	5E03078	0.0010	0.010	ND	0.971	05/03/05	05/04/05	
<i>Surrogate: Decachlorobiphenyl (45-120%)</i>					75 %				
<i>Surrogate: Tetrachloro-m-xylene (35-115%)</i>					64 %				

Del Mar Analytical, Irvine
Michele Harper
Project Manager



Del Mar Analytical

17461 Derian Ave., Suite 100, Irvine, CA 92614 (949) 261-1022 FAX (949) 260-3297
 1014 E. Cooley Dr., Suite A, Colton, CA 92324 (909) 370-4667 FAX (909) 370-1046
 9484 Chesapeake Dr., Suite 805, San Diego, CA 92123 (858) 503-8596 FAX (858) 505-9689
 9830 South 51st St., Suite B-120, Phoenix, AZ 85044 (480) 785-0043 FAX (480) 785-0851
 2520 E. Sunset Rd. #3, Las Vegas, NV 89120 (702) 798-3620 FAX (702) 798-3621

MWH-Pasadena/Boeing
 300 North Lake Avenue, Suite 1200
 Pasadena, CA 91101
 Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD2044

Sampled: 04/28/05

Received: 04/28/05

METALS

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD2044-01 (Outfall 002 - Water) - cont.									
Reporting Units: ug/l									
Copper	EPA 200.8	5D28117	0.49	2.0	4.6	1	04/28/05	04/29/05	
Lead	EPA 200.8	5D28117	0.13	1.0	2.4	1	04/28/05	04/29/05	
Mercury	EPA 245.1	5D29061	0.063	0.20	ND	1	04/29/05	04/29/05	

Del Mar Analytical, Irvine
 Michele Harper
 Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from Del Mar Analytical.



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD2044

Sampled: 04/28/05

Received: 04/28/05

INORGANICS

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD2044-01 (Outfall 002 - Water) - cont.									
Reporting Units: mg/l									
Ammonia-N (Distilled)	EPA 350.2	5E02067	0.30	0.50	0.84	1	05/02/05	05/02/05	
Biochemical Oxygen Demand	EPA 405.1	5D29091	0.59	2.0	5.2	1	04/29/05	05/04/05	
Chloride	EPA 300.0	5D28116	0.52	1.0	31	2	04/28/05	04/29/05	
Nitrate/Nitrite-N	EPA 300.0	5D28116	0.072	0.26	0.23	1	04/28/05	04/28/05	J
Oil & Grease	EPA 413.1	5D29041	0.94	5.0	ND	1	04/29/05	04/29/05	
Sulfate	EPA 300.0	5D28116	0.36	1.0	89	2	04/28/05	04/29/05	
Surfactants (MBAS)	SM5540-C	5D28122	0.044	0.10	ND	1	04/28/05	04/28/05	
Total Dissolved Solids	SM2540C	5D29129	10	10	530	1	04/29/05	04/29/05	
Total Suspended Solids	EPA 160.2	5E04071	10	10	91	1	05/04/05	05/04/05	
Sample ID: IOD2044-01 (Outfall 002 - Water)									
Reporting Units: ml/hr									
Total Settleable Solids	EPA 160.5	5D29094	0.10	0.10	ND	1	04/29/05	04/29/05	
Sample ID: IOD2044-01 (Outfall 002 - Water)									
Reporting Units: NTU									
Turbidity	EPA 180.1	5D29110	0.080	2.0	79	2	04/29/05	04/29/05	
Sample ID: IOD2044-01 (Outfall 002 - Water)									
Reporting Units: ug/l									
Total Cyanide	EPA 335.2	5D29078	2.2	5.0	ND	1	04/29/05	04/29/05	
Perchlorate	EPA 314.0	5D29065	0.80	4.0	ND	1	04/29/05	04/30/05	
Sample ID: IOD2044-01 (Outfall 002 - Water)									
Reporting Units: umhos/cm									
Specific Conductance	EPA 120.1	5D29130	1.0	1.0	590	1	04/29/05	04/29/05	

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD2044

Sampled: 04/28/05

Received: 04/28/05

SHORT HOLD TIME DETAIL REPORT

	Hold Time (in days)	Date/Time Sampled	Date/Time Received	Date/Time Extracted	Date/Time Analyzed
Sample ID: Outfall 002 (IOD2044-01) - Water					
EPA 160.5	2	04/28/2005 14:06	04/28/2005 18:15	04/29/2005 13:29	04/29/2005 15:00
EPA 180.1	2	04/28/2005 14:06	04/28/2005 18:15	04/29/2005 15:00	04/29/2005 16:00
EPA 300.0	2	04/28/2005 14:06	04/28/2005 18:15	04/28/2005 21:10	04/28/2005 21:58
EPA 405.1	2	04/28/2005 14:06	04/28/2005 18:15	04/29/2005 13:10	05/04/2005 10:00
SM5540-C	2	04/28/2005 14:06	04/28/2005 18:15	04/28/2005 21:00	04/28/2005 21:40

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD2044

Sampled: 04/28/05

Received: 04/28/05

METHOD BLANK/QC DATA

PURGEABLES BY GC/MS (EPA 624)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC %REC	RPD Limits	RPD RPD	Data Qualifiers
---------	--------	-----------------	-----	-------	-------------	---------------	-----------	------------	---------	-----------------

Batch: 5E04019 Extracted: 05/04/05

Blank Analyzed: 05/04/2005 (5E04019-BLK1)

Benzene	ND	2.0	0.28	ug/l						
Trichlorotrifluoroethane (Freon 113)	ND	5.0	1.2	ug/l						
Carbon tetrachloride	ND	5.0	0.28	ug/l						
Chloroform	ND	2.0	0.33	ug/l						
1,1-Dichloroethane	ND	2.0	0.27	ug/l						
1,2-Dichloroethane	ND	2.0	0.28	ug/l						
1,1-Dichloroethene	ND	3.0	0.32	ug/l						
Ethylbenzene	ND	2.0	0.25	ug/l						
Tetrachloroethene	ND	2.0	0.32	ug/l						
Toluene	ND	2.0	0.36	ug/l						
1,1,1-Trichloroethane	ND	2.0	0.30	ug/l						
1,1,2-Trichloroethane	ND	2.0	0.30	ug/l						
Trichloroethene	ND	5.0	0.26	ug/l						
Trichlorofluoromethane	ND	5.0	0.34	ug/l						
Vinyl chloride	ND	5.0	0.26	ug/l						
Xylenes, Total	ND	4.0	0.52	ug/l						
Surrogate: Dibromofluoromethane	28.0			ug/l	25.0		112	80-120		
Surrogate: Toluene-d8	27.0			ug/l	25.0		108	80-120		
Surrogate: 4-Bromofluorobenzene	24.4			ug/l	25.0		98	80-120		

LCS Analyzed: 05/04/2005 (5E04019-BS1)

Benzene	27.2	2.0	0.28	ug/l	25.0		109	70-120		
Carbon tetrachloride	24.4	5.0	0.28	ug/l	25.0		98	70-140		
Chloroform	26.2	2.0	0.33	ug/l	25.0		105	75-130		
1,1-Dichloroethane	28.4	2.0	0.27	ug/l	25.0		114	70-135		
1,2-Dichloroethane	24.1	2.0	0.28	ug/l	25.0		96	60-150		
1,1-Dichloroethene	29.8	3.0	0.32	ug/l	25.0		119	75-135		
Ethylbenzene	26.7	2.0	0.25	ug/l	25.0		107	80-120		
Tetrachloroethene	25.6	2.0	0.32	ug/l	25.0		102	75-125		
Toluene	26.5	2.0	0.36	ug/l	25.0		106	75-120		
1,1,1-Trichloroethane	25.7	2.0	0.30	ug/l	25.0		103	75-140		
1,1,2-Trichloroethane	28.0	2.0	0.30	ug/l	25.0		112	70-125		
Trichloroethene	24.0	5.0	0.26	ug/l	25.0		96	80-120		
Trichlorofluoromethane	25.2	5.0	0.34	ug/l	25.0		101	65-145		
Vinyl chloride	24.5	5.0	0.26	ug/l	25.0		98	50-130		
Surrogate: Dibromofluoromethane	28.2			ug/l	25.0		113	80-120		

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD2044

Sampled: 04/28/05
Received: 04/28/05

METHOD BLANK/QC DATA

PURGEABLES BY GC/MS (EPA 624)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
---------	--------	-----------------	-----	-------	-------------	---------------	------	-------------	-----	-----------	-----------------

Batch: 5E04019 Extracted: 05/04/05

LCS Analyzed: 05/04/2005 (5E04019-BS1)

Surrogate: Toluene-d8	27.8			ug/l	25.0		111	80-120			
Surrogate: 4-Bromofluorobenzene	27.1			ug/l	25.0		108	80-120			

Matrix Spike Analyzed: 05/05/2005 (5E04019-MS1)

Source: IOD2043-01

Benzene	27.6	2.0	0.28	ug/l	25.0	ND	110	70-120			
Carbon tetrachloride	23.1	5.0	0.28	ug/l	25.0	ND	92	70-145			
Chloroform	25.1	2.0	0.33	ug/l	25.0	ND	100	70-135			
1,1-Dichloroethane	27.2	2.0	0.27	ug/l	25.0	ND	109	65-135			
1,2-Dichloroethane	21.6	2.0	0.28	ug/l	25.0	ND	86	60-150			
1,1-Dichloroethene	29.5	3.0	0.32	ug/l	25.0	ND	118	65-140			
Ethylbenzene	27.8	2.0	0.25	ug/l	25.0	ND	111	70-130			
Tetrachloroethene	26.5	2.0	0.32	ug/l	25.0	ND	106	70-130			
Toluene	25.1	2.0	0.36	ug/l	25.0	ND	100	70-120			
1,1,1-Trichloroethane	24.0	2.0	0.30	ug/l	25.0	ND	96	75-140			
1,1,2-Trichloroethane	25.8	2.0	0.30	ug/l	25.0	ND	103	60-135			
Trichloroethene	24.8	5.0	0.26	ug/l	25.0	ND	99	70-125			
Trichlorofluoromethane	23.2	5.0	0.34	ug/l	25.0	ND	93	55-145			
Vinyl chloride	23.2	5.0	0.26	ug/l	25.0	ND	93	40-135			
Surrogate: Dibromofluoromethane	25.9			ug/l	25.0		104	80-120			
Surrogate: Toluene-d8	26.8			ug/l	25.0		107	80-120			
Surrogate: 4-Bromofluorobenzene	27.0			ug/l	25.0		108	80-120			

Matrix Spike Dup Analyzed: 05/05/2005 (5E04019-MSD1)

Source: IOD2043-01

Benzene	27.9	2.0	0.28	ug/l	25.0	ND	112	70-120	1	20	
Carbon tetrachloride	23.5	5.0	0.28	ug/l	25.0	ND	94	70-145	2	25	
Chloroform	24.9	2.0	0.33	ug/l	25.0	ND	100	70-135	1	20	
1,1-Dichloroethane	26.7	2.0	0.27	ug/l	25.0	ND	107	65-135	2	20	
1,2-Dichloroethane	20.4	2.0	0.28	ug/l	25.0	ND	82	60-150	6	20	
1,1-Dichloroethene	29.3	3.0	0.32	ug/l	25.0	ND	117	65-140	1	20	
Ethylbenzene	28.0	2.0	0.25	ug/l	25.0	ND	112	70-130	1	20	
Tetrachloroethene	26.6	2.0	0.32	ug/l	25.0	ND	106	70-130	0	20	
Toluene	26.6	2.0	0.36	ug/l	25.0	ND	106	70-120	6	20	
1,1,1-Trichloroethane	24.1	2.0	0.30	ug/l	25.0	ND	96	75-140	0	20	
1,1,2-Trichloroethane	24.6	2.0	0.30	ug/l	25.0	ND	98	60-135	5	25	
Trichloroethene	25.0	5.0	0.26	ug/l	25.0	ND	100	70-125	1	20	
Trichlorofluoromethane	22.7	5.0	0.34	ug/l	25.0	ND	91	55-145	2	25	

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing Project ID: Routine Outfall 002
300 North Lake Avenue, Suite 1200 Report Number: IOD2044
Pasadena, CA 91101 Sampled: 04/28/05
Attention: Bronwyn Kelly Received: 04/28/05

METHOD BLANK/QC DATA

PURGEABLES BY GC/MS (EPA 624)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5E04019 Extracted: 05/04/05											
Matrix Spike Dup Analyzed: 05/05/2005 (5E04019-MSD1)						Source: IOD2043-01					
Vinyl chloride	23.3	5.0	0.26	ug/l	25.0	ND	93	40-135	0	30	
Surrogate: Dibromofluoromethane	25.0			ug/l	25.0		100	80-120			
Surrogate: Toluene-d8	27.0			ug/l	25.0		108	80-120			
Surrogate: 4-Bromofluorobenzene	25.3			ug/l	25.0		101	80-120			

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002
Report Number: IOD2044

Sampled: 04/28/05
Received: 04/28/05

METHOD BLANK/QC DATA

ACID & BASE/NEUTRALS BY GC/MS (EPA 625)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC %REC	Limit	RPD	RPD Limit	Data Qualifiers
---------	--------	-----------------	-----	-------	-------------	---------------	-----------	-------	-----	-----------	-----------------

Batch: 5E01024 Extracted: 05/01/05

Blank Analyzed: 05/04/2005 (5E01024-BLK1)

Bis(2-ethylhexyl)phthalate	ND	5.0	1.1	ug/l							
2,4-Dinitrotoluene	ND	9.0	0.23	ug/l							
N-Nitrosodimethylamine	ND	8.0	0.22	ug/l							
Pentachlorophenol	ND	8.0	0.78	ug/l							
2,4,6-Trichlorophenol	ND	6.0	0.10	ug/l							
Surrogate: 2-Fluorophenol	12.2			ug/l	20.0		61	30-120			
Surrogate: Phenol-d6	12.4			ug/l	20.0		62	35-120			
Surrogate: 2,4,6-Tribromophenol	14.9			ug/l	20.0		74	45-120			
Surrogate: Nitrobenzene-d5	6.02			ug/l	10.0		60	45-120			
Surrogate: 2-Fluorobiphenyl	6.54			ug/l	10.0		65	45-120			
Surrogate: Terphenyl-d14	7.54			ug/l	10.0		75	45-120			

LCS Analyzed: 05/04/2005 (5E01024-BS1)

Bis(2-ethylhexyl)phthalate	7.94	5.0	1.1	ug/l	10.0		79	60-130			M-NRI
2,4-Dinitrotoluene	6.92	9.0	0.23	ug/l	10.0		69	60-120			J
N-Nitrosodimethylamine	6.10	8.0	0.22	ug/l	10.0		61	40-120			J
Pentachlorophenol	7.58	8.0	0.78	ug/l	10.0		76	50-120			J
2,4,6-Trichlorophenol	7.66	6.0	0.10	ug/l	10.0		77	60-120			
Surrogate: 2-Fluorophenol	11.4			ug/l	20.0		57	30-120			
Surrogate: Phenol-d6	12.1			ug/l	20.0		60	35-120			
Surrogate: 2,4,6-Tribromophenol	15.6			ug/l	20.0		78	45-120			
Surrogate: Nitrobenzene-d5	6.30			ug/l	10.0		63	45-120			
Surrogate: 2-Fluorobiphenyl	7.26			ug/l	10.0		73	45-120			
Surrogate: Terphenyl-d14	7.76			ug/l	10.0		78	45-120			

LCS Dup Analyzed: 05/04/2005 (5E01024-BS1)

Bis(2-ethylhexyl)phthalate	8.48	5.0	1.1	ug/l	10.0		85	60-130	7	20	
2,4-Dinitrotoluene	7.22	9.0	0.23	ug/l	10.0		72	60-120	4	20	J
N-Nitrosodimethylamine	6.54	8.0	0.22	ug/l	10.0		65	40-120	7	20	J
Pentachlorophenol	8.02	8.0	0.78	ug/l	10.0		80	50-120	6	25	
2,4,6-Trichlorophenol	8.36	6.0	0.10	ug/l	10.0		84	60-120	9	20	
Surrogate: 2-Fluorophenol	12.9			ug/l	20.0		64	30-120			
Surrogate: Phenol-d6	13.5			ug/l	20.0		68	35-120			
Surrogate: 2,4,6-Tribromophenol	16.4			ug/l	20.0		82	45-120			
Surrogate: Nitrobenzene-d5	6.78			ug/l	10.0		68	45-120			
Surrogate: 2-Fluorobiphenyl	7.78			ug/l	10.0		78	45-120			

Del Mar Analytical, Irvine
Michele Harper
Project Manager

MWH-Pasadena/Boeing
 300 North Lake Avenue, Suite 1200
 Pasadena, CA 91101
 Attention: Bronwyn Kelly

Project ID: Routine Outfall 002
 Report Number: IOD2044

Sampled: 04/28/05
 Received: 04/28/05

METHOD BLANK/QC DATA

ACID & BASE/NEUTRALS BY GC/MS (EPA 625)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5E01024 Extracted: 05/01/05											
LCS Dup Analyzed: 05/04/2005 (5E01024-BSD1)											
Surrogate: Terphenyl-d14	8.06			ug/l	10.0		81	45-120			

Del Mar Analytical, Irvine
 Michele Harper
 Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002
Report Number: IOD2044

Sampled: 04/28/05
Received: 04/28/05

METHOD BLANK/QC DATA

ORGANOCHLORINE PESTICIDES (EPA 608)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5E03078 Extracted: 05/03/05											
Blank Analyzed: 05/04/2005 (5E03078-BLK1)											
alpha-BHC	ND	0.010	0.0010	ug/l							
Surrogate: Decachlorobiphenyl	0.451			ug/l	0.500		90	45-120			
Surrogate: Tetrachloro-m-xylene	0.350			ug/l	0.500		70	35-115			
LCS Analyzed: 05/04/2005 (5E03078-BS1)											
alpha-BHC	0.336	0.010	0.0010	ug/l	0.500		67	45-115			M-NR1
Surrogate: Decachlorobiphenyl	0.425			ug/l	0.500		85	45-120			
Surrogate: Tetrachloro-m-xylene	0.322			ug/l	0.500		64	35-115			
LCS Dup Analyzed: 05/04/2005 (5E03078-BSD1)											
alpha-BHC	0.364	0.010	0.0010	ug/l	0.500		73	45-115	8	30	
Surrogate: Decachlorobiphenyl	0.415			ug/l	0.500		83	45-120			
Surrogate: Tetrachloro-m-xylene	0.340			ug/l	0.500		68	35-115			

Del Mar Analytical, Irvine
Michele Harper
Project Manager



Del Mar Analytical

17461 Derian Ave., Suite 100, Irvine, CA 92614 (949) 261-1022 FAX (949) 260-3297
 1014 E. Cooley Dr., Suite A, Colton, CA 92324 (909) 370-4667 FAX (909) 370-1046
 9484 Chesapeake Dr., Suite 805, San Diego, CA 92123 (858) 505-8596 FAX (858) 505-9689
 9830 South 51st St., Suite B-120, Phoenix, AZ 85044 (480) 785-0043 FAX (480) 785-0851
 2520 E. Sunset Rd. #3, Las Vegas, NV 89120 (702) 798-3620 FAX (702) 798-3621

MWH-Pasadena/Boeing
 300 North Lake Avenue, Suite 1200
 Pasadena, CA 91101
 Attention: Bronwyn Kelly

Project ID: Routine Outfall 002
 Report Number: IOD2044

Sampled: 04/28/05
 Received: 04/28/05

METHOD BLANK/QC DATA

METALS

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC %REC	Limit	RPD	RPD Limit	Data Qualifiers
---------	--------	-----------------	-----	-------	-------------	---------------	-----------	-------	-----	-----------	-----------------

Batch: 5D28117 Extracted: 04/28/05

Blank Analyzed: 04/29/2005 (5D28117-BLK1)

Copper	ND	2.0	0.49	ug/l							
Lead	ND	1.0	0.13	ug/l							

LCS Analyzed: 04/29/2005 (5D28117-BS1)

Copper	79.5	2.0	0.49	ug/l	80.0		99	85-115			
Lead	89.7	1.0	0.13	ug/l	80.0		112	85-115			

Matrix Spike Analyzed: 04/29/2005 (5D28117-MS1)

					Source: IOD2044-01						
Copper	85.0	2.0	0.49	ug/l	80.0	4.6	100	70-130			
Lead	91.0	1.0	0.13	ug/l	80.0	2.4	111	70-130			

Matrix Spike Dup Analyzed: 04/29/2005 (5D28117-MSD1)

					Source: IOD2044-01						
Copper	79.3	2.0	0.49	ug/l	80.0	4.6	93	70-130	7	20	
Lead	89.6	1.0	0.13	ug/l	80.0	2.4	109	70-130	2	20	

Batch: 5D29061 Extracted: 04/29/05

Blank Analyzed: 04/29/2005 (5D29061-BLK1)

Mercury	ND	0.20	0.063	ug/l							
---------	----	------	-------	------	--	--	--	--	--	--	--

LCS Analyzed: 04/29/2005 (5D29061-BS1)

Mercury	8.06	0.20	0.063	ug/l	8.00		101	85-115			
---------	------	------	-------	------	------	--	-----	--------	--	--	--

Matrix Spike Analyzed: 04/29/2005 (5D29061-MS1)

					Source: IOD2033-03						
Mercury	7.76	0.20	0.063	ug/l	8.00	ND	97	70-130			

Matrix Spike Dup Analyzed: 04/29/2005 (5D29061-MSD1)

					Source: IOD2033-03						
Mercury	7.82	0.20	0.063	ug/l	8.00	ND	98	70-130	1	20	

Del Mar Analytical, Irvine
 Michele Harper
 Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from Del Mar Analytical.



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD2044

Sampled: 04/28/05
Received: 04/28/05

METHOD BLANK/QC DATA

INORGANICS

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC %REC	RPD Limits	RPD RPD	Data Qualifiers
---------	--------	-----------------	-----	-------	-------------	---------------	-----------	------------	---------	-----------------

Batch: 5D28116 Extracted: 04/28/05

Blank Analyzed: 04/28/2005 (5D28116-BLK1)

Chloride	ND	0.50	0.26	mg/l						
Nitrate/Nitrite-N	ND	0.26	0.072	mg/l						
Sulfate	ND	0.50	0.18	mg/l						

LCS Analyzed: 04/28/2005 (5D28116-BS1)

Chloride	4.82	0.50	0.26	mg/l	5.00		96	90-110		M-3
Sulfate	9.63	0.50	0.18	mg/l	10.0		96	90-110		M-3

Batch: 5D28122 Extracted: 04/28/05

Blank Analyzed: 04/28/2005 (5D28122-BLK1)

Surfactants (MBAS)	ND	0.10	0.044	mg/l						
--------------------	----	------	-------	------	--	--	--	--	--	--

LCS Analyzed: 04/28/2005 (5D28122-BS1)

Surfactants (MBAS)	0.252	0.10	0.044	mg/l	0.250		101	90-110		
--------------------	-------	------	-------	------	-------	--	-----	--------	--	--

Matrix Spike Analyzed: 04/28/2005 (5D28122-MS1)

Surfactants (MBAS)	0.276	0.10	0.044	mg/l	0.250	ND	110	50-125		
--------------------	-------	------	-------	------	-------	----	-----	--------	--	--

Source: IOD1996-01

Matrix Spike Dup Analyzed: 04/28/2005 (5D28122-MSD1)

Surfactants (MBAS)	0.277	0.10	0.044	mg/l	0.250	ND	111	50-125	0	20
--------------------	-------	------	-------	------	-------	----	-----	--------	---	----

Source: IOD1996-01

Batch: 5D29041 Extracted: 04/29/05

Blank Analyzed: 04/29/2005 (5D29041-BLK1)

Oil & Grease	ND	5.0	0.94	mg/l						
--------------	----	-----	------	------	--	--	--	--	--	--

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing Project ID: Routine Outfall 002
300 North Lake Avenue, Suite 1200 Report Number: IOD2044
Pasadena, CA 91101
Attention: Bronwyn Kelly
Sampled: 04/28/05
Received: 04/28/05

METHOD BLANK/QC DATA

INORGANICS

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC %REC	Limit	RPD	RPD Limit	Data Qualifiers
Batch: 5D29041 Extracted: 04/29/05											
LCS Analyzed: 04/29/2005 (5D29041-BS1)											
Oil & Grease	18.3	5.0	0.94	mg/l	20.0		92	65-120			M-NRI
LCS Dup Analyzed: 04/29/2005 (5D29041-BSD1)											
Oil & Grease	18.9	5.0	0.94	mg/l	20.0		94	65-120	3	20	
Batch: 5D29065 Extracted: 04/29/05											
Blank Analyzed: 04/29/2005 (5D29065-BLK1)											
Perchlorate	ND	4.0	0.80	ug/l							
LCS Analyzed: 04/29/2005 (5D29065-BS1)											
Perchlorate	51.0	4.0	0.80	ug/l	50.0		102	85-115			
Matrix Spike Analyzed: 04/29/2005 (5D29065-MS1)											
Perchlorate	53.1	4.0	0.80	ug/l	50.0	Source: IOD1955-04 ND	106	80-120			
Matrix Spike Dup Analyzed: 04/29/2005 (5D29065-MSD1)											
Perchlorate	52.9	4.0	0.80	ug/l	50.0	Source: IOD1955-04 ND	106	80-120	0	20	
Batch: 5D29078 Extracted: 04/29/05											
Blank Analyzed: 04/29/2005 (5D29078-BLK1)											
Total Cyanide	ND	5.0	2.2	ug/l							
LCS Analyzed: 04/29/2005 (5D29078-BS1)											
Total Cyanide	181	5.0	2.2	ug/l	200		90	90-110			

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
 300 North Lake Avenue, Suite 1200
 Pasadena, CA 91101
 Attention: Bronwyn Kelly

Project ID: Routine Outfall 002
 Report Number: IOD2044

Sampled: 04/28/05
 Received: 04/28/05

METHOD BLANK/QC DATA

INORGANICS

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5D29078 Extracted: 04/29/05											
Matrix Spike Analyzed: 04/29/2005 (5D29078-MS1)						Source: IOD1888-02					
Total Cyanide	162	5.0	2.2	ug/l	200	ND	81	70-115			
Matrix Spike Dup Analyzed: 04/29/2005 (5D29078-MSD1)						Source: IOD1888-02					
Total Cyanide	156	5.0	2.2	ug/l	200	ND	78	70-115	4	15	
Batch: 5D29091 Extracted: 04/29/05											
Blank Analyzed: 05/04/2005 (5D29091-BLK1)											
Biochemical Oxygen Demand	ND	2.0	0.59	mg/l							
LCS Analyzed: 05/04/2005 (5D29091-BS1)											
Biochemical Oxygen Demand	209	100	30	mg/l	198		106	85-115			
LCS Dup Analyzed: 05/04/2005 (5D29091-BSD1)											
Biochemical Oxygen Demand	208	100	30	mg/l	198		105	85-115	1	20	
Batch: 5D29110 Extracted: 04/29/05											
Blank Analyzed: 04/29/2005 (5D29110-BLK1)											
Turbidity	ND	1.0	0.040	NTU							
Duplicate Analyzed: 04/29/2005 (5D29110-DUP1)						Source: IOD2066-01					
Turbidity	135	5.0	0.20	NTU		130			4	20	
Batch: 5D29129 Extracted: 04/29/05											
Blank Analyzed: 04/29/2005 (5D29129-BLK1)											
Total Dissolved Solids	ND	10	10	mg/l							

Del Mar Analytical, Irvine
 Michele Harper
 Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002
Report Number: IOD2044

Sampled: 04/28/05
Received: 04/28/05

METHOD BLANK/QC DATA

INORGANICS

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC %REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
<u>Batch: 5D29129 Extracted: 04/29/05</u>											
LCS Analyzed: 04/29/2005 (5D29129-BS1)											
Total Dissolved Solids	930	10	10	mg/l	1000		93	90-110			
Duplicate Analyzed: 04/29/2005 (5D29129-DUP1)											
Total Dissolved Solids	334	10	10	mg/l		Source: IOD2033-01 360			7	10	
<u>Batch: 5D29130 Extracted: 04/29/05</u>											
Duplicate Analyzed: 04/29/2005 (5D29130-DUP1)											
Specific Conductance	625	1.0	1.0	umhos/cm		Source: IOD2023-01 640			2	5	
<u>Batch: 5E02067 Extracted: 05/02/05</u>											
Blank Analyzed: 05/02/2005 (5E02067-BLK1)											
Ammonia-N (Distilled)	ND	0.50	0.30	mg/l							
LCS Analyzed: 05/02/2005 (5E02067-BS1)											
Ammonia-N (Distilled)	10.4	0.50	0.30	mg/l	10.0		104	80-115			
Matrix Spike Analyzed: 05/02/2005 (5E02067-MS1)											
Ammonia-N (Distilled)	10.9	0.50	0.30	mg/l	10.0	1.1	98	70-120			
Matrix Spike Dup Analyzed: 05/02/2005 (5E02067-MSD1)											
Ammonia-N (Distilled)	11.2	0.50	0.30	mg/l	10.0	1.1	101	70-120	3	15	
<u>Batch: 5E04071 Extracted: 05/04/05</u>											
Blank Analyzed: 05/04/2005 (5E04071-BLK1)											
Total Suspended Solids	ND	10	10	mg/l							

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002
Report Number: IOD2044

Sampled: 04/28/05
Received: 04/28/05

METHOD BLANK/QC DATA

INORGANICS

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5E04071 Extracted: 05/04/05											
LCS Analyzed: 05/04/2005 (5E04071-BS1)											
Total Suspended Solids	1000	10	10	mg/l	1000		100	85-115			
Duplicate Analyzed: 05/04/2005 (5E04071-DUP1)											
Total Suspended Solids	ND	10	10	mg/l		ND				10	

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD2044

Sampled: 04/28/05
Received: 04/28/05

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits appear in bold on this page.

LabNumber	Analysis	Analyte	Units	Result	MRL	Compliance Limit
IOD2044-01	413.1 Oil and Grease	Oil & Grease	mg/l	-1	5.0	10.00
IOD2044-01	608-Pest Boeing 001/002 Q (LL)	alpha-BHC	ug/l	0	0.010	0.0100
IOD2044-01	624-Boeing 001/002 Q (Fr113+X)	1,1-Dichloroethene	ug/l	0	3.0	3.20
IOD2044-01	624-Boeing 001/002 Q (Fr113+X)	Trichloroethene	ug/l	0.27	5.0	5.00
IOD2044-01	625-Boeing 001/002 Q-LL	2,4,6-Trichlorophenol	ug/l	0	6.0	6.50
IOD2044-01	625-Boeing 001/002 Q-LL	2,4-Dinitrotoluene	ug/l	0.12	9.0	9.10
IOD2044-01	625-Boeing 001/002 Q-LL	Bis(2-ethylhexyl)phthalate	ug/l	0.45	5.0	4.00
IOD2044-01	625-Boeing 001/002 Q-LL	N-Nitrosodimethylamine	ug/l	0	8.0	8.10
IOD2044-01	625-Boeing 001/002 Q-LL	Pentachlorophenol	ug/l	0	8.0	8.20
IOD2044-01	BOD	Biochemical Oxygen Demand	mg/l	5.20	2.0	20
IOD2044-01	Chloride - 300.0	Chloride	mg/l	31	1.0	150
IOD2044-01	Copper-200.8	Copper	ug/l	4.60	2.0	7.10
IOD2044-01	Cyanide-335.2 5ppb	Total Cyanide	ug/l	0	5.0	4.30
IOD2044-01	Lead-200.8	Lead	ug/l	2.40	1.0	2.60
IOD2044-01	MBAS - SM5540-C	Surfactants (MBAS)	mg/l	0.028	0.10	0.50
IOD2044-01	Mercury - 245.1	Mercury	ug/l	0.025	0.20	0.20
IOD2044-01	Nitrogen, NO3+NO2 -N	Nitrate/Nitrite-N	mg/l	0.23	0.26	8.00
IOD2044-01	Perchlorate 314.0	Perchlorate	ug/l	0	4.0	6.00
IOD2044-01	Sulfate-300.0	Sulfate	mg/l	89	1.0	300
IOD2044-01	TDS - SM 2540C	Total Dissolved Solids	mg/l	530	10	950
IOD2044-02	624-Boeing 001/002 Q (Fr113+X)	1,1-Dichloroethene	ug/l	0	3.0	3.20
IOD2044-02	624-Boeing 001/002 Q (Fr113+X)	Trichloroethene	ug/l	0	5.0	5.00

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD2044

Sampled: 04/28/05

Received: 04/28/05

DATA QUALIFIERS AND DEFINITIONS

- J** Estimated value. Analyte detected at a level less than the Reporting Limit (RL) and greater than or equal to the Method Detection Limit (MDL). The user of this data should be aware that this data is of limited reliability.
- M-3** Results exceeded the linear range in the MS/MSD and therefore are not available for reporting. The batch was accepted based on acceptable recovery in the Blank Spike (LCS).
- M-NR1** There was no MS/MSD analyzed with this batch due to insufficient sample volume. See Blank Spike/Blank Spike Duplicate.
- ND** Analyte NOT DETECTED at or above the reporting limit or MDL, if MDL is specified.
- RPD** Relative Percent Difference

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Routine Outfall 002

Report Number: IOD2044

Sampled: 04/28/05

Received: 04/28/05

Certification Summary

Del Mar Analytical, Irvine

Method	Matrix	Nelac	California
EPA 120.1	Water	X	X
EPA 160.2	Water	X	X
EPA 160.5	Water	X	X
EPA 180.1	Water	X	X
EPA 200.8	Water	X	X
EPA 245.1	Water	X	X
EPA 300.0	Water	X	X
EPA 314.0	Water	N/A	X
EPA 335.2	Water	X	X
EPA 350.2	Water		X
EPA 405.1	Water	X	X
EPA 413.1	Water	X	X
EPA 608	Water	X	X
EPA 624	Water	X	X
EPA 625	Water	X	X
SM2540C	Water	X	X
SM5540-C	Water	X	X

Nevada and NELAP provide analyte specific accreditations. Analyte specific information for Del Mar Analytical may be obtained by contacting the laboratory or visiting our website at www.dmalabs.com.

Subcontracted Laboratories

Alta Analytical California Cert #1640, Nevada Cert #CA-413

1104 Windfield Way - El Dorado Hills, CA 95762

Analysis Performed: 1613-Dioxin-HR

Samples: IOD2044-01

Analysis Performed: EDD + Level 4

Samples: IOD2044-01

Del Mar Analytical, Irvine

Michele Harper
Project Manager



17461 Derian Ave., Irvine CA 92606 (949) 261-1022 FAX (949) 261-1228
1014 E. Cooley Dr., Suite A, Colton, CA 92324 (909) 370-4667 FAX (949) 370-1046
9484 Chesapeake Dr., Suite 805, San Diego, CA 92123 (858) 505-8596 FAX (858) 505-9689
9830 South 51st St., Suite B-120, Phoenix, AZ 85044 (480) 785-0043 FAX (480) 785-0851
2520 E. Sunset Rd. #3, Las Vegas, NV 89120 (702) 798-3620 FAX (702) 798-3621

June 20, 2005

MWH- Pasadena / Boeing
300 North Lake Avenue, Suite 1200
Pasadena , CA 91101

Attention: Bronwyn Kelly
Project: Routine Outfall 002
Sampled: 04/28/05
Del Mar Analytical Number: IOD2044

Dear Ms. Kelly:

Alta Analytical Laboratories performed the EPA Method 1613 for tetra-through-octa chlorinated dioxins and furans for the project referenced above. Please use the following cross-reference table when reviewing your results.

MWH ID	Del Mar ID	Alta ID
Outfall 002	IOD2044-01	26112-001

Attached is the original report from the subcontract laboratory. If you have any questions or require further assistance, please do not hesitate to contact me at (949) 261-1022, extension 215.

Sincerely yours,

DEL MAR ANALYTICAL


Michele Harper
Project Manager

Enclosure



May 20, 2005

Alta Project I.D.: 26112

Ms. Michele Harper
Del Mar Analytical, Irvine
17461 Derian Avenue, Suite 100
Irvine, CA 92614

Dear Ms. Harper,

Enclosed are the results for the one aqueous sample received at Alta Analytical Laboratory on April 30, 2005 under your Project Name "IOD2044". This sample was extracted and analyzed using EPA Method 1613 for tetra-through-octa chlorinated dioxins and furans. A standard turnaround time was provided for this work.

An "A" qualifier indicates that the result is greater than the low point in the calibration curve, but lower than the EPA Method 1613 Minimum Level.

The following report consists of a Sample Inventory (Section I), Analytical Results (Section II) and the Appendix, which contains the chain-of-custody, a list of data qualifiers and abbreviations, Alta's current certifications, and copies of the raw data (if requested).

Alta Analytical Laboratory is committed to serving you effectively. If you require additional information, please contact me at 916-933-1640 or by email at mmaier@altalab.com. Thank you for choosing Alta as part of your analytical support team.

Sincerely,

Martha M. Maier
Director of HRMS Services



For information on laboratory accreditation, visit our website at www.nelap.com. For more information on our services, visit our website at www.altalab.com. This report should not be reproduced or used in any manner without the written approval of ALTA.



Alta Analytical Laboratory Inc.

1104 Windfield Way
El Dorado Hills, CA 95762

FAX (916) 673-0106
(916) 933-1640

Section I: Sample Inventory Report

Date Received: 4/30/2005

Alta Lab. ID

Client Sample ID

26112-001

IOD2044-01

SECTION II



Method Blank		EPA Method 1613			
Matrix: Aqueous	QC Batch No.: 6789	Lab Sample: 0-MB001	Date Analyzed DB-5: 19-May-05	Date Analyzed DB-225: NA	
Sample Size: 1.000 L	Date Extracted: 17-May-05				
Analyte	Conc. (ug/L)	DL ^a	EMPC ^b	%R	LCL-UCL ^d Qualifiers
2,3,7,8-TCDD	ND	0.00000124		69.9	25 - 164
1,2,3,7,8-PeCDD	ND	0.00000166		84.1	25 - 181
1,2,3,4,7,8-HxCDD	ND	0.00000186		72.5	32 - 141
1,2,3,6,7,8-HxCDD	ND	0.00000179		75.3	28 - 130
1,2,3,7,8,9-HxCDD	ND	0.00000186		65.8	23 - 140
1,2,3,4,6,7,8-HpCDD	ND	0.00000303		58.4	17 - 157
OCDD	ND	0.00000677		81.1	24 - 169
2,3,7,8-TCDF	ND	0.000000924		79.5	24 - 185
1,2,3,7,8-PeCDF	ND	0.00000226		82.4	21 - 178
2,3,4,7,8-PeCDF	ND	0.00000193		72.6	26 - 152
1,2,3,4,7,8-HxCDF	ND	0.000000785		75.4	26 - 123
1,2,3,6,7,8-HxCDF	ND	0.000000731		92.3	28 - 136
2,3,4,6,7,8-HxCDF	ND	0.000000672		68.4	29 - 147
1,2,3,7,8,9-HxCDF	ND	0.00000158		63.5	28 - 143
1,2,3,4,6,7,8-HpCDF	ND	0.000000969		52.9	26 - 138
1,2,3,4,7,8,9-HpCDF	ND	0.00000192		49.2	17 - 157
OCDF	ND	0.00000476		89.9	35 - 197
Totals					
Total TCDD	ND	0.00000124			
Total PeCDD	ND	0.00000166			
Total HxCDD	ND	0.00000183			
Total HpCDD	ND	0.00000303			
Total TCDF	ND	0.000000924			
Total PeCDF	ND	0.00000209			
Total HxCDF	ND	0.000000872			
Total HpCDF	ND	0.00000132			
Footnotes					
a. Sample specific estimated detection limit.					
b. Estimated maximum possible concentration.					
c. Method detection limit.					
d. Lower control limit - upper control limit.					

Analyst: RAS

Approved By: William J. Luksemburg

20-May-2005 10:57



EPA Method 1613

OPR Results		Lab Sample: 0-OPR001		Date Analyzed DB-5: 19-May-05		Date Analyzed DB-225: NA	
Matrix:	Aqueous	QC Batch No.:	6789	Date Analyzed DB-5:	17-May-05	Date Analyzed DB-225:	NA
Sample Size:	1.000 L	Date Extracted:		Date Analyzed DB-5:	19-May-05	Date Analyzed DB-225:	NA
Analyte	Spike Conc.	Conc. (ng/mL)	OPR Limits	Labeled Standard	%R	LCL-UCL	
2,3,7,8-TCDD	10.0	10.3	6.7 - 15.8	<u>IS</u> 13C-2,3,7,8-TCDD	66.3	25 - 164	
1,2,3,7,8-PeCDD	50.0	51.8	35 - 71	13C-1,2,3,7,8-PeCDD	82.1	25 - 181	
1,2,3,4,7,8-HxCDD	50.0	50.1	35 - 82	13C-1,2,3,4,7,8-HxCDD	69.4	32 - 141	
1,2,3,6,7,8-HxCDD	50.0	52.2	38 - 67	13C-1,2,3,6,7,8-HxCDD	74.5	28 - 130	
1,2,3,7,8,9-HxCDD	50.0	54.3	32 - 81	13C-1,2,3,4,6,7,8-HpCDD	64.6	23 - 140	
1,2,3,4,6,7,8-HpCDD	50.0	49.7	35 - 70	13C-OCDD	40.2	17 - 157	
OCDD	100	99.1	78 - 144	13C-2,3,7,8-TCDF	71.3	24 - 169	
2,3,7,8-TCDF	10.0	10.1	7.5 - 15.8	13C-1,2,3,7,8-PeCDF	78.8	24 - 185	
1,2,3,7,8-PeCDF	50.0	49.0	40 - 67	13C-2,3,4,7,8-PeCDF	85.0	21 - 178	
2,3,4,7,8-PeCDF	50.0	49.2	34 - 80	13C-1,2,3,4,7,8-HxCDF	72.8	26 - 152	
1,2,3,4,7,8-HxCDF	50.0	48.2	36 - 67	13C-1,2,3,6,7,8-HxCDF	78.4	26 - 123	
1,2,3,6,7,8-HxCDF	50.0	48.8	42 - 65	13C-2,3,4,6,7,8-HxCDF	82.5	28 - 136	
2,3,4,6,7,8-HxCDF	50.0	48.4	35 - 78	13C-1,2,3,7,8,9-HxCDF	69.8	29 - 147	
1,2,3,7,8,9-HxCDF	50.0	49.7	39 - 65	13C-1,2,3,4,6,7,8-HpCDF	58.1	28 - 143	
1,2,3,4,6,7,8-HpCDF	50.0	49.7	41 - 61	13C-1,2,3,4,7,8,9-HpCDF	45.9	26 - 138	
1,2,3,4,7,8,9-HpCDF	50.0	50.6	39 - 69	13C-OCDF	36.3	17 - 157	
OCDF	100	93.6	63 - 170	<u>CRS</u> 37Cl-2,3,7,8-TCDD	85.6	35 - 197	

Analyst: RAS

Approved By: William J. Luksemburg 20-May-2005 10:57



Sample ID: IOD2044-01

EPA Method 1613

Client Data		Sample Data		Laboratory Data			
Name: Project: Date Collected: Time Collected:	Del Mar Analytical, Irvine IOD2044 28-Apr-05 1406	Matrix: Sample Size:	Aqueous 0.950 L	Lab Sample: QC Batch No.: Date Analyzed DB-5:	26112-001 6789 19-May-05	Date Received: Date Extracted: Date Analyzed DB-225: NA	30-Apr-05 17-May-05 NA
Analyte	Conc. (ug/L)	DL ^a	EMPC ^b	Qualifiers	Labeled Standard	%R	LCL-UCL ^d Qualifiers
2,3,7,8-TCDD	ND	0.00000199			IS 13C-2,3,7,8-TCDD	61.2	25 - 164
1,2,3,7,8-PeCDD	ND	0.00000294			13C-1,2,3,7,8-PeCDD	65.5	25 - 181
1,2,3,4,7,8-HxCDD	ND	0.00000400			13C-1,2,3,4,7,8-HxCDD	63.8	32 - 141
1,2,3,6,7,8-HxCDD	ND	0.00000399			13C-1,2,3,6,7,8-HxCDD	65.8	28 - 130
1,2,3,7,8,9-HxCDD	ND	0.00000409			13C-1,2,3,4,6,7,8-HpCDD	61.5	23 - 140
1,2,3,4,6,7,8-HpCDD	0.0000557				13C-OCDD	45.0	17 - 157
OCDD	0.000706				13C-2,3,7,8-TCDF	66.5	24 - 169
2,3,7,8-TCDF	ND	0.00000200			13C-1,2,3,7,8-PeCDF	63.6	24 - 185
1,2,3,7,8-PeCDF	ND	0.00000362			13C-2,3,4,7,8-PeCDF	66.3	21 - 178
2,3,4,7,8-PeCDF	ND	0.00000288			13C-1,2,3,4,7,8-HxCDF	65.2	26 - 152
1,2,3,4,7,8-HxCDF	ND	0.00000117			13C-1,2,3,6,7,8-HxCDF	69.0	26 - 123
1,2,3,6,7,8-HxCDF	ND	0.00000165			13C-2,3,4,6,7,8-HxCDF	70.5	28 - 136
2,3,4,6,7,8-HxCDF	ND	0.00000118			13C-1,2,3,7,8,9-HxCDF	62.6	29 - 147
1,2,3,7,8,9-HxCDF	ND	0.00000214			13C-1,2,3,4,6,7,8-HpCDF	58.0	28 - 143
1,2,3,4,6,7,8-HpCDF	0.00000968			A	13C-1,2,3,4,7,8,9-HpCDF	49.7	26 - 138
1,2,3,4,7,8,9-HpCDF	ND				13C-OCDF	43.8	17 - 157
OCDF	0.0000306			A	CRS 37Cl-2,3,7,8-TCDD	78.7	35 - 197
Totals							
Total TCDD	ND	0.00000199					
Total PeCDD	ND	0.00000294					
Total HxCDD	0.00000660		0.0000135				
Total HpCDD	0.000114						
Total TCDF	0.00000366						
Total PeCDF	ND	0.00000322					
Total HxCDF	0.00000666						
Total HpCDF	0.0000253		0.00000980				

Footnotes
a. Sample specific estimated detection limit.
b. Estimated maximum possible concentration.
c. Method detection limit.
d. Lower control limit - upper control limit.

Analyst: RAS

Approved By: William J. Luksemburg 20-May-2005 10:57

APPENDIX

DATA QUALIFIERS & ABBREVIATIONS

B	This compound was also detected in the method blank.
D	The amount reported is the maximum possible concentration due to possible chlorinated diphenylether interference.
H	The signal-to-noise ratio is greater than 10:1.
I	Chemical Interference
J	The amount detected is below the Lower Calibration Limit of the instrument.
P	Homologue totals include any coplanar PCBs detected at concentrations less than the reporting limit.
*	See Cover Letter
Conc.	Concentration
DL	Sample-specific estimated detection limit
MDL	The minimum concentration of a substance that can be measured and reported with 99% confidence that the analyte concentration is greater than zero in the matrix tested.
EMPC	Estimated Maximum Possible Concentration
NA	Not applicable
RL	Reporting Limit – concentrations that correspond to low calibration point
ND	Not Detected
TEQ	Toxic Equivalency

Unless otherwise noted, solid sample results are reported in dry weight. Tissue samples are reported in wet weight.

The control limits are “interim limits only” until in-house limits are utilized.

CURRENT CERTIFICATIONS

NELAP — (Primary AA: California, Certificate No. 02102CA)

Department of the Navy

U.S. Army Corps of Engineers

U.S. EPA Region 5

Bureau of Reclamation — Mid-Pacific Region — (MP-470, Res-1.10)

Commonwealth of Kentucky — (Certificate No. 90063)

Commonwealth of Virginia — (Certificate No. 00013)

State of Alaska, Department of Environmental Conservation — (Certificate No. OS-00197)

State of Arizona — (Certificate No. AZ0639)

State of Arkansas, Department of Health — (Approval granted through CA certification)

State of Arkansas, Department of Environmental Quality

State of California — (Certificate No. 1640)

State of Colorado

State of Connecticut — (Certificate No. PH-0182)

State of Florida — (Certificate No. 87456)

State of Louisiana, Department of Health and Hospitals — (Certificate No. LA000014)

State of Louisiana, Department of Environmental Quality

State of Maine

State of Michigan (Certificate No. 81178087)

State of Mississippi — (Approval granted through CA certification)

State of Nevada — (Certificate No. CA413)

State of New Jersey — (Certificate No. CA003)

State of New York, Department of Health — (Certificate No. 11411)

State of North Carolina — (Certification No. 06700)

State of North Dakota, Department of Health — (Certificate No. R-078)

State of New Mexico

State of Oklahoma — (D9919)

State of Oregon — (Certificate No. CA413)

State of Pennsylvania — (Certificate No. 68-490)

State of South Carolina — (Certificate No. 87002001)

State of Tennessee — (Certificate No. 02996)

State of Texas — (Certificate No. TX247-1000A)

State of Utah — (Certificate No. E-201)

State of Washington — (Certification No. C091)

State of Wisconsin — (Certificate No. 998036160)

State of Wyoming — (USEPA Region 8 Ref: 8TMS-Q)



17461 Derian Ave. Suite 100, Irvine, CA 92614 Ph (949) 261-1022 Fax (949) 261-1228
 1014 E. Cooley Dr., Suite A, Colton, CA 92324 Ph (909) 370-4667 Fax (909) 370-1046
 9484 Chesapeake Drive, Suite 805, San Diego, CA 92123 Ph (619) 505-9598 Fax (619) 505-9689
 9830 South 51st Street, Suite B-120, Phoenix, AZ 85044 Ph (480) 785-0043 Fax (480) 785-0851
 2520 E. Sunset Rd., Suite #3, Las Vegas, NV 89120 Ph (702) 798-3820 Fax (702) 798-3821

SUBCONTRACT ORDER - PROJECT # IOD2044

SENDING LABORATORY:	RECEIVING LABORATORY:
Del Mar Analytical, Irvine 17461 Derian Avenue. Suite 100 Irvine, CA 92614 Phone: (949) 261-1022 Fax: (949) 261-1228 Project Manager: Michele Harper	Alta Analytical 1104 Windfield Way El Dorado Hills, CA 95762 Phone : (916) 933-1640 Fax: (916) 673-0106 <div style="text-align: right; font-size: 1.2em; font-family: cursive;"> 26112 1.1°C </div>

Standard TAT is requested unless specific due date is requested => Due Date: _____ Initials: _____

Analysis	Expiration	Comments
Sample ID: IOD2044-01 Water	Sampled: 04/28/05 14:06	Instant Notification
1613-Dioxin-HR	05/05/05 14:06	J flags, 17 congeners, no TEQ, sub=Alta, DP to AMEC
EDD + Level 4	05/26/05 14:06	Excel EDD email to pm, Include Std logs for Lvl IV
Containers Supplied:		
1 L Amber (IOD2044-01G)		
1 L Amber (IOD2044-01H)		

SAMPLE INTEGRITY:

All containers intact: Yes No
 Sample labels/COC agree: Yes No
 Samples Received On Ice: Yes No
 Custody Seals Present: Yes No
 Samples Preserved Properly: Yes No
 Samples Received at (temp): _____

Released By:
 Date: 4/29/07
 Time: 17:00
 Received By: M Jollent
 Date: 4/30/05
 Time: 0915

Released By _____ Date _____ Time _____ Received By _____ Date _____ Time _____

STANDARD OPERATING PROCEDURE

Attachment 10.B.1

SAMPLE LOG-IN CHECKLIST

ALTA Project No.: 26112

1. Date Samples Arrived: <u>4/30/05 0915</u> Initials: <u>MU</u> Location: <u>WR-2</u>			
2. Time / Date logged in: <u>0915 5/2/05</u> Initials: <u>BBB</u> Location: <u>WR-2</u>			
3. Samples Arrived By: (circle) <u>FedEx</u> UPS World Courier Other:			
4. Shipping Preservation: (circle) <u>Ice</u> <u>Blue Ice</u> / Dry Ice / None Temp °C <u>1.1</u>			
5. Shipping Container(s) Intact? If not, describe condition in comment section.	YES ✓	NO	NA
6. Shipping Container(s) Custody Seals Present? Intact? If not intact, describe condition in comment section.		✓	✓
7. Shipping Documentation Present? (circle) Shipping Label <u>Airbill</u> Tracking Number <u>7916 1353 5260</u>	✓		
8. Sample Custody Seal(s) Present? No. of Seals _____ or Seal No. Intact? If not intact, describe condition in comment section.		✓	✓
9. Sample Container Intact? If no, indicate sample condition in comment section.	✓		
10. Chain of Custody (COC) or other Sample Documentation Present?	✓		
11. COC/Documentation Acceptable? If no, complete COC Anomaly Form.	✓		
12. Shipping Container (circle): ALTA <u>Client</u> Retain or <u>Return</u> or Disposed			
13. Container(s) and/or Bottle(s) Requested?		✓	
14. Drinking Water Sample? (HRMS Only) If yes, Acceptable Preservation? Y or N Preservation Info From? (circle) COC or Sample Container or None Noted			✓

Comments: Sampler's initials found on sample labels

ALTA Analytical Laboratory
El Dorado Hills, CA 95762

