

Table 1. Sample identification

Client ID	EPA ID	Lab No.	Matrix	Method
Outfall 012	Outfall 012	IOC2360-01	water	8015B/EFH

2. DATA VALIDATION FINDINGS

2.1 SAMPLE MANAGEMENT

The following are findings associated with sample management:

2.1.1 Sample Preservation, Handling, and Transport

The sample in this SDG was received at Del Mar Analytical laboratory on ice within the temperature limits of $4^{\circ}\text{C} \pm 2^{\circ}\text{C}$. The Del Mar Analytical case narrative noted that the sample containers were received intact. No qualifications were required.

2.1.2 Chain of Custody

The COC was signed and dated by both field and laboratory personnel, and accounted for the analysis presented in this SDG. The EFH analysis (rather than the GRO analysis) was requested in error on the COC for the Trip Blank sample. The sample was analyzed correctly. As the site sample was couriered directly to the laboratory, custody seals were not required. No qualifications were required.

2.1.3 Holding Times

The sample was extracted within seven days of sample collection and analyzed within 40 days of extraction. No qualifications were required.

2.2 CALIBRATION

The initial calibration associated with the sample analysis was analyzed on 03/11/05. The %RSD was within the QC limit of $\leq 20\%$. The %Ds for the initial calibration verification (ICV) and continuing calibrations associated with the sample analysis were $\leq 15\%$. The %RSD and %Ds were recalculated from the raw data and no transcription or calculation errors were noted. No qualifications were required.

2.4 METHOD BLANKS

One method blank (5C31011-BLK1) was extracted and analyzed with the sample in this SDG. EFH (C13-C22) was not present above the MDL in the method blank or in the instrument blank analyzed at the beginning of the analytical sequence. Review of the chromatograms showed no false negatives. No qualifications were required.

2.5 BLANK SPIKES AND LABORATORY CONTROL SAMPLES

One method blank spike/blank spike duplicate pair (5C31011-BS1/BSD1) was extracted and analyzed with the sample in this SDG. The laboratory reported recoveries of alkane range C13-C28 from spiked diesel. The recoveries were within the laboratory-established QC limits of 40-120%,

and the RPD was within the QC limit of $\leq 25\%$. The recoveries and RPD were checked from the raw data, and no calculation or transcription errors were noted. No qualifications were required.

2.6 SURROGATE RECOVERY

The sample was fortified with the surrogate compound n-octacosane. The sample surrogate recovery was within the laboratory-established QC limits of 40-125%. The recovery was calculated from the raw data and no transcription or calculation errors were noted. No qualifications were required.

2.7 MATRIX SPIKE/MATRIX SPIKE DUPLICATE

There were no MS/MSD analyses associated with the sample of this SDG. Evaluation of method accuracy and precision was based on the BS/BSD results. No qualifications were required.

2.8 FIELD QC SAMPLES

Field QC samples are evaluated, and if necessary, qualified based on method blanks and laboratory QC samples for usability. Any remaining detects are used to evaluate the associated sample. The following are findings associated with field QC samples:

2.9.1 Field Blanks and Equipment Rinsates

There were no field blank or equipment rinsate samples associated with the site sample in this SDG. No qualifications were required.

2.9.2 Field Duplicates

There were no field duplicate samples associated with this SDG.

2.10 COMPOUND IDENTIFICATION

The laboratory analyzed for EFH n-alkane range C13-C22 by EPA SW846 Method 8015M. Compound identification is verified at a Level IV validation. Review of chromatograms and retention times indicated no problems with compound identification for this SDG. No qualifications were required.

2.11 COMPOUND QUANTIFICATION AND REPORTED DETECTION LIMITS

Compound quantification was verified for this SDG by recalculating any sample detect, blank spike recoveries, and a representative number of surrogate recoveries. Reporting limits were supported by the low level standard of the initial calibration and by the laboratory MDL. Results were reported in mg/L (ppm). No qualifications were required.



Del Mar Analytical

17461 Derian Ave., Suite 100, Irvine, CA 92614 (949) 261-1022 FAX (949) 260-0297
 1014 E. Cooley Dr., Suite A, Colton, CA 92324 (909) 370-4667 FAX (941) 370-1046
 9484 Chesapeake Dr., Suite 805, San Diego, CA 92123 (619) 505-8596 FAX (619) 505-9689
 9830 South 51st St., Suite B-120, Phoenix, AZ 85044 (480) 785-0043 FAX (480) 785-0851
 2320 E. Sunset Rd. #3, Las Vegas, NV 89120 (702) 798-3620 FAX (702) 798-3621

MWH-Pasadena/Boeing
 300 North Lake Avenue, Suite 1200
 Pasadena, CA 91101
 Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOC2360

Sampled: 03/30/05
 Received: 03/30/05

DRAFT: EXTRACTABLE FUEL HYDROCARBONS (CADHS/8015 Modified)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Analyzed	Date Analyzed	Data Qualifiers
Sample ID: IOC2360-01 (DRAFT: Outfall 012 - Water) - cont.									
Reporting Units: mg/l									
EFH (C13 - C22)	EPA 8015B	5C31011	0.082	0.50	0.66	0.962	03/31/05	03/31/05	rel qual Code
Surrogate: n-Octacosane (40-125%)					94 %				

AMEC VALIDATED
LEVEL IV

DRAFT REPORT
 DRAFT REPORT
 DATA SUBJECT TO CHANGE

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from Del Mar Analytical.

CONTRACT COMPLIANCE SCREENING FORM FOR HARDCOPY DATA

AMEC Earth & Environmental
 550 South Wadsworth Boulevard
 Suite 500
 Lakewood, CO 80226


Package ID T711VO93
 Task Order 313150010
 SDG No. IOC2360

No. of Analyses 1

Laboratory Del Mar

Reviewer M. Pokorny

Analysis/Method Volatiles (1,4-dioxane)

Date: April 13, 2005
 Reviewer's Signature


ACTION ITEMS*	
1. Case Narrative Deficiencies	
2. Out of Scope Analyses	
3. Analyses Not Conducted	
4. Missing Hardcopy Deliverables	
5. Incorrect Hardcopy Deliverables	
6. Deviations from Analysis Protocol, e.g., Holding Times GC/MS Tune/Inst. Perform Calibrations Blanks Surrogates Matrix Spike/Dup LCS Field QC Internal Standard Performance Compound Identification and Quantitation System Performance	Qualification required for continuing calibration outlier.
COMMENTS ^b	
^a Subcontracted analytical laboratory is not meeting contract and/or method requirements. ^b Differences in protocol have been adopted by the laboratory but no action against the laboratory is required.	



DATA VALIDATION REPORT

NPDES Monitoring

ANALYSIS: VOLATILES

SAMPLE DELIVERY GROUP: IOC2360

Prepared by

AMEC—Denver Operations
550 South Wadsworth Boulevard, Suite 500
Lakewood, Colorado 80226

1. INTRODUCTION

Task Order Title: NPDES Monitoring
Contract Task Order #: 313150010
Sample Delivery Group #: IOC2360
Project Manager: B. McIlvaine
Matrix: Water
Analysis: Volatiles (1,4-dioxane)
QC Level: Level IV
No. of Samples: 1
No. of Reanalyses/Dilutions: 0
Reviewer: M. Pokorny
Date of Review: April 13, 2005

The samples listed in Table 1 were validated based on the guidelines outlined in *the AMEC Data Validation Procedure for Levels C and D Volatile Organics (DVP-2, Rev. 2)*, *EPA Method SW-846 8260B* and the *National Functional Guidelines For Organic Data Review (2/94)*. Any deviations from these procedures and guidelines are documented herein. Qualifiers were applied in cases where the data did not meet the required QC criteria or where special consideration by the data user is required. Data qualifiers were placed on Form Is with the associated qualification codes. Analytes that were rejected for any reason are denoted on the Form I as having only the "R" data qualifier and associated qualification code(s) denoting the reason for rejection. Any additional problems with the data that may have resulted in an estimated value were not denoted by a qualification code since the data had already been rejected.

Table 1. Sample identification

Client ID	EPA ID	Lab No.	Matrix	Method
Outfall 012	Outfall 012	IOC2360-01	water	624

2. DATA VALIDATION FINDINGS

2.1 SAMPLE MANAGEMENT

Following are findings associated with sample management:

2.1.1 Sample Preservation, Handling, and Transport

The sample in this SDG was received at Del Mar, and at the subcontractor laboratory, within the temperature limits of $4^{\circ}\text{C} \pm 2^{\circ}\text{C}$. The sample was properly preserved. The COCs noted that the sample was received intact; however, information regarding absence of headspace was not provided. No qualifications were required.

2.1.2 Chain of Custody

The original and transfer COCs were signed by field and laboratory personnel. The COCs accounted for the analysis presented in this SDG. According to the sample login sheet, custody seals were not present on the cooler. The sample summary form did not have the Outfall 012 ID printed on it; the ID was added by the reviewer. No qualifications were required.

2.1.3 Holding Times

The sample was analyzed within 14 days of collection. No qualifications were required.

2.2 GC/MS TUNING

The ion abundance windows were consistent with those specified in EPA Method 8260B. All ion abundances were within the established windows, and the sample was analyzed within 12 hours of the BFB injection time. No qualifications were required.

2.3 CALIBRATION

One initial calibration, dated 03/21/05, was associated with this SDG. The average RRF for 1,4-dioxane was ≥ 0.05 and the r^2 value was ≤ 0.995 . One continuing calibration, dated 04/02/05 was associated with this SDG. The RRF for 1,4-dioxane was ≥ 0.05 and the %D was $> 20\%$. 1,4-Dioxane was qualified as an estimated nondetect, "UJ," in sample Outfall 012. The r^2 value and average RRF for 1,4-dioxane in the initial calibration, and the %D and RRF for 1,4-dioxane in the continuing calibration were recalculated from the raw data, and no calculation or transcription errors were found. No further qualifications were required.

2.4 BLANKS

One water method blank (P5D0201-BLK1) was associated with this SDG. Target compound 1,4-dioxane was not detected in the method blank. The method blank raw data showed no evidence of a false negative. No qualifications were required.

2.5 BLANK SPIKES AND LABORATORY CONTROL SAMPLES

The laboratory analyzed a blank spike/blank spike duplicate pair (P5D0201-BS1/BS1D) with this SDG. The recoveries and RPD for 1,4-dioxane were within the laboratory QC limits. A representative recovery was recalculated from the raw data and no calculation or transcription errors were found. No qualifications were required.

2.6 SURROGATE RECOVERY

The samples and QC were fortified with dibromofluoromethane. The surrogate was recovered within the laboratory QC limits of 80-125%. The surrogate recovery for this sample was recalculated from the raw data and no calculation or transcription errors were found. No qualifications were required.

2.7 MATRIX SPIKE/MATRIX SPIKE DUPLICATE

No MS/MSD analyses were associated with this SDG. Evaluation of method accuracy and precision were based on blank spike and blank spike duplicate results. No qualifications were required.

2.8 FIELD QC SAMPLES

Field QC samples were evaluated, and if necessary, qualified based on method blanks and other laboratory QC results affecting the usability of the field QC data. Any remaining detects were used to evaluate the associated site sample. Following are findings associated with field QC samples:

2.8.1 Trip Blanks

The sample in this SDG had no associated trip blank. No qualifications were required.

2.8.1 Field Blanks and Equipment Rinsates

The site sample in this SDG had no associated field QC samples. No qualifications were required.

2.8.2 Field Duplicates

There were no field duplicate samples associated with this SDG.

2.9 INTERNAL STANDARDS PERFORMANCE

Internal standard area counts and retention times for the sample were within the control limits established by the continuing calibration standards: +100%/-50% for internal standard areas and ± 0.50 minutes for retention times. Internal standard areas and retention times were verified from the raw data, and no calculation or transcription errors were noted. No qualifications were required.

2.10 COMPOUND IDENTIFICATION

Target compound identification was verified at a Level IV data validation. The laboratory analyzed for 1,4-dioxane by Method 8260B/SIM. Chromatograms, retention times, and spectra for the sample and QC were examined and no target compound identification problems were noted. No qualifications were required.

2.11 COMPOUND QUANTIFICATION AND REPORTED DETECTION LIMITS

Compound quantification is verified at a Level IV data validation. The reporting limit was supported by the lowest concentration of the initial calibration standards and by the undated MDL supplied by the laboratory. Compound quantitation was verified by recalculating blank spike and surrogate recoveries from the raw data. No calculation or transcription errors were noted. No qualifications were required.

2.12 TENTATIVELY IDENTIFIED COMPOUNDS

TICs are not typically reported for SIM methods.

2.13 SYSTEM PERFORMANCE

A review of the chromatograms and other raw data showed no identifiable problems with system performance. No qualifications were required.



Del Mar Analytical

1746 Derian Ave., Suite 100, Irvine, CA 92614 (949) 261-1022 FAX (949) 260-3
 1014 E. Cooley Dr., Suite A, Colton, CA 92324 (909) 370-4667 FAX (949) 370-1
 9484 Chesapeake Dr., Suite 805, San Diego, CA 92123 (858) 505-8596 FAX (619) 505-9
 9830 South 51st St., Suite B-126, Phoenix, AZ 85044 (480) 785-0043 FAX (480) 785-0
 2520 E. Sunset Rd. #3, Las Vegas, NV 89120 (702) 798-3620 FAX (702) 798-3

Del Mar Analytical - Irvine
 17461 Derian Ave. Suite 100
 Irvine, CA 92614
 Attention: Michele Harper

Project ID: IOC2360

Report Number: POD0005

Sampled: 03/30/05
 Received: 04/01/05

1,4-DIOXANE BY GC/MS (EPA 5030B/8260B)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: POD0005-01 (IOC2360-01 - Water)									
Reporting Units: ug/l									
1,4-Dioxane	EPA 8260B	P5D0201	0.49	1.0	ND 116%	1	04/02/05	04/02/05	REV QUAL UJC
Surrogate: Dibromofluoromethane (80-125%)									

AMEC VALIDATED

LEVEL IV

Del Mar Analytical - Phoenix
 Karen Maxwell
 Project Manager


The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from Del Mar Analytical.

CONTRACT COMPLIANCE SCREENING FORM FOR HARDCOPY DATA

AMEC Earth & Environmental
 550 South Wadsworth Boulevard
 Suite 500
 Lakewood, CO 80226

Package ID T711VO94
 Task Order 313150010
 SDG No. IOC2360
 No. of Analyses 2

Laboratory Del Mar
 Reviewer M. Pokorny
 Analysis/Method Volatiles

Date: April 13, 2005
 Reviewer's Signature


ACTION ITEMS^a	
1. Case Narrative	
Deficiencies	
2. Out of Scope	
Analyses	
3. Analyses Not Conducted	
4. Missing Hardcopy	
Deliverables	
5. Incorrect Hardcopy	
Deliverables	
6. Deviations from Analysis	
Protocol, e.g.,	
Holding Times	
GC/MS Tune/Inst. Perform	
Calibrations	
Blanks	
Surrogates	
Matrix Spike/Dup LCS	
Field QC	
Internal Standard Performance	
Compound Identification and	
Quantitation	
System Performance	
COMMENTS^b	Acceptable as reviewed.
^a Subcontracted analytical laboratory is not meeting contract and/or method requirements. ^b Differences in protocol have been adopted by the laboratory but no action against the laboratory is required.	



DATA VALIDATION REPORT

NPDES Monitoring

ANALYSIS: VOLATILES

SAMPLE DELIVERY GROUP: IOC2360

Prepared by

AMEC Denver Operations
550 South Wadsworth Boulevard, Suite 500
Lakewood, Colorado 80226

1. INTRODUCTION

Task Order Title: NPDES Monitoring
Contract Task Order #: 313150010
SDG#: IOC2360
Project Manager: B. McIlvaine
Matrix: Water
Analysis: Volatiles
QC Level: Level IV
No. of Samples: 2
No. of Reanalyses/Dilutions: 0
Reviewer: M. Pokorny
Date of Review: April 13, 2005

The samples listed in Table 1 were validated based on the guidelines outlined in the *AMEC Data Validation Procedure for Levels C and D Volatile Organics (DVP-2, Rev. 2)*, *EPA Method 624*, *EPA SW-846 Method 8260B*, and the *National Functional Guidelines For Organic Data Review (2/94)*. Any deviations from these procedures are documented herein. Qualifiers were applied in cases where the data did not meet the required QC criteria or where special consideration by the data user is required. Data qualifiers were placed on Form Is with the associated qualification codes. Analytes that were rejected for any reason are denoted on the summary forms as having only the "R" data qualifier and associated qualification code(s) denoting the reason for rejection. Any additional problems with the data that may have resulted in an estimated value were not denoted by a qualification code since the data had already been rejected.

Table 1. Sample identification

Client ID	EPA ID	Lab No.	Matrix	Method
Outfall 012	Outfall 012	IOC2360-01	water	624
Trip Blank	Trip Blank	IOC2360-02	water	624

2. DATA VALIDATION FINDINGS

2.1 SAMPLE MANAGEMENT

The following are findings associated with sample management:

2.1.1 Sample Preservation, Handling, and Transport

The samples in this SDG were received at the laboratory within the temperature limits of 4°C ±2°C. The COC noted that the samples were received intact; however, information regarding absence of headspace was not provided. No qualifications were required.

2.1.2 Chain of Custody

The COC was signed and dated by both field and laboratory personnel. The COCs accounted for the analyses presented in this SDG. As the samples were couriered directly to the laboratory, custody seals were not required. No qualifications were required.

2.1.3 Holding Times

The samples were analyzed within 14 days of collection. No qualifications were required.

2.2 GC/MS TUNING

The ion abundance windows shown on the quantitation reports were consistent with those specified in EPA Method 624, and all ion abundances were within the established windows. The samples and associated QC were analyzed within 12 hours of the BFB injection times. The Form Vs were verified from the raw data and no discrepancies between the summary forms and the raw data were noted. No qualifications were required.

2.3 CALIBRATION

One initial calibration dated 03/04/05 was associated with this SDG. The average RRFs were ≥0.05 for all compounds listed on the sample result summaries. The %RSDs were ≤35% for the target compounds analyzed by EPA Method 624. One continuing calibration associated with the sample analyses was analyzed 03/31/05. The RRFs were ≥0.05 in the continuing calibration. The %Ds for the continuing calibrations associated with the site samples were all ≤20%. A representative number of %RSDs and average RRFs from the initial calibrations, and %Ds and RRFs from the continuing calibrations were recalculated from the raw data, and no calculation or transcription errors were found. No qualifications were required.

2.4 BLANKS

One water method blank (5C31014-BLK1) were associated with the sample analyses. There were no detects above the MDLs for the target compounds listed on the sample result summaries. The method blank raw data showed no evidence of false negatives. No qualifications were required.

2.5 BLANK SPIKES AND LABORATORY CONTROL SAMPLES

One water blank spike (5C31014-BS1) was associated with the sample analyses. All recoveries were within the laboratory-established QC limits. A representative number of recoveries were recalculated from the raw data and no calculation or transcription errors were found. No qualifications were required.

2.6 SURROGATE RECOVERY

The surrogates were recovered within the QC limits of 80-120% in the samples and associated QC. A representative number of surrogate recoveries were recalculated from the raw data and no calculation or transcription errors were found. No qualifications were required.

2.7 MATRIX SPIKE/MATRIX SPIKE DUPLICATE

No MS/MSD analyses were associated with this SDG. Evaluation of method accuracy and precision were based on blank spike and blank spike duplicate results. No qualifications were required.

2.8 FIELD QC SAMPLES

Field QC samples were evaluated, and if necessary, qualified based on method blanks and other laboratory QC results affecting the usability of the field QC data. Any remaining detects were used to evaluate the associated site sample. Following are findings associated with field QC samples:

2.8.1 Trip Blanks

Sample Trip Blank (IOC2360-02) was the trip blank associated with this SDG. No target compounds were reported in the Trip Blank. No qualifications were required.

2.8.2 Field Blanks and Equipment Rinsates

There were no field QC samples associated with this SDG. No qualifications were required.

2.8.3 Field Duplicates

There were no field duplicate samples associated with this SDG.

2.9 INTERNAL STANDARDS PERFORMANCE

Internal standard area counts and retention times for the samples in this SDG were within the control limits established by the continuing calibration standards: +100%/-50% for internal standard areas and ± 0.50 minutes for retention times. A representative number of internal standard areas and retention times were verified from the raw data, and no calculation or transcription errors were noted. No qualifications were required.

2.10 COMPOUND IDENTIFICATION

Target compound identification was verified at a Level IV data validation. The laboratory analyzed for five volatile target compounds by EPA Method 624. Chromatograms, retention times, and spectra for the samples and QC were examined and no target compound identification problems were noted. No qualifications were required.

2.11 COMPOUND QUANTIFICATION AND REPORTED DETECTION LIMITS

Compound quantification is verified at a Level IV data validation. The reporting limits were supported by the lowest concentrations of the initial calibration standards and by the MDL study. Compound quantitation was verified by recalculating any sample detects and a representative number of blank spike and surrogate recoveries from the raw data. Results were reported in $\mu\text{g/L}$ (ppb). No calculation or transcription errors were noted. No qualifications were required.

2.12 TENTATIVELY IDENTIFIED COMPOUNDS

The laboratory did not provide TICs for this SDG. No qualifications were required.

2.13 SYSTEM PERFORMANCE

A review of the chromatograms and other raw data showed no identifiable problems with system performance. No qualifications were required.



17461 Derian Ave., Suite 100, Irvine, CA 92614 (949) 261-1022 FAX (949) 260-3297
 1014 E. Cooley Dr., Suite A, Colton, CA 92324 (909) 370-4667 FAX (949) 370-1046
 9484 Chesapeake Dr., Suite 805, San Diego, CA 92123 (658) 503-8596 FAX (658) 503-9689
 9830 South 51st St., Suite B-120, Phoenix, AZ 85044 (480) 785-0011 FAX (480) 785-0851
 2520 E. Sunset Rd. #3, Las Vegas, NV 89120 (702) 798-3620 FAX (702) 798-3621

MWH-Pasadena/Boeing
 300 North Lake Avenue, Suite 1200
 Pasadena, CA 91101
 Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOC2360

Sampled: 03/30/05
 Received: 03/30/05

DRAFT: PURGEABLES BY GC/MS (EPA 624)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOC2360-01 (DRAFT: Outfall 012 - Water)									
Reporting Units: ug/l									
1,2-Dibromoethane (EDB)	EPA 624	5C31014	0.32	2.0	ND	1	03/31/05	03/31/05	REV QUAL
Methyl-tert-butyl Ether (MTBE)	EPA 624	5C31014	0.32	5.0	ND	1	03/31/05	03/31/05	QUAL CODE
1,2,3-Trichloropropane	EPA 624	5C31014	0.85	10	ND	1	03/31/05	03/31/05	↓
Di-isopropyl Ether (DIPE)	EPA 624	5C31014	0.25	5.0	ND	1	03/31/05	03/31/05	↓
tert-Butanol (TBA)	EPA 624	5C31014	3.1	25	ND	1	03/31/05	03/31/05	↓
Surrogate: Dibromofluoromethane (80-120%)					111 %				
Surrogate: Toluene-d8 (80-120%)					100 %				
Surrogate: 4-Bromofluorobenzene (80-120%)					100 %				
Sample ID: IOC2360-02 (DRAFT: Trip Blank - Water)									
Reporting Units: ug/l									
1,2-Dibromoethane (EDB)	EPA 624	5C31014	0.32	2.0	ND	1	03/31/05	03/31/05	U
Methyl-tert-butyl Ether (MTBE)	EPA 624	5C31014	0.32	5.0	ND	1	03/31/05	03/31/05	↓
1,2,3-Trichloropropane	EPA 624	5C31014	0.85	10	ND	1	03/31/05	03/31/05	↓
Di-isopropyl Ether (DIPE)	EPA 624	5C31014	0.25	5.0	ND	1	03/31/05	03/31/05	↓
tert-Butanol (TBA)	EPA 624	5C31014	3.1	25	ND	1	03/31/05	03/31/05	↓
Surrogate: Dibromofluoromethane (80-120%)					109 %				
Surrogate: Toluene-d8 (80-120%)					100 %				
Surrogate: 4-Bromofluorobenzene (80-120%)					98 %				

AMEC VALIDATED

LEVEL IV

DRAFT REPORT
 DRAFT REPORT
 DATA SUBJECT TO CHANGE

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from Del Mar Analytical.

CONTRACT COMPLIANCE SCREENING FORM FOR HARDCOPY DATA

AMEC Earth & Environmental
 550 South Wadsworth Boulevard
 Suite 500
 Lakewood, CO 80226

Package ID T711WC133
 Task Order 313150010
 SDG No. IOC2360

No. of Analyses 1


Laboratory Del Mar Analytical

Reviewer L. Jarusewic

Analysis/Method General Minerals

Date: 04/08/05

Reviewer's Signature



ACTION ITEMS^a

1. Case Narrative Deficiencies
2. Out of Scope Analyses
3. Analyses Not Conducted
4. Missing Hardcopy Deliverables
5. Incorrect Hardcopy Deliverables
6. Deviations from Analysis Protocol, e.g.,
 - Holding Times
 - GC/MS Tune/Inst. Performance
 - Calibrations
 - Blanks
 - Surrogates
 - Matrix Spike/Dup LCS
 - Field QC
 - Internal Standard Performance
 - Compound Identification and Quantitation
 - System Performance

COMMENTS^b

Acceptable as reviewed.

^a Subcontracted analytical laboratory is not meeting contract and/or method requirements.

^b Differences in protocol have been adopted by the laboratory but no action against the laboratory is required.

Data Qualifier Reference Table

Qualifier	Organics	Inorganics
U	The analyte was analyzed for, but was not detected above the reported sample quantitation limit.	The material was analyzed for, but was not detected above the level of the associated value. The associated value is either the sample quantitation limit or the sample detection limit.
J	The analyte was positively identified; the associated numerical value is the approximate concentration of the analyte in the sample.	The associated value is an estimated quantity.
N	The analysis indicates the presence of an analyte for which there is presumptive evidence to make a "tentative identification."	Not applicable.
NJ	The analysis indicates the presence of an analyte that has been "tentatively identified" and the associated numerical value represents its approximate concentration.	Not applicable.
UJ	The analyte was not deemed above the reported sample quantitation limit. However, the reported quantitation limit is approximate and may or may not represent the actual limit of quantitation necessary to accurately and precisely measure the analyte in the sample.	The material was analyzed for, but was not detected. The associated value is an estimate and may be inaccurate or imprecise.
R	The sample results are rejected due to serious deficiencies in the ability to analyze the sample and to meet quality control criteria. The presence or absence of the analyte cannot be verified.	The data are unusable. (Note: Analyte may or may not be present).

Qualification Code Reference Table

Qualifier	Organics	Inorganics
H	Holding times were exceeded.	Holding times were exceeded.
S	Surrogate recovery was outside QC limits.	The sequence or number of standards used for the calibration was incorrect
C	Calibration %RSD or %D were noncompliant.	Correlation coefficient is <0.995.
R	Calibration RRF was <0.05.	%R for calibration is not within control limits.
B	Presumed contamination from preparation (method) blank.	Presumed contamination from preparation (method) or calibration blank.
L	Laboratory Blank Spike/Blank Spike Duplicate %R was not within control limits.	Laboratory Control Sample %R was not within control limits.
Q	MS/MSD recovery was poor or RPD high.	MS recovery was poor.
E	Not applicable.	Duplicates showed poor agreement.
I	Internal standard performance was unsatisfactory.	ICP ICS results were unsatisfactory.
A	Not applicable.	ICP Serial Dilution %D were not within control limits.
M	Tuning (BFB or DFTPP) was noncompliant.	Not applicable.
T	Presumed contamination from trip blank.	Not applicable.
+	False positive – reported compound was not present. Not applicable.	
-	False negative – compound was present but not reported.	Not applicable.
F	Presumed contamination from FB, or ER.	Presumed contamination from FB or ER.
\$	Reported result or other information was incorrect.	Reported result or other information was incorrect.
?	TIC identity or reported retention time has been changed.	Not applicable.
D	The analysis with this flag should not be used because another more technically sound analysis is available.	The analysis with this flag should not be used because another more technically sound analysis is available.
P	Instrument performance for pesticides was poor.	Post Digestion Spike recovery was not within control limits.
DNQ	The compound was detected between the MDL and the RL and, by definition, is considered an estimated value.	The compound was detected between the MDL and the RL and, by definition, is considered an estimated value.

*#

Unusual problems found with the data that have been described in Section 2.#, "Data Validation Findings." The number following the asterisk (*) will indicate the subsection where a description of the problem can be found (eg. *1 would indicate a sample was not within temperature limits).

Unusual problems found with the data that have been described in Section 2.#, "Data Validation Findings." The number following the asterisk (*) will indicate the subsection where a description of the problem can be found (eg. *1 would indicate a sample was not within temperature limits).



DATA VALIDATION REPORT

NPDES Monitoring

ANALYSIS: GENERAL MINERALS

SAMPLE DELIVERY GROUP: IOC2360

Prepared by

AMEC—Denver Operations
550 South Wadsworth Boulevard, Suite 500
Lakewood, Colorado 80226

1. INTRODUCTION

Task Order Title: NPDES Monitoring
Contract Task Order #: 313150010
Sample Delivery Group #: IOC2360
Project Manager: B. McIlvaine
Matrix: Water
Analysis: General Minerals
QC Level: Level IV
No. of Samples: 1
Reviewer: L. Jarusewic
Date of Review: April 8, 2005

The samples listed in Table 1 was validated based on the guidelines outlined in the AMEC *Data Validation Procedures SOP DVP-6, Rev. 2, USEPA Methods for Chemical Analysis of Water and Wastes Method 350.2, 405.1, 413.1, 418.1, 160.2, 160.5, and 180.1, Standard Methods for the Examination of Water and Wastewater Method SM2540C*, and validation guidelines outlined in the USEPA *Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (2/94)*. Any deviations from these procedures and guidelines are documented herein. Qualifiers were applied in cases where the data did not meet the required QC criteria or where special consideration by the data user is required. Data qualifiers were placed on Form Is with the associated qualification codes. Analytes that were rejected for any reason are denoted on the Form I as having only the "R" data qualifier and associated qualification code(s) denoting the reason for rejection. Any additional problems with the data that may have resulted in an estimated value were not denoted by a qualification code since the data had already been rejected.

Table 1. Sample identification

Client ID	EPA ID	Laboratory ID	Matrix	COC Method
Outfall 012	Outfall 012	IOC2360-01	Water	General Minerals

2. DATA VALIDATION FINDINGS

2.1 SAMPLE MANAGEMENT

Following are findings associated with sample management:

2.1.1 Sample Preservation, Handling, and Transport

The sample in this SDG was received at the laboratory within the temperature limits of $4^{\circ}\text{C} \pm 2^{\circ}\text{C}$. No preservation problems were noted by the laboratory. No qualifications were required.

2.1.2 Chain of Custody

The COC was signed and dated by field and laboratory personnel. The COC accounted for all analyses present in this SDG. No sample qualifications were required.

2.1.3 Holding Times

The holding times were assessed by comparing the date of collection with the dates of analyses. The 28-day analytical holding time for ammonia, total recoverable hydrocarbons, and oil and grease, the seven-day holding time for total suspended solids and total dissolved solids, and the 48-hour holding time for turbidity, biological oxygen demand, and total settleable solids holding times were met. No qualifications were required.

2.2 CALIBRATION

For the applicable analyses, the initial calibration correlation coefficients were ≥ 0.995 . The initial and continuing calibration information was acceptable with recoveries within the control limits of 90-110%. For ammonia, no information regarding the standardization of the titrant was provided; however, as the LCS recovery was within the CCV control limits, no qualifications were required. For BOD, no information regarding the calibration of the oxygen meter was provided; however, as the LCS recovery was within the CCV control limits, no qualifications were required. Calibration is not applicable to oil and grease, total dissolved solids, total suspended solids, or total settleable solids. No qualifications were required.

2.3 BLANKS

The method blank and CCB results reported on the summary forms and in the raw data for blank analyses associated with the sample were nondetects at the reporting limit. No qualifications were required.

2.4 BLANK SPIKES AND LABORATORY CONTROL SAMPLES

The laboratory control sample and laboratory control sample duplicate (BOD, total recoverable hydrocarbons, and oil and grease only) recoveries and RPDs were within the laboratory-established control

limits. The LCS is not applicable to turbidity, conductivity, or settleable solids. No qualifications were required.

2.5 SURROGATES RECOVERY

Surrogate recovery is not applicable to the analyses presented in this SDG.

2.6 LABORATORY DUPLICATES

No MS/MSD analyses were performed in association with the sample in this SDG; therefore no assessment was made with respect to this criterion.

2.7 MATRIX SPIKE/MATRIX SPIKE DUPLICATE

No MS/MSD analyses were performed in association with the sample in this SDG; therefore, no assessment was made with respect to this criterion. Method accuracy was assessed based on LCS results.

2.8 FURNACE ATOMIC ABSORPTION QC

Furnace atomic absorption was not utilized for the analyses of this sample; therefore, furnace atomic absorption QC is not applicable.

2.9 ICP SERIAL DILUTION

ICP serial dilution is not applicable to the analyses presented in this data validation report.

2.10 SAMPLE RESULT VERIFICATION

A Level IV review was performed for the sample in this data package. Calculations were verified, and the sample results reported on the Form Is were verified against the raw data. The oil and grease blank value was not subtracted from the site sample raw data as required by the method. The reviewer corrected the raw data and the changed the results on the sample Form I. The data was qualified as an estimated detect ("J"), between the MDL and the reporting limit with a "DNQ" qualification code. No further transcription errors or calculation errors were noted. No further qualifications were required.

2.11 FIELD QC SAMPLES

Field QC samples are evaluated, and if necessary, qualified based only on laboratory blanks. Any remaining detects are used to evaluate the associated sample. The following are findings associated with field QC samples:

2.11.1 Field Blanks and Equipment Rinsates

The sample in this SDG had no associated field QC samples. No qualifications were required.

2.11.2 Field Duplicates

There were no field duplicate pairs associated with this SDG.



Del Mar Analytical

17461 Denian Ave., Suite 100, Irvine, CA 92614 (949) 261-1022 FAX (949) 260-1297
 1014 E. Cooley Dr., Suite A, Colton, CA 92324 (909) 370-4607 FAX (949) 370-1046
 9484 Chesapeake Dr., Suite 805, San Diego, CA 92123 (619) 505-8596 FAX (619) 505-0486
 9800 South 51st St., Suite B-120, Phoenix, AZ 85044 (480) 785-0043 FAX (480) 785-0053
 2520 E. Sunset Rd. #3, Las Vegas, NV 89120 (702) 798-3620 FAX (702) 798-3623

MWH-Pasadena/Boeing
 300 North Lake Avenue, Suite 1200
 Pasadena, CA 91101
 Attention: Bronwyn Kelly

- Project ID: Alfa Outfall 012 - During Test

Report Number: IOC2360

Sampled: 03/30/05
 Received: 03/30/05

DRAFT: TOTAL RECOVERABLE PETROLEUM HYDROCARBONS (EPA 418.1)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOC2360-01 (DRAFT: Outfall 012 - Water)									
Reporting Units: mg/l									
Total Recoverable Hydrocarbons	EPA 418.1	5C31088	0.31	1.0	4.1	1	03/31/05	03/31/05	REV QUAL TOTAL CODE

AMEC VALIDATED

LEVEL IV

DRAFT REPORT
 DRAFT REPORT
 DATA SUBJECT TO CHANGE

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from Del Mar Analytical.



Del Mar Analytical

17601 Derian Ave., Suite 100, Irvine, CA 92614 (949) 261-5021 FAX (949) 261-1207
 1014 E. Cooley Dr., Suite A, Colton, CA 92324 (909) 371-4607 FAX (949) 370-1046
 5404 Chatsapeake Dr., Suite 805, San Diego, CA 92123 (619) 515-8596 FAX (619) 505-0689
 1870 South 51st St., Suite B-120, Phoenix, AZ 85044 (480) 782-0041 FAX (480) 785-0871
 2320 E. Sunset Rd., #3, Las Vegas, NV 89120 (702) 798-1020 FAX (702) 798-1021

MWH-Pasadena/Boeing
 300 North Lake Avenue, Suite 1200
 Pasadena, CA 91101
 Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOC2360

Sampled: 03/30/05
 Received: 03/30/05

DRAFT: INORGANICS

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOC2360-01 (DRAFT: Outfall 012 - Water) - cont. Reporting Units: mg/l									
Ammonia-N (Distilled)	EPA 350.2	5C31085	0.30	0.50	0.56	1	03/31/05	03/31/05	
Biochemical Oxygen Demand	EPA 405.1	5C31066	0.59	2.0	4.2	1	03/31/05	04/05/05	
Oil & Grease	EPA 413.1	5C31087	0.94	5.0	5.8	1	03/31/05	03/31/05	J
Total Dissolved Solids	SM2540C	5C30093	10	10	180	1	03/30/05	03/30/05	
Total Suspended Solids	EPA 160.2	5C30091	10	10	14	1	03/30/05	03/30/05	
Sample ID: IOC2360-01 (DRAFT: Outfall 012 - Water) Reporting Units: ml/l/hr									
Total Settleable Solids	EPA 160.5	5C31067	0.10	0.10	ND	1	03/31/05	03/31/05	U
Sample ID: IOC2360-01 (DRAFT: Outfall 012 - Water) Reporting Units: NTU									
Turbidity	EPA 180.1	5D01075	0.040	1.0	20	1	04/01/05	04/01/05	
Sample ID: IOC2360-01 (DRAFT: Outfall 012 - Water) Reporting Units: ug/l									
Perchlorate	EPA 314.0	5C31059	0.80	4.0	ND	1	03/31/05	03/31/05	*

REV
 OUTL
 CODE

B#, DV

AMEC VALIDATED

LEVEL IV

*Analysis Not Validated

DRAFT REPORT
 DRAFT REPORT
 DATA SUBJECT TO CHANGE

EAW
 5/12/05


The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from Del Mar Analytical.

CONTRACT COMPLIANCE SCREENING FORM FOR HARDCOPY DATA

AMEC Earth & Environmental
 550 South Wadsworth Boulevard
 Suite 500
 Lakewood, CO 80226

Package ID T711WC134
 Task Order 313150010
 SDG No. IOC2360
 No. of Analyses 1

Laboratory Del Mar Analytical
 Reviewer L. Jarusewic
 Analysis/Method Perchlorate

Date: 04/08/05
 Reviewer's Signature


ACTION ITEMS^a

1. **Case Narrative Deficiencies**
2. **Out of Scope Analyses**
3. **Analyses Not Conducted**
4. **Missing Hardcopy Deliverables**
5. **Incorrect Hardcopy Deliverables**
6. **Deviations from Analysis Protocol, e.g.,**
 - Holding Times
 - GC/MS Tune/Inst. Performance
 - Calibrations
 - Blanks
 - Surrogates
 - Matrix Spike/Dup LCS
 - Field QC
 - Internal Standard Performance
 - Compound Identification and Quantitation
 - System Performance

COMMENTS^b Acceptable as reviewed.

^a Subcontracted analytical laboratory is not meeting contract and/or method requirements.
^b Differences in protocol have been adopted by the laboratory but no action against the laboratory is required.

Data Qualifier Reference Table

Qualifier	Organics	Inorganics
U	The analyte was analyzed for, but was not detected above the reported sample quantitation limit.	The material was analyzed for, but was not detected above the level of the associated value. The associated value is either the sample quantitation limit or the sample detection limit.
J	The analyte was positively identified; the associated numerical value is the approximate concentration of the analyte in the sample.	The associated value is an estimated quantity.
N	The analysis indicates the presence of an analyte for which there is presumptive evidence to make a "tentative identification."	Not applicable.
NJ	The analysis indicates the presence of an analyte that has been "tentatively identified" and the associated numerical value represents its approximate concentration.	Not applicable.
UJ	The analyte was not deemed above the reported sample quantitation limit. However, the reported quantitation limit is approximate and may or may not represent the actual limit of quantitation necessary to accurately and precisely measure the analyte in the sample.	The material was analyzed for, but was not detected. The associated value is an estimate and may be inaccurate or imprecise.
R	The sample results are rejected due to serious deficiencies in the ability to analyze the sample and to meet quality control criteria. The presence or absence of the analyte cannot be verified.	The data are unusable. (Note: Analyte may or may not be present).

Qualification Code Reference Table

Qualifier	Organics	Inorganics
H	Holding times were exceeded.	Holding times were exceeded.
S	Surrogate recovery was outside QC limits.	The sequence or number of standards used for the calibration was incorrect
C	Calibration %RSD or %D were noncompliant.	Correlation coefficient is <0.995.
R	Calibration RRF was <0.05.	%R for calibration is not within control limits.
B	Presumed contamination from preparation (method) blank.	Presumed contamination from preparation (method) or calibration blank.
L	Laboratory Blank Spike/Blank Spike Duplicate %R was not within control limits.	Laboratory Control Sample %R was not within control limits.
Q	MS/MSD recovery was poor or RPD high.	MS recovery was poor.
E	Not applicable.	Duplicates showed poor agreement.
I	Internal standard performance was unsatisfactory.	ICP ICS results were unsatisfactory.
A	Not applicable.	ICP Serial Dilution %D were not within control limits.
M	Tuning (BFB or DFTPP) was noncompliant.	Not applicable.
T	Presumed contamination from trip blank.	Not applicable.
+	False positive -- reported compound was not present. Not applicable.	
-	False negative -- compound was present but not reported.	Not applicable.
F	Presumed contamination from FB, or ER.	Presumed contamination from FB or ER.
\$	Reported result or other information was incorrect.	Reported result or other information was incorrect.
?	TIC identity or reported retention time has been changed.	Not applicable.
D	The analysis with this flag should not be used because another more technically sound analysis is available.	The analysis with this flag should not be used because another more technically sound analysis is available.
P	Instrument performance for pesticides was poor.	Post Digestion Spike recovery was not within control limits.
DNQ	The compound was detected between the MDL and the RL and, by definition, is considered an estimated value.	The compound was detected between the MDL and the RL and, by definition, is considered an estimated value.

*#

Unusual problems found with the data that have been described in Section 2.#, "Data Validation Findings." The number following the asterisk (*) will indicate the subsection where a description of the problem can be found (eg. *1 would indicate a sample was not within temperature limits).

Unusual problems found with the data that have been described in Section 2.#, "Data Validation Findings." The number following the asterisk (*) will indicate the subsection where a description of the problem can be found (eg. *1 would indicate a sample was not within temperature limits).



DATA VALIDATION REPORT

NPDES Monitoring

ANALYSIS: PERCHLORATE

SAMPLE DELIVERY GROUP: IOC2360

Prepared by

AMEC—Denver Operations
550 South Wadsworth Boulevard, Suite 500
Lakewood, Colorado 80226

1. INTRODUCTION

Task Order Title: NPDES Monitoring
Contract Task Order #: 313150010
Sample Delivery Group #: IOC2360
Project Manager: B. McIlvaine
Matrix: Water
Analysis: Perchlorate
QC Level: Level IV
No. of Samples: 1
Reviewer: L. Jarusewic
Date of Review: April 8, 2005

The samples listed in Table 1 was validated based on the guidelines outlined in the AMEC *Data Validation Procedures SOP DVP-6, Rev. 2, USEPA Methods for Chemical Analysis of Water and Wastes Method 314.0*, and validation guidelines outlined in the USEPA *Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (2/94)*. Any deviations from these procedures and guidelines are documented herein. Qualifiers were applied in cases where the data did not meet the required QC criteria or where special consideration by the data user is required. Data qualifiers were placed on Form Is with the associated qualification codes. Analytes that were rejected for any reason are denoted on the Form I as having only the "R" data qualifier and associated qualification code(s) denoting the reason for rejection. Any additional problems with the data that may have resulted in an estimated value were not denoted by a qualification code since the data had already been rejected.

Table 1. Sample identification

Client ID	EPA ID	Laboratory ID	Matrix	COC Method
Outfall 012	Outfall 012	IOC2360-01	Water	Perchlorate

2. DATA VALIDATION FINDINGS

2.1 SAMPLE MANAGEMENT

Following are findings associated with sample management:

2.1.1 Sample Preservation, Handling, and Transport

The sample in this SDG was received at the laboratory within the temperature limits of $4^{\circ}\text{C} \pm 2^{\circ}\text{C}$. The analysis did not require preservation and no preservation was noted in the field. No qualifications were required.

2.1.2 Chain of Custody

The COC was signed and dated by field and laboratory personnel, and accounted for the sample and analysis presented in this SDG. No qualifications were required.

2.1.3 Holding Times

The holding time was assessed by comparing the date of collection with the date of analysis. The 28-day analytical holding time for perchlorate was met, and no qualifications were required.

2.2 CALIBRATION

The initial calibration correlation coefficient was ≥ 0.995 . The IPC-MA recovery was within the control limits of 80-120%. The ICV, CCV, ICCS, and IPC recoveries were within the control limits of 90-110%. No qualifications were required.

2.3 BLANKS

The method blank and CCB results reported on the summary forms and in the raw data for blank analyses associated with the sample were nondetects at the reporting limit. No qualifications were required.

2.4 BLANK SPIKES AND LABORATORY CONTROL SAMPLES

The laboratory control sample recovery was within the method control limits of 85-115%. No qualifications were required.

2.5 SURROGATES RECOVERY

Surrogate recovery is not applicable to the analysis presented in this SDG.

2.6 LABORATORY DUPLICATES

No MS/MSD or duplicate analyses were performed in association with the sample in this SDG; therefore, no assessment was made with respect to this criterion.

2.7 MATRIX SPIKE/MATRIX SPIKE DUPLICATE

No MS/MSD analyses were performed in association with the sample in this SDG; therefore, no assessment was made with respect to this criterion. Method accuracy was assessed based on LCS results.

2.8 FURNACE ATOMIC ABSORPTION QC

Furnace atomic absorption was not utilized for the analysis of this sample; therefore, furnace atomic absorption QC is not applicable.

2.9 ICP SERIAL DILUTION

ICP serial dilution is not applicable to the analysis presented in this data validation report.

2.10 SAMPLE RESULT VERIFICATION

A Level IV review was performed for the sample in this data package. Calculations were verified, and the sample result reported on the Form I was verified against the raw data. No transcription errors or calculation errors were noted. No qualifications were required.

2.11 FIELD QC SAMPLES

Field QC samples are evaluated, and if necessary, qualified based only on laboratory blanks. Any remaining detects are used to evaluate the associated sample. The following are findings associated with field QC samples:

2.11.1 Field Blanks and Equipment Rinsates

The sample in this SDG had no associated field QC samples. No qualifications were required.

2.11.2 Field Duplicates

There were no field duplicate pairs associated with this SDG.



Del Mar Analytical

17601 Doran Ave., Suite 100, Irvine, CA 92614 (949) 261-7100 FAX (949) 260-1200
 1614 E. Conley Dr., Suite A, Colton, CA 92314 (909) 370-4667 FAX (909) 370-1040
 9404 Chesapeake Dr., Suite 805, San Diego, CA 92123 (619) 505-8500 FAX (619) 505-0685
 9810 South 51st St., Suite B-120, Phoenix, AZ 85044 (480) 783-0043 FAX (480) 783-0811
 2520 E. Sunset Rd. #3, Las Vegas, NV 89120 (702) 798-1620 FAX (702) 798-3621

MWH-Pasadena/Boeing
 300 North Lake Avenue, Suite 1200
 Pasadena, CA 91101
 Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOC2360

Sampled: 03/30/05
 Received: 03/30/05

DRAFT: INORGANICS

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOC2360-01 (DRAFT: Outfall 012 - Water) - cont.									
Reporting Units: mg/l									
Ammonia-N (Distilled)	EPA 350.2	5C31085	0.30	0.50	0.56	1	03/31/05	03/31/05	* 3
Biochemical Oxygen Demand	EPA 405.1	5C31066	0.59	2.0	4.2	1	03/31/05	04/05/05	
Oil & Grease	EPA 413.1	5C31087	0.94	5.0	5.8	1	03/31/05	03/31/05	
Total Dissolved Solids	SM2540C	5C30093	10	10	180	1	03/30/05	03/30/05	
Total Suspended Solids	EPA 160.2	5C30091	10	10	14	1	03/30/05	03/30/05	
Sample ID: IOC2360-01 (DRAFT: Outfall 012 - Water)									
Reporting Units: ml/hr									
Total Settleable Solids	EPA 160.5	5C31067	0.10	0.10	ND	1	03/31/05	03/31/05	
Sample ID: IOC2360-01 (DRAFT: Outfall 012 - Water)									
Reporting Units: NTU									
Turbidity	EPA 180.1	SD01075	0.040	1.0	20	1	04/01/05	04/01/05	
Sample ID: IOC2360-01 (DRAFT: Outfall 012 - Water)									
Reporting Units: ug/l									
Perchlorate	EPA 314.0	5C31059	0.80	4.0	ND	1	03/31/05	03/31/05	u

AMEC VALIDATED

LEVEL IV

Analysis Not Validated

DRAFT REPORT
 DRAFT REPORT
 DATA SUBJECT TO CHANGE

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from Del Mar Analytical.



LABORATORY REPORT

Prepared For: MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project: Alfa Outfall 012 - During Test

Sampled: 04/04/05
Received: 04/05/05
Issued: 05/18/05 11:44

NELAP #01108CA California ELAP#1197 CSDLAC #10117

*The results listed within this Laboratory Report pertain only to the samples tested in the laboratory. The analyses contained in this report were performed in accordance with the applicable certifications as noted. All soil samples are reported on a wet weight basis unless otherwise noted in the report. This Laboratory Report is confidential and is intended for the sole use of Del Mar Analytical and its client. This report shall not be reproduced, except in full, without written permission from Del Mar Analytical. The Chain(s) of Custody, 2 pages, are included and are an integral part of this report.
This entire report was reviewed and approved for release.*

CASE NARRATIVE

- SAMPLE RECEIPT: Samples were received intact, at 4°C, on ice and with chain of custody documentation.
- HOLDING TIMES: Not all holding times were met. Results were qualified where the sample analysis did not occur within method specified holding time requirements. The Turbidity analysis was performed one day past the method specified hold time due to laboratory oversight.
- PRESERVATION: Samples requiring preservation were verified prior to sample analysis.
- QA/QC CRITERIA: All analyses met method criteria, except as noted in the report with data qualifiers.
- COMMENTS: Results that fall between the MDL and RL are 'J' flagged.
- SUBCONTRACTED: Refer to the last page for specific subcontract laboratory information included in this report.

LABORATORY ID	CLIENT ID	MATRIX
IOD0229-01	Outfall 012	Water
IOD0229-02	Trip Blank	Water

Reviewed By:

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS (EPA 418.1)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD0229-01 (Outfall 012 - Water)					Sampled: 04/04/05				
Reporting Units: mg/l									
Total Recoverable Hydrocarbons	EPA 418.1	5D07088	0.31	1.0	7.7	1	04/07/05	04/07/05	

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

EXTRACTABLE FUEL HYDROCARBONS (CADHS/8015 Modified)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD0229-01 (Outfall 012 - Water) - cont.					Sampled: 04/04/05				
Reporting Units: mg/l									
EFH (C13 - C22)	EPA 8015B	5D08084	0.082	0.50	0.81	0.962	04/08/05	04/09/05	
<i>Surrogate: n-Octacosane (40-125%)</i>					<i>80 %</i>				

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

VOLATILE FUEL HYDROCARBONS (EPA 5030/CADHS Mod. 8015)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD0229-01 (Outfall 012 - Water) - cont.					Sampled: 04/04/05				
Reporting Units: mg/l									
GRO (C4 - C12)	EPA 8015 Mod.	5D13113	0.25	0.50	1.2	5	04/13/05	04/13/05	
Surrogate: 4-BFB (FID) (65-140%)					92 %				
Sample ID: IOD0229-02 (Trip Blank - Water)					Sampled: 04/04/05				
Reporting Units: mg/l									
GRO (C4 - C12)	EPA 8015 Mod.	5D12097	0.050	0.10	ND	1	04/12/05	04/12/05	
Surrogate: 4-BFB (FID) (65-140%)					92 %				

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

PURGEABLES BY GC/MS (EPA 624)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD0229-01 (Outfall 012 - Water)					Sampled: 04/04/05				
Reporting Units: ug/l									
1,2-Dibromoethane (EDB)	EPA 624	5D10007	0.32	2.0	ND	1	04/10/05	04/10/05	
Methyl-tert-butyl Ether (MTBE)	EPA 624	5D10007	0.32	5.0	ND	1	04/10/05	04/10/05	
1,2,3-Trichloropropane	EPA 624	5D10007	0.85	10	ND	1	04/10/05	04/10/05	
Di-isopropyl Ether (DIPE)	EPA 624	5D10007	0.25	5.0	ND	1	04/10/05	04/10/05	
tert-Butanol (TBA)	EPA 624	5D10007	3.1	25	ND	1	04/10/05	04/10/05	
Surrogate: Dibromofluoromethane (80-120%)					103 %				
Surrogate: Toluene-d8 (80-120%)					102 %				
Surrogate: 4-Bromofluorobenzene (80-120%)					104 %				
Sample ID: IOD0229-02 (Trip Blank - Water)					Sampled: 04/04/05				
Reporting Units: ug/l									
1,2-Dibromoethane (EDB)	EPA 624	5D10007	0.32	2.0	ND	1	04/10/05	04/10/05	
Methyl-tert-butyl Ether (MTBE)	EPA 624	5D10007	0.32	5.0	ND	1	04/10/05	04/10/05	
1,2,3-Trichloropropane	EPA 624	5D10007	0.85	10	ND	1	04/10/05	04/10/05	
Di-isopropyl Ether (DIPE)	EPA 624	5D10007	0.25	5.0	ND	1	04/10/05	04/10/05	
tert-Butanol (TBA)	EPA 624	5D10007	3.1	25	ND	1	04/10/05	04/10/05	
Surrogate: Dibromofluoromethane (80-120%)					109 %				
Surrogate: Toluene-d8 (80-120%)					111 %				
Surrogate: 4-Bromofluorobenzene (80-120%)					107 %				

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

ACID & BASE/NEUTRALS BY GC/MS (EPA 625)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD0229-01 (Outfall 012 - Water)					Sampled: 04/04/05				
Reporting Units: ug/l									
Naphthalene	EPA 625	5D06042	4.5	10	22	0.962	04/06/05	04/14/05	
N-Nitrosodimethylamine	EPA 625	5D06042	3.7	20	ND	0.962	04/06/05	04/14/05	
Surrogate: 2-Fluorophenol (30-120%)					60 %				
Surrogate: Phenol-d6 (35-120%)					66 %				
Surrogate: 2,4,6-Tribromophenol (45-120%)					84 %				
Surrogate: Nitrobenzene-d5 (45-120%)					83 %				
Surrogate: 2-Fluorobiphenyl (45-120%)					72 %				
Surrogate: Terphenyl-d14 (45-120%)					92 %				

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

INORGANICS

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD0229-01 (Outfall 012 - Water) - cont.					Sampled: 04/04/05				
Reporting Units: mg/l									
Ammonia-N (Distilled)	EPA 350.2	5D08083	0.30	0.50	ND	1	04/08/05	04/08/05	
Biochemical Oxygen Demand	EPA 405.1	5D06055	0.59	2.0	3.6	1	04/06/05	04/11/05	
Oil & Grease	EPA 413.1	5D11073	0.94	5.0	5.5	1	04/11/05	04/11/05	
Total Dissolved Solids	SM2540C	5D07121	10	10	220	1	04/07/05	04/07/05	
Total Suspended Solids	EPA 160.2	5D07079	10	10	20	1	04/07/05	04/07/05	
Sample ID: IOD0229-01 (Outfall 012 - Water)					Sampled: 04/04/05				
Reporting Units: ml/l/hr									
Total Settleable Solids	EPA 160.5	5D06057	0.10	0.10	ND	1	04/06/05	04/06/05	
Sample ID: IOD0229-01 (Outfall 012 - Water)					Sampled: 04/04/05				
Reporting Units: NTU									
Turbidity	EPA 180.1	5D07071	0.040	1.0	32	1	04/07/05	04/07/05	H
Sample ID: IOD0229-01 (Outfall 012 - Water)					Sampled: 04/04/05				
Reporting Units: ug/l									
Perchlorate	EPA 314.0	5D06050	0.80	4.0	ND	1	04/06/05	04/06/05	

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

1,4-DIOXANE BY GC/MS (EPA 5030B/8260B)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD0229-01 (Outfall 012 - Water) - cont.					Sampled: 04/04/05				
Reporting Units: ug/l									
1,4-Dioxane	EPA 8260B	P5D1317	0.49	1.0	ND	1	04/13/05	04/13/05	
<i>Surrogate: Dibromofluoromethane (80-125%)</i>					109 %				

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05

Received: 04/05/05

SHORT HOLD TIME DETAIL REPORT

Sample ID: Outfall 012 (IOD0229-01) - Water	Hold Time (in days)	Date/Time Sampled	Date/Time Received	Date/Time Extracted	Date/Time Analyzed
EPA 160.5	2	04/04/2005 15:30	04/05/2005 17:00	04/06/2005 09:51	04/06/2005 12:00
EPA 180.1	2	04/04/2005 15:30	04/05/2005 17:00	04/07/2005 08:30	04/07/2005 09:45
EPA 405.1	2	04/04/2005 15:30	04/05/2005 17:00	04/06/2005 18:00	04/11/2005 10:00

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

METHOD BLANK/QC DATA

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS (EPA 418.1)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC %REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5D07088 Extracted: 04/07/05											
Blank Analyzed: 04/07/2005 (5D07088-BLK1)											
Total Recoverable Hydrocarbons	ND	1.0	0.31	mg/l							
LCS Analyzed: 04/07/2005 (5D07088-BS1)											
Total Recoverable Hydrocarbons	4.89	1.0	0.31	mg/l	5.00		98	65-120			M-NR1
LCS Dup Analyzed: 04/07/2005 (5D07088-BSD1)											
Total Recoverable Hydrocarbons	5.23	1.0	0.31	mg/l	5.00		105	65-120	7	20	

Del Mar Analytical, Irvine
Michele Harper
Project Manager

MWH-Pasadena/Boeing
 300 North Lake Avenue, Suite 1200
 Pasadena, CA 91101
 Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
 Received: 04/05/05

METHOD BLANK/QC DATA

EXTRACTABLE FUEL HYDROCARBONS (CADHS/8015 Modified)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC %REC	Limit	RPD	RPD Limit	Data Qualifiers
Batch: 5D08084 Extracted: 04/08/05											
Blank Analyzed: 04/08/2005 (5D08084-BLK1)											
EFH (C13 - C22)	ND	0.50	0.082	mg/l							
EFH (C13 - C40)	ND	0.50	0.082	mg/l							
Surrogate: n-Octacosane	0.184			mg/l	0.200		92	40-125			
LCS Analyzed: 04/08/2005 (5D08084-BS1)											
EFH (C13 - C40)	0.598	0.50	0.082	mg/l	0.775		77	40-120			M-NRI
Surrogate: n-Octacosane	0.176			mg/l	0.200		88	40-125			
LCS Dup Analyzed: 04/08/2005 (5D08084-BSD1)											
EFH (C13 - C40)	0.619	0.50	0.082	mg/l	0.775		80	40-120	3	25	
Surrogate: n-Octacosane	0.184			mg/l	0.200		92	40-125			

Del Mar Analytical, Irvine
 Michele Harper
 Project Manager

MWH-Pasadena/Boeing
 300 North Lake Avenue, Suite 1200
 Pasadena, CA 91101
 Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05

Received: 04/05/05

METHOD BLANK/QC DATA

VOLATILE FUEL HYDROCARBONS (EPA 5030/CADHS Mod. 8015)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5D12097 Extracted: 04/12/05											
Blank Analyzed: 04/12/2005 (5D12097-BLK1)											
GRO (C4 - C12)	ND	0.10	0.050	mg/l							
Surrogate: 4-BFB (FID)	0.00846			mg/l	0.0100		85	65-140			
LCS Analyzed: 04/12/2005 (5D12097-BS1)											
GRO (C4 - C12)	0.616	0.10	0.050	mg/l	0.800		77	70-140			
Surrogate: 4-BFB (FID)	0.0328			mg/l	0.0300		109	65-140			
Matrix Spike Analyzed: 04/12/2005 (5D12097-MS1)						Source: IOD0502-10					
GRO (C4 - C12)	0.233	0.10	0.050	mg/l	0.220	ND	106	60-140			
Surrogate: 4-BFB (FID)	0.0109			mg/l	0.0100		109	65-140			
Matrix Spike Dup Analyzed: 04/12/2005 (5D12097-MSD1)						Source: IOD0502-10					
GRO (C4 - C12)	0.221	0.10	0.050	mg/l	0.220	ND	100	60-140	5	20	
Surrogate: 4-BFB (FID)	0.0104			mg/l	0.0100		104	65-140			
Batch: 5D13113 Extracted: 04/13/05											
Blank Analyzed: 04/13/2005 (5D13113-BLK1)											
GRO (C4 - C12)	ND	0.10	0.050	mg/l							
Surrogate: 4-BFB (FID)	0.0103			mg/l	0.0100		103	65-140			
LCS Analyzed: 04/13/2005 (5D13113-BS1)											
GRO (C4 - C12)	0.774	0.10	0.050	mg/l	0.800		97	70-140			
Surrogate: 4-BFB (FID)	0.0288			mg/l	0.0300		96	65-140			
Matrix Spike Analyzed: 04/13/2005 (5D13113-MS1)						Source: IOD0422-01					
GRO (C4 - C12)	0.230	0.10	0.050	mg/l	0.220	ND	105	60-140			
Surrogate: 4-BFB (FID)	0.0124			mg/l	0.0100		124	65-140			

Del Mar Analytical, Irvine
 Michele Harper
 Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

METHOD BLANK/QC DATA

VOLATILE FUEL HYDROCARBONS (EPA 5030/CADHS Mod. 8015)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5D13113 Extracted: 04/13/05											
Matrix Spike Dup Analyzed: 04/13/2005 (5D13113-MSD1)						Source: IOD0422-01					
GRO (C4 - C12)	0.222	0.10	0.050	mg/l	0.220	ND	101	60-140	4	20	
Surrogate: 4-BFB (FID)	0.0113			mg/l	0.0100		113	65-140			

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

METHOD BLANK/QC DATA

PURGEABLES BY GC/MS (EPA 624)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC %REC	Limit	RPD RPD	Limit	Data Qualifiers
Batch: 5D10007 Extracted: 04/10/05											
Blank Analyzed: 04/10/2005 (5D10007-BLK1)											
1,2-Dibromoethane (EDB)	ND	2.0	0.32	ug/l							
Methyl-tert-butyl Ether (MTBE)	ND	5.0	0.32	ug/l							
1,2,3-Trichloropropane	ND	10	0.85	ug/l							
Di-isopropyl Ether (DIPE)	ND	5.0	0.25	ug/l							
tert-Butanol (TBA)	ND	25	3.1	ug/l							
Surrogate: Dibromofluoromethane	27.6			ug/l	25.0		110	80-120			
Surrogate: Toluene-d8	26.2			ug/l	25.0		105	80-120			
Surrogate: 4-Bromofluorobenzene	26.0			ug/l	25.0		104	80-120			
LCS Analyzed: 04/10/2005 (5D10007-BS1)											
1,2-Dibromoethane (EDB)	29.0	2.0	0.32	ug/l	25.0		116	75-125			
Methyl-tert-butyl Ether (MTBE)	29.0	5.0	0.32	ug/l	25.0		116	55-145			
1,2,3-Trichloropropane	27.8	10	0.85	ug/l	25.0		111	60-130			
Di-isopropyl Ether (DIPE)	33.8	5.0	0.25	ug/l	25.0		135	65-135			
tert-Butanol (TBA)	129	25	3.1	ug/l	125		103	70-140			
Surrogate: Dibromofluoromethane	27.4			ug/l	25.0		110	80-120			
Surrogate: Toluene-d8	27.2			ug/l	25.0		109	80-120			
Surrogate: 4-Bromofluorobenzene	26.6			ug/l	25.0		106	80-120			
Matrix Spike Analyzed: 04/10/2005 (5D10007-MS1) Source: IOD0511-02											
1,2-Dibromoethane (EDB)	26.6	2.0	0.32	ug/l	25.0	ND	106	70-130			
Methyl-tert-butyl Ether (MTBE)	26.4	5.0	0.32	ug/l	25.0	ND	106	50-155			
1,2,3-Trichloropropane	26.5	10	0.85	ug/l	25.0	ND	106	55-140			
Di-isopropyl Ether (DIPE)	30.6	5.0	0.25	ug/l	25.0	ND	122	65-140			
tert-Butanol (TBA)	142	25	3.1	ug/l	125	ND	114	65-145			
Surrogate: Dibromofluoromethane	26.5			ug/l	25.0		106	80-120			
Surrogate: Toluene-d8	26.9			ug/l	25.0		108	80-120			
Surrogate: 4-Bromofluorobenzene	25.4			ug/l	25.0		102	80-120			

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

METHOD BLANK/QC DATA

PURGEABLES BY GC/MS (EPA 624)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5D10007 Extracted: 04/10/05											
Matrix Spike Dup Analyzed: 04/10/2005 (5D10007-MSD1)						Source: IOD0511-02					
1,2-Dibromoethane (EDB)	25.2	2.0	0.32	ug/l	25.0	ND	101	70-130	5	25	
Methyl-tert-butyl Ether (MTBE)	26.1	5.0	0.32	ug/l	25.0	ND	104	50-155	1	25	
1,2,3-Trichloropropane	25.4	10	0.85	ug/l	25.0	ND	102	55-140	4	30	
Di-isopropyl Ether (DIPE)	28.9	5.0	0.25	ug/l	25.0	ND	116	65-140	6	25	
tert-Butanol (TBA)	118	25	3.1	ug/l	125	ND	94	65-145	18	25	
Surrogate: Dibromofluoromethane	26.3			ug/l	25.0		105	80-120			
Surrogate: Toluene-d8	27.2			ug/l	25.0		109	80-120			
Surrogate: 4-Bromofluorobenzene	26.0			ug/l	25.0		104	80-120			

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

METHOD BLANK/QC DATA

ACID & BASE/NEUTRALS BY GC/MS (EPA 625)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC %REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
---------	--------	-----------------	-----	-------	-------------	---------------	-----------	-------------	-----	-----------	-----------------

Batch: 5D06042 Extracted: 04/06/05

Blank Analyzed: 04/14/2005 (5D06042-BLK1)

Naphthalene	ND	10	4.5	ug/l							
N-Nitrosodimethylamine	ND	20	3.7	ug/l							
Surrogate: 2-Fluorophenol	82.3			ug/l	200		41	30-120			
Surrogate: Phenol-d6	103			ug/l	200		52	35-120			
Surrogate: 2,4,6-Tribromophenol	142			ug/l	200		71	45-120			
Surrogate: Nitrobenzene-d5	60.3			ug/l	100		60	45-120			
Surrogate: 2-Fluorobiphenyl	60.3			ug/l	100		60	45-120			
Surrogate: Terphenyl-d14	101			ug/l	100		101	45-120			

LCS Analyzed: 04/13/2005 (5D06042-BS1)

Naphthalene	116	10	4.5	ug/l	100		116	50-120			M-NR1
N-Nitrosodimethylamine	84.0	20	3.7	ug/l	100		84	40-120			
Surrogate: 2-Fluorophenol	116			ug/l	200		58	30-120			
Surrogate: Phenol-d6	140			ug/l	200		70	35-120			
Surrogate: 2,4,6-Tribromophenol	183			ug/l	200		92	45-120			
Surrogate: Nitrobenzene-d5	74.5			ug/l	100		74	45-120			
Surrogate: 2-Fluorobiphenyl	74.9			ug/l	100		75	45-120			
Surrogate: Terphenyl-d14	88.6			ug/l	100		89	45-120			

LCS Dup Analyzed: 04/13/2005 (5D06042-BSD1)

Naphthalene	111	10	4.5	ug/l	100		111	50-120	4	20	
N-Nitrosodimethylamine	69.1	20	3.7	ug/l	100		69	40-120	19	20	
Surrogate: 2-Fluorophenol	107			ug/l	200		54	30-120			
Surrogate: Phenol-d6	130			ug/l	200		65	35-120			
Surrogate: 2,4,6-Tribromophenol	182			ug/l	200		91	45-120			
Surrogate: Nitrobenzene-d5	71.6			ug/l	100		72	45-120			
Surrogate: 2-Fluorobiphenyl	71.3			ug/l	100		71	45-120			
Surrogate: Terphenyl-d14	95.4			ug/l	100		95	45-120			

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05

Received: 04/05/05

METHOD BLANK/QC DATA

INORGANICS

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC %REC	Limit	RPD	RPD Limit	Data Qualifiers
Batch: 5D06050 Extracted: 04/06/05											
Blank Analyzed: 04/06/2005 (5D06050-BLK1)											
Perchlorate	ND	4.0	0.80	ug/l							
LCS Analyzed: 04/06/2005 (5D06050-BS1)											
Perchlorate	51.5	4.0	0.80	ug/l	50.0		103	85-115			
Matrix Spike Analyzed: 04/06/2005 (5D06050-MS1) Source: IOD0241-01											
Perchlorate	53.1	4.0	0.80	ug/l	50.0	1.7	103	80-120			
Matrix Spike Dup Analyzed: 04/06/2005 (5D06050-MSD1) Source: IOD0241-01											
Perchlorate	53.6	4.0	0.80	ug/l	50.0	1.7	104	80-120	1	20	
Batch: 5D06055 Extracted: 04/06/05											
Blank Analyzed: 04/11/2005 (5D06055-BLK1)											
Biochemical Oxygen Demand	ND	2.0	0.59	mg/l							
LCS Analyzed: 04/11/2005 (5D06055-BS1)											
Biochemical Oxygen Demand	213	100	30	mg/l	198		108	85-115			
LCS Dup Analyzed: 04/11/2005 (5D06055-BSD1)											
Biochemical Oxygen Demand	214	100	30	mg/l	198		108	85-115	1	20	
Batch: 5D07071 Extracted: 04/07/05											
Blank Analyzed: 04/07/2005 (5D07071-BLK1)											
Turbidity	ND	1.0	0.040	NTU							

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing Project ID: Alfa Outfall 012 - During Test
300 North Lake Avenue, Suite 1200 Report Number: IOD0229
Pasadena, CA 91101
Attention: Bronwyn Kelly
Sampled: 04/04/05
Received: 04/05/05

METHOD BLANK/QC DATA

INORGANICS

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5D07071 Extracted: 04/07/05											
Duplicate Analyzed: 04/07/2005 (5D07071-DUP1)						Source: IOD0296-03					
Turbidity	0.120	1.0	0.040	NTU		0.11			9	20	J
Batch: 5D07079 Extracted: 04/07/05											
Blank Analyzed: 04/07/2005 (5D07079-BLK1)											
Total Suspended Solids	ND	10	10	mg/l							
LCS Analyzed: 04/07/2005 (5D07079-BS1)											
Total Suspended Solids	941	10	10	mg/l	1000		94	85-115			
Duplicate Analyzed: 04/07/2005 (5D07079-DUP1)						Source: IOD0149-02					
Total Suspended Solids	ND	10	10	mg/l		ND				10	
Batch: 5D07121 Extracted: 04/07/05											
Blank Analyzed: 04/07/2005 (5D07121-BLK1)											
Total Dissolved Solids	ND	10	10	mg/l							
LCS Analyzed: 04/07/2005 (5D07121-BS1)											
Total Dissolved Solids	960	10	10	mg/l	1000		96	90-110			
Duplicate Analyzed: 04/07/2005 (5D07121-DUP1)						Source: IOD0483-01					
Total Dissolved Solids	186	10	10	mg/l		180			3	10	
Batch: 5D08083 Extracted: 04/08/05											
Blank Analyzed: 04/08/2005 (5D08083-BLK1)											
Ammonia-N (Distilled)	ND	0.50	0.30	mg/l							

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

METHOD BLANK/QC DATA

INORGANICS

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 5D08083 Extracted: 04/08/05											
LCS Analyzed: 04/08/2005 (5D08083-BS1)											
Ammonia-N (Distilled)	9.80	0.50	0.30	mg/l	10.0		98	80-115			
Matrix Spike Analyzed: 04/08/2005 (5D08083-MS1)											
Ammonia-N (Distilled)	9.80	0.50	0.30	mg/l	10.0	0.56	92	70-120			
Matrix Spike Dup Analyzed: 04/08/2005 (5D08083-MSD1)											
Ammonia-N (Distilled)	9.52	0.50	0.30	mg/l	10.0	0.56	90	70-120	3	15	
Batch: 5D11073 Extracted: 04/11/05											
Blank Analyzed: 04/11/2005 (5D11073-BLK1)											
Oil & Grease	ND	5.0	0.94	mg/l							
LCS Analyzed: 04/11/2005 (5D11073-BS1)											
Oil & Grease	15.9	5.0	0.94	mg/l	20.0		80	65-120			M-NR1
LCS Dup Analyzed: 04/11/2005 (5D11073-BSD1)											
Oil & Grease	18.9	5.0	0.94	mg/l	20.0		94	65-120	17	20	

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

METHOD BLANK/QC DATA

1,4-DIOXANE BY GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	MDL	Units	Spike Level	Source Result	%REC %REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: P5D1317 Extracted: 04/13/05											
Blank Analyzed: 04/13/2005 (P5D1317-BLK1)											
I,4-Dioxane	ND	1.0	0.49	ug/l							
Surrogate: Dibromofluoromethane	1.06			ug/l	1.00		106	80-125			
LCS Analyzed: 04/13/2005 (P5D1317-BS1)											
I,4-Dioxane	8.92	1.0	0.49	ug/l	10.0		89	70-130			
Surrogate: Dibromofluoromethane	1.08			ug/l	1.00		108	80-125			
LCS Dup Analyzed: 04/13/2005 (P5D1317-BSD1)											
I,4-Dioxane	8.63	1.0	0.49	ug/l	10.0		86	70-130	3	20	
Surrogate: Dibromofluoromethane	1.07			ug/l	1.00		107	80-125			
Matrix Spike Analyzed: 04/13/2005 (P5D1317-MS1) Source: POD0203-01											
I,4-Dioxane	17.2	1.0	0.49	ug/l	10.0	6.2	110	70-150			
Surrogate: Dibromofluoromethane	1.03			ug/l	1.00		103	80-125			
Matrix Spike Dup Analyzed: 04/13/2005 (P5D1317-MSD1) Source: POD0203-01											
I,4-Dioxane	16.5	1.0	0.49	ug/l	10.0	6.2	103	70-150	4	25	
Surrogate: Dibromofluoromethane	1.02			ug/l	1.00		102	80-125			

Del Mar Analytical, Irvine
Michele Harper
Project Manager



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05

Received: 04/05/05

DATA QUALIFIERS AND DEFINITIONS

- H** Sample analysis performed past method-specified holding time.
- J** Estimated value. Analyte detected at a level less than the Reporting Limit (RL) and greater than or equal to the Method Detection Limit (MDL). The user of this data should be aware that this data is of unknown quality.
- M-NRI** There was no MS/MSD analyzed with this batch due to insufficient sample volume. See Blank Spike/Blank Spike Duplicate.
- ND** Analyte NOT DETECTED at or above the reporting limit or MDL, if MDL is specified.
- RPD** Relative Percent Difference

ADDITIONAL COMMENTS

For GRO (C4-C12):

GRO (C4-C12) is quantitated against a gasoline standard. Quantitation begins immediately following the methanol peak.

For Extractable Fuel Hydrocarbons (EFH, DRO, ORO) :

Unless otherwise noted, Extractable Fuel Hydrocarbons (EFH, DRO, ORO) are quantitated against a Diesel Fuel Standard.



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

Certification Summary

Del Mar Analytical, Irvine

Method	Matrix	Nelac	California
EPA 160.2	Water	X	X
EPA 160.5	Water	X	X
EPA 180.1	Water	X	X
EPA 314.0	Water	N/A	X
EPA 350.2	Water		X
EPA 405.1	Water	X	X
EPA 413.1	Water	X	X
EPA 418.1	Water	X	X
EPA 624	Water	X	X
EPA 625	Water	X	X
EPA 8015 Mod.	Water	X	X
EPA 8015B	Water	X	X
SM2540C	Water	X	X

Nevada and NELAP provide analyte specific accreditations. Analyte specific information for Del Mar Analytical may be obtained by contacting the laboratory or visiting our website at www.dmalabs.com.

Subcontracted Laboratories

Del Mar Analytical - Phoenix NELAC Cert #01109CA, California Cert #2446

9830 S. 51st Street, Suite B-120 - Phoenix, AZ 85044

Method Performed: EPA 8260B

Samples: IOD0229-01

Del Mar Analytical, Irvine
Michele Harper
Project Manager

1779 IODD0229

CHAIN OF CUSTODY FORM

Del Mar Analytical Version 5/8/12/04

Client Name/Address:		Project:		ANALYSIS REQUIRED												Field readings:						
MWH-Pasadena 300 North Lake Avenue, Suite 1200 Pasadena, CA 91101		Boeing-SSFL NPDES During Test -- Outfall 012 Alfa Test Stand		Oil & Grease (EPA 413.1)	8015-gas	8015-diesel/jet fuel	1,4-Dioxane-826B	TRPH, Total Rec. (EPA 418.1)	624 (EDB, 1,2,3-TCF, MTBE, DPE, TBA)	BOD5(20 degrees C)	625 Naphthalene +NDMA analysis	Ammonia-N, Titr. (350.2) w/ dist	Perchlorate	Turbidity, TDS, TSS	Settleable Solids	Temp = 66.2	pH = 7.5					
Sample Description	Sample Matrix	Container Type	# of Cont.	Preservative	Bottle #	Sampling Date/Time	Phone Number: (626) 568-6691	Fax Number: (626) 568-6515	Oil & Grease (EPA 413.1)	8015-gas	8015-diesel/jet fuel	1,4-Dioxane-826B	TRPH, Total Rec. (EPA 418.1)	624 (EDB, 1,2,3-TCF, MTBE, DPE, TBA)	BOD5(20 degrees C)	625 Naphthalene +NDMA analysis	Ammonia-N, Titr. (350.2) w/ dist	Perchlorate	Turbidity, TDS, TSS	Settleable Solids	Comments	
Outfall 012	W	1L Amber	1	HCl	1A	4-4-05 15:30			X													
Outfall 012 duplicate	W	1L Amber	1	HCl	1B				X													
Outfall 012	W	VOAs	1	HCl	2A					X												
Outfall 012 duplicate	W	VOAs	2	HCl	2B, 2C				X													
Outfall 012	W	1L Amber	1	None	3A					X												
Outfall 012 duplicate	W	1L Amber	1	None	3B					X												
Outfall 012	W	VOAs	1	HCl	4A						X											
Outfall 012 duplicate	W	VOAs	2	HCl	4B, 4C						X											
Outfall 012	W	1L Amber	1	HCl	5A							X										
Outfall 012 duplicate	W	1L Amber	1	HCl	5B							X										
Outfall 012	W	VOAs	1	HCl	6A								X									
Outfall 012 duplicate	W	VOAs	2	HCl	6B, 6C							X										
Outfall 012	W	1L Poly	1	None	7A								X									
Outfall 012	W	1L Amber	1	None	8A									X								
Outfall 012 duplicate	W	1L Amber	1	None	8B										X							
Outfall 012	W	500ml Poly	1	H2S04	9A											X						
Outfall 012	W	1L Poly	1	None	10A												X					
Outfall 012	W	1L Poly	1	None	11A													X				
Trip Blank	W	VOAs	6	HCl	12A, 12B, 12C, 12D, 12E, 12F																X	

Turn around Time: (check)
 24 Hours _____ 5 Days _____
 48 Hours _____ 10 Days _____
 72 Hours _____ Normal _____
 Perchlorate Only 72 Hours _____
 Metals Only 72 Hours _____
 Sample Integrity... (Check) On Ice 4°C

Received By: *[Signature]* Date/Time: 4/5/05 10:25
 Received By: *[Signature]* Date/Time: 4/5/05 17:00
 Received By: *[Signature]* Date/Time: 4/5/05 17:00

Relinquished By: *[Signature]* Date/Time: 4/5/05 10:25
 Relinquished By: *[Signature]* Date/Time: 4/5/05 17:00
 Relinquished By: *[Signature]* Date/Time: 4/5/05 17:00



DATA VALIDATION REPORT

NPDES Monitoring

ANALYSIS: SEMIVOLATILES

SAMPLE DELIVERY GROUP: IOD0229

Prepared by

AMEC Denver Operations
550 South Wadsworth Boulevard, Suite 500
Lakewood, Colorado 80226

1. INTRODUCTION

Task Order Title: NPDES Monitoring
Contract Task Order #: 313150010
SDG#: IOD0229
Project Manager: B. McIlvaine
Matrix: Water
Analysis: Semivolatiles
QC Level: Level IV
No. of Samples: 1
No. of Reanalyses/Dilutions: 0
Reviewer: M. Pokorny
Date of Review: May 11, 2005

The samples listed in Table 1 were validated based on the guidelines outlined in the *AMEC Data Validation Procedure for Levels C and D Semivolatile Organics (DVP-3, Rev. 2)*, *EPA Method 625*, and the *National Functional Guidelines For Organic Data Review (2/94)*. Any deviations from these procedures are documented herein. Qualifiers were applied in cases where the data did not meet the required QC criteria or where special consideration by the data user is required. Data qualifiers were placed on Form Is with the associated qualification codes. Analytes that were rejected for any reason are denoted on the Form I as having only the "R" data qualifier and associated qualification code(s) denoting the reason for rejection. Any additional problems with the data that may have resulted in an estimated value were not denoted by a qualification code since the data had already been rejected.

Table 1. Sample identification

Client ID	EPA ID	Lab No.	Matrix	Method
Outfall 012	Outfall 012	IOD0229-01	water	625

2. DATA VALIDATION FINDINGS

2.1 SAMPLE MANAGEMENT

The sample in this SDG was received at the laboratory within the temperature limits of 4°C ±2°C. The analysis did not require preservation, and no preservation was noted in the field. The COC noted that the sample was received intact. No qualifications were required.

2.1.2 Chain of Custody

The COC was signed and dated by both field and laboratory personnel. The COC accounted for the analysis presented in this SDG. As the sample was couriered directly to the laboratory, custody seals were not required. No qualifications were required.

2.1.3 Holding Times

The water sample was extracted within seven days of collection and analyzed within 40 days of collection. No qualifications were required.

2.2 GC/MS TUNING

The DFTPP tunes met the criteria specified in Method 625, and the sample was analyzed within 12 hours of the DFTPP injection time. No qualifications were required.

2.3 CALIBRATION

The initial calibration associated with this SDG was dated 04/12/05. The average RRFs were ≥ 0.05 and the %RSDs were $\leq 35\%$ or $r^2 \geq 0.995$ for both target compounds listed on the sample summary form. A representative number of average RRFs and %RSDs were checked from the raw data, and no calculation or transcription errors were noted. The continuing calibration associated with the sample analysis was analyzed 04/13/05. The RRFs for both target compounds were ≥ 0.05 , and the %Ds were $\leq 20\%$. A representative number of RRFs, r^2 values, and %Ds were checked from the raw data, and no calculation or transcription errors were noted. No qualifications were required.

2.4 BLANKS

One method blank (5D06042-BLK1) was extracted and analyzed with this SDG. No target compounds were reported in the method blank. Review of the raw data indicated no reportable false negatives. No qualifications were required.

2.5 BLANK SPIKES AND LABORATORY CONTROL SAMPLES

One blank spike/blank spike duplicate pair (5D06042-BS1/5D06042-BSD1) was extracted and analyzed with this SDG. All percent recoveries and RPDs were within the laboratory QC limits. A

representative number of recoveries and RPDs were calculated from the raw data and no calculation or transcription errors were found. No qualifications were required.

2.6 SURROGATE RECOVERY

The sample surrogate recoveries were within the laboratory QC limits. A representative number of recoveries were calculated from the raw data, and no transcription or calculation errors were noted. No qualifications were required.

2.7 MATRIX SPIKE/MATRIX SPIKE DUPLICATE

No MS/MSD analyses were associated with this SDG. Evaluation of method accuracy and precision was based on blank spike/blank spike duplicate results. No qualifications were required.

2.8 FIELD QC SAMPLES

Field QC samples were evaluated, and if necessary, qualified based on method blanks and other laboratory QC results affecting the usability of the field QC data. Any remaining detects were used to evaluate the associated site sample. Following are findings associated with field QC samples:

2.8.1 Field Blanks and Equipment Rinsates

There were no field QC samples associated with this SDG. No qualifications were required.

2.8.2 Field Duplicates

There were no field duplicate samples associated with this SDG. No qualifications were required.

2.9 INTERNAL STANDARDS PERFORMANCE

The internal standard area counts and retention times were within the control limits established by the continuing calibration standards: -50%/+100% for internal standard areas and ± 30 seconds for retention times. A representative number of recoveries were checked from the raw data, and no transcription or calculation errors were noted. No qualifications were required.

2.10 COMPOUND IDENTIFICATION

The laboratory analyzed for two semivolatile target compounds by EPA Method 625. Review of the sample chromatogram, retention times, and spectra indicated no problems with target compound identification. No qualifications were required.

2.11 COMPOUND QUANTIFICATION AND REPORTED DETECTION LIMITS

Compound quantification is verified at a Level IV data validation. No calculation or transcription errors were found. The reporting limits were supported by the low level of the initial and the method detection limit study. No qualifications were required.

2.12 TENTATIVELY IDENTIFIED COMPOUNDS

TICs were not reported by the laboratory for this SDG. No qualifications were required.

2.13 SYSTEM PERFORMANCE

Review of the raw data indicated no problems with system performance. No qualifications were required.



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

DRAFT: ACID & BASE/NEUTRALS BY GC/MS (EPA 625)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers	
Sample ID: IOD0229-01 (DRAFT: Outfall 012 - Water)					Sampled: 04/04/05					REV QUAL
Reporting Units: ug/l										CODE
Naphthalene	EPA 625	5D06042	4.5	10	22	0.962	04/06/05	04/14/05		
N-Nitrosodimethylamine	EPA 625	5D06042	3.7	20	ND	0.962	04/06/05	04/14/05	U	
Surrogate: 2-Fluorophenol (30-120%)					60 %					
Surrogate: Phenol-d5 (35-120%)					66 %					
Surrogate: 2,4,6-Tribromophenol (45-120%)					84 %					
Surrogate: Nitrobenzene-d5 (45-120%)					83 %					
Surrogate: 2-Fluorobiphenyl (45-120%)					72 %					
Surrogate: Terphenyl-d14 (45-120%)					92 %					

AMEC VALIDATED

LEVEL IV

DRAFT REPORT
DRAFT REPORT
DATA SUBJECT TO CHANGE

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from Del Mar Analytical.



DATA VALIDATION REPORT

NPDES Monitoring

ANALYSIS: TPH/Purgeable

SAMPLE DELIVERY GROUP: IOD0229

Prepared by

AMEC Denver Operations
550 South Wadsworth Boulevard, Suite 500
Lakewood, Colorado 80226

1. INTRODUCTION

Task Order Title: NPDES Monitoring
Contract Task Order #: 313150010
SDG#: IOD0229
Project Manager: B. McIlvaine
Matrix: Water
Analysis: TPH-Purgeable
QC Level: Level IV
No. of Samples: 2
No. of Reanalyses/Dilutions: 0
Reviewer: L. Calvin
Date of Review: May 13, 2005

The samples listed in Table 1 were validated based on the general guidelines outlined in the *AMEC Data Validation Procedure for Levels C and D Extractable Total Fuel Hydrocarbons by GC (DVP-8, Rev. 2)*, USEPA SW-846 Method 8015M, and validation guidelines outlined in the *USEPA CLP National Functional Guidelines for Organic Data Review (2/94)*. Any deviations from these procedures are documented herein. Qualifiers were applied in cases where the data did not meet the required QC criteria or where special consideration by the data user is required. Data qualifiers were placed on Form Is with the associated qualification codes. Analytes that were rejected for any reason are denoted on the Form I as having only the "R" data qualifier and associated qualification code(s) denoting the reason for rejection. Any additional problems with the data that may have resulted in an estimated value were not denoted by a qualification code since the data had already been rejected.

Table 1. Sample identification

Client ID	EPA ID	Lab No.	Matrix	Method
Outfall 012	Outfall 012	IOD0229-01	water	8015M/GRO
Trip Blank	Trip Blank	IOD0229-02	water	8015M/GRO

2. DATA VALIDATION FINDINGS

2.1 SAMPLE MANAGEMENT

The following are findings associated with sample management:

2.1.1 Sample Preservation, Handling, and Transport

The samples in this SDG were received at Del Mar Analytical on ice within the temperature limits of $4^{\circ}\text{C} \pm 2^{\circ}\text{C}$, at 4°C . The Del Mar Analytical case narrative noted that the samples were received intact, and the COC indicated the samples were properly preserved. No qualifications were required.

2.1.2 Chain of Custody

The COC was signed and dated by both field and laboratory personnel. As the samples were couriered directly to the laboratory, custody seals were not required. No qualifications were required.

2.1.3 Holding Times

The water samples were analyzed within 14 days of collection. No qualifications were required.

2.2 CALIBRATION

Two gasoline standard initial calibrations dated 08/20/04 and 08/26/04 were associated with the sample analyses. The %RSDs for GRO (C4-C12) were within the QC limit of $\leq 20\%$. Initial calibration verifications (ICVs) were not provided in the data package. The %Ds for all CCVs bracketing the sample analyses were within the Method QC limit of $\leq 15\%$. The %RSD and %Ds were recalculated from the raw data and no transcription or calculation errors were noted. No qualifications were required.

2.4 METHOD BLANKS

Two water method blanks (5D12097-BLK1 and 5D13113-BLK1) were associated with the sample analyses. GRO (C4-C12) was not detected above the MDL in either method blank. Review of the raw data indicated no false negative results. No qualifications were necessary.

2.5 BLANK SPIKES AND LABORATORY CONTROL SAMPLES

Two water method blank spikes (5D12097-BS1 and 5D13113-BS1) were associated with the sample analyses. GRO (C4-C12) was recovered within the laboratory-established QC limits of 70-140% in both blank spikes. The recoveries were checked from the raw data, and no calculation or transcription errors were noted. No qualifications were required.

2.6 SURROGATE RECOVERY

The samples were fortified with the surrogate compounds 4-bromofluorobenzene (BFB). Surrogate recoveries were within the laboratory-established QC limits of 65-140%. Recoveries were calculated from the raw data and no transcription or calculation errors were noted. No qualifications were required.

2.7 MATRIX SPIKE/MATRIX SPIKE DUPLICATE

MS/MSD analyses were not performed on the site sample in this SDG; therefore, evaluation of method accuracy was based on the blank spike results. No qualifications were required.

2.8 FIELD QC SAMPLES

Field QC samples are evaluated, and if necessary, qualified based on method blanks and laboratory QC samples for usability. Any remaining detects are used to evaluate the associated samples. The following are findings associated with field QC samples:

2.9.1 Trip Blanks, Field Blanks, and Equipment Rinsates

Sample Trip Blank was the trip blank associated with site sample Outfall 012. GRO (C4-C12) was not detected above the MDL in the trip blank. Review of the raw data indicated no false negative result. There were no field blank or equipment rinsate samples associated with this SDG. No qualifications were necessary.

2.9.2 Field Duplicates

There were no field duplicate samples in this SDG.

2.10 COMPOUND IDENTIFICATION

The laboratory analyzed for GRO (C4-C12) by Method 8015M. Compound identification is verified at a Level IV validation. Review of chromatograms and retention times indicated no problems with compound identification for the samples in this SDG. No qualifications were required.

2.11 COMPOUND QUANTIFICATION AND REPORTED DETECTION LIMITS

Compound quantification was verified for this SDG by recalculating any sample detects, blank spike recoveries, and a representative number of surrogate recoveries. Reporting limits were supported by the low level standard of the initial calibration and by the laboratory MDL. The results were reported in mg/L (ppm). No qualifications were required.



MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

DRAFT: VOLATILE FUEL HYDROCARBONS (EPA 5030/CADHS Mod. 8015)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD0229-01 (DRAFT: Outfall 012 - Water) - cont.					Sampled: 04/04/05				
Reporting Units: mg/l									
GRO (C4 - C12)	EPA 8015 Mod.	5D13113	0.25	0.50	1.2	5	04/13/05	04/13/05	rev qual qual code
Surrogate: 4-BFB (FID) (65-140%)					92 %				
Sample ID: IOD0229-02 (DRAFT: Trip Blank - Water)					Sampled: 04/04/05				
Reporting Units: mg/l									
GRO (C4 - C12)	EPA 8015 Mod.	5D12097	0.050	0.10	ND	1	04/12/05	04/12/05	u
Surrogate: 4-BFB (FID) (65-140%)					92 %				

**AMEC VALIDATED
LEVEL IV**

DRAFT REPORT
DRAFT REPORT
DATA SUBJECT TO CHANGE

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from Del Mar Analytical.



DATA VALIDATION REPORT

NPDES Monitoring

ANALYSIS: TPH/Extractable

SAMPLE DELIVERY GROUP: IOD0229

Prepared by

AMEC Denver Operations
550 South Wadsworth Boulevard, Suite 500
Lakewood, Colorado 80226

1. INTRODUCTION

Task Order Title: NPDES Monitoring
Contract Task Order #: 313150010
SDG#: IOD0229
Project Manager: B. McIlvaine
Matrix: Water
Analysis: TPH-Extractable
QC Level: Level IV
No. of Samples: 1
No. of Reanalyses/Dilutions: 0
Reviewer: L. Calvin
Date of Review: May 13, 2005

The samples listed in Table 1 were validated based on the general guidelines outlined in the *AMEC Data Validation Procedure for Levels C and D Extractable Total Fuel Hydrocarbons by GC (DVP-8, Rev. 2)*, USEPA SW-846 Method 8015B, and validation guidelines outlined in the USEPA *CLP National Functional Guidelines for Organic Data Review (2/94)*. Any deviations from these procedures are documented herein. Qualifiers were applied in cases where the data did not meet the required QC criteria or where special consideration by the data user is required. Data qualifiers were placed on Form Is with the associated qualification codes. Analytes that were rejected for any reason are denoted on the Form I as having only the "R" data qualifier and associated qualification code(s) denoting the reason for rejection. Any additional problems with the data that may have resulted in an estimated value were not denoted by a qualification code since the data had already been rejected.

Table 1. Sample identification

Client ID	EPA ID	Lab No.	Matrix	Method
Outfall 012	Outfall 012	IOD0229-01	water	8015B

2. DATA VALIDATION FINDINGS

2.1 SAMPLE MANAGEMENT

The following are findings associated with sample management:

2.1.1 Sample Preservation, Handling, and Transport

The sample in this SDG was received at Del Mar Analytical laboratory on ice within the temperature limits of $4^{\circ}\text{C} \pm 2^{\circ}\text{C}$. The Del Mar Analytical case narrative noted that the sample containers were received intact. No qualifications were required.

2.1.2 Chain of Custody

The COC was signed and dated by both field and laboratory personnel, and accounted for the analysis presented in this SDG. The sample was analyzed correctly. As the site sample was couriered directly to the laboratory, custody seals were not required. No qualifications were required.

2.1.3 Holding Times

The sample was extracted within seven days of sample collection and analyzed within 40 days of extraction. No qualifications were required.

2.2 CALIBRATION

The initial calibration associated with the sample analysis was analyzed on 04/05/05. The %RSD was within the QC limit of $\leq 20\%$. The %Ds for the initial calibration verification (ICV) and continuing calibrations associated with the sample analysis were $\leq 15\%$. The %RSD and %Ds were recalculated from the raw data and no transcription or calculation errors were noted. No qualifications were required.

2.4 METHOD BLANKS

One method blank (5D08084-BLK1) was extracted and analyzed with the sample in this SDG. EFH (C13-C22) was not present above the MDL in the method blank or in the instrument blank analyzed at the beginning of the analytical sequence. Review of the chromatograms showed no false negatives. No qualifications were required.

2.5 BLANK SPIKES AND LABORATORY CONTROL SAMPLES

One method blank spike/blank spike duplicate pair (5D08084-BS1/BSD1) was extracted and analyzed with the sample in this SDG. The laboratory reported recoveries of alkane range C13-C28 from spiked diesel. The recoveries were within the laboratory-established QC limits of 40-120%, and the RPD was within the QC limit of $\leq 25\%$. The recoveries and RPD were checked from the raw data, and no calculation or transcription errors were noted. No qualifications were required.

2.6 SURROGATE RECOVERY

The sample was fortified with the surrogate compound n-octacosane. The sample surrogate recovery was within the laboratory-established QC limits of 40-125%. The recovery was calculated from the raw data and no transcription or calculation errors were noted. No qualifications were required.

2.7 MATRIX SPIKE/MATRIX SPIKE DUPLICATE

There were no MS/MSD analyses associated with the sample of this SDG. Evaluation of method accuracy and precision was based on the BS/BSD results. No qualifications were required.

2.8 FIELD QC SAMPLES

Field QC samples are evaluated, and if necessary, qualified based on method blanks and laboratory QC samples for usability. Any remaining detects are used to evaluate the associated sample. The following are findings associated with field QC samples:

2.9.1 Field Blanks and Equipment Rinsates

There were no field blank or equipment rinsate samples associated with the site sample in this SDG. No qualifications were required.

2.9.2 Field Duplicates

There were no field duplicate samples associated with this SDG.

2.10 COMPOUND IDENTIFICATION

The laboratory analyzed for EFH n-alkane range C13-C22 by EPA SW-846 Method 8015B. Compound identification is verified at a Level IV validation. Review of chromatograms and retention times indicated no problems with compound identification for this SDG. No qualifications were required.

2.11 COMPOUND QUANTIFICATION AND REPORTED DETECTION LIMITS

Compound quantification was verified for this SDG by recalculating any sample detect, blank spike recoveries, and a representative number of surrogate recoveries. Reporting limits were supported by the low level standard of the initial calibration and by the laboratory MDL. Results were reported in mg/L (ppm). No qualifications were required.



Del Mar Analytical

17461 Deran Ave., Suite 100, Irvine, CA 92614 (949) 261-1022 FAX (949) 260-3297
1014 E. Cooley Dr., Suite A, Colton, CA 92324 (909) 370-4667 FAX (909) 370-1046
9484 Chesapeake Dr., Suite 805, San Diego, CA 92123 (858) 505-8596 FAX (858) 505-9689
9530 South 51st St., Suite B-120, Phoenix, AZ 85044 (480) 785-0043 FAX (480) 785-0851
2520 E. Sunset Rd. #3, Las Vegas, NV 89120 (702) 798-3620 FAX (702) 798-3621

MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

DRAFT: EXTRACTABLE FUEL HYDROCARBONS (CADHS/8015 Modified)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD0229-01 (DRAFT: Outfall 012 - Water) - cont.									
Reporting Units: mg/l									
EFH (C13 - C22)	EPA 8015B	5D08084	0.082	0.50	0.81	0.962	04/08/05	04/09/05	view qual qual code
Surrogate: n-Octacosane (40-125%)					80 %				

**AMEC VALIDATED
LEVEL IV**

DRAFT REPORT
DRAFT REPORT
DATA SUBJECT TO CHANGE

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from Del Mar Analytical.

CONTRACT COMPLIANCE SCREENING FORM FOR HARDCOPY DATA

AMEC Earth & Environmental
 550 South Wadsworth Boulevard
 Suite 500
 Lakewood, CO 80226

Package ID T711TF65
 Task Order 313150010
 SDG No. IOD0229

No. of Analyses 1

Laboratory Del Mar

Reviewer L. Calvin

Analysis/Method EPH by Method 8015B

Date: May 13, 2005
 Reviewer's Signature: *L. Calvin*

ACTION ITEMS^a	
1. Case Narrative Deficiencies	
2. Out of Scope Analyses	
3. Analyses Not Conducted	
4. Missing Hardcopy Deliverables	
5. Incorrect Hardcopy Deliverables	
6. Deviations from Analysis Protocol, e.g., Holding Times GC/MS Tune/Inst. Performance Calibration Method blanks Surrogates Matrix Spike/Dup LCS Field QC Internal Standard Performance Compound Identification Quantitation System Performance	
COMMENTS^b	Acceptable as reviewed.
^a Subcontracted analytical laboratory is not meeting contract and/or method requirements. ^b Differences in protocol have been adopted by the laboratory but no action against the laboratory is required.	



DATA VALIDATION REPORT

NPDES Monitoring

ANALYSIS: TPH/Extractable

SAMPLE DELIVERY GROUP: IOD0229

Prepared by

AMEC Denver Operations
550 South Wadsworth Boulevard, Suite 500
Lakewood, Colorado 80226

1. INTRODUCTION

Task Order Title: NPDES Monitoring
Contract Task Order #: 313150010
SDG#: IOD0229
Project Manager: B. McIlvaine
Matrix: Water
Analysis: TPH-Extractable
QC Level: Level IV
No. of Samples: 1
No. of Reanalyses/Dilutions: 0
Reviewer: L. Calvin
Date of Review: May 13, 2005

The samples listed in Table 1 were validated based on the general guidelines outlined in the *AMEC Data Validation Procedure for Levels C and D Extractable Total Fuel Hydrocarbons by GC (DVP-8, Rev. 2)*, USEPA SW-846 Method 8015B, and validation guidelines outlined in the *USEPA CLP National Functional Guidelines for Organic Data Review (2/94)*. Any deviations from these procedures are documented herein. Qualifiers were applied in cases where the data did not meet the required QC criteria or where special consideration by the data user is required. Data qualifiers were placed on Form Is with the associated qualification codes. Analytes that were rejected for any reason are denoted on the Form I as having only the "R" data qualifier and associated qualification code(s) denoting the reason for rejection. Any additional problems with the data that may have resulted in an estimated value were not denoted by a qualification code since the data had already been rejected.

Table 1. Sample identification

Client ID	EPA ID	Lab No.	Matrix	Method
Outfall 012	Outfall 012	IOD0229-01	water	8015B

2. DATA VALIDATION FINDINGS

2.1 SAMPLE MANAGEMENT

The following are findings associated with sample management:

2.1.1 Sample Preservation, Handling, and Transport

The sample in this SDG was received at Del Mar Analytical laboratory on ice within the temperature limits of $4^{\circ}\text{C} \pm 2^{\circ}\text{C}$. The Del Mar Analytical case narrative noted that the sample containers were received intact. No qualifications were required.

2.1.2 Chain of Custody

The COC was signed and dated by both field and laboratory personnel, and accounted for the analysis presented in this SDG. The sample was analyzed correctly. As the site sample was couriered directly to the laboratory, custody seals were not required. No qualifications were required.

2.1.3 Holding Times

The sample was extracted within seven days of sample collection and analyzed within 40 days of extraction. No qualifications were required.

2.2 CALIBRATION

The initial calibration associated with the sample analysis was analyzed on 04/05/05. The %RSD was within the QC limit of $\leq 20\%$. The %Ds for the initial calibration verification (ICV) and continuing calibrations associated with the sample analysis were $\leq 15\%$. The %RSD and %Ds were recalculated from the raw data and no transcription or calculation errors were noted. No qualifications were required.

2.4 METHOD BLANKS

One method blank (5D08084-BLK1) was extracted and analyzed with the sample in this SDG. EFH (C13-C22) was not present above the MDL in the method blank or in the instrument blank analyzed at the beginning of the analytical sequence. Review of the chromatograms showed no false negatives. No qualifications were required.

2.5 BLANK SPIKES AND LABORATORY CONTROL SAMPLES

One method blank spike/blank spike duplicate pair (5D08084-BS1/BSD1) was extracted and analyzed with the sample in this SDG. The laboratory reported recoveries of alkane range C13-C28 from spiked diesel. The recoveries were within the laboratory-established QC limits of 40-120%, and the RPD was within the QC limit of $\leq 25\%$. The recoveries and RPD were checked from the raw data, and no calculation or transcription errors were noted. No qualifications were required.

2.6 SURROGATE RECOVERY

The sample was fortified with the surrogate compound n-octacosane. The sample surrogate recovery was within the laboratory-established QC limits of 40-125%. The recovery was calculated from the raw data and no transcription or calculation errors were noted. No qualifications were required.

2.7 MATRIX SPIKE/MATRIX SPIKE DUPLICATE

There were no MS/MSD analyses associated with the sample of this SDG. Evaluation of method accuracy and precision was based on the BS/BSD results. No qualifications were required.

2.8 FIELD QC SAMPLES

Field QC samples are evaluated, and if necessary, qualified based on method blanks and laboratory QC samples for usability. Any remaining detects are used to evaluate the associated sample. The following are findings associated with field QC samples:

2.9.1 Field Blanks and Equipment Rinsates

There were no field blank or equipment rinsate samples associated with the site sample in this SDG. No qualifications were required.

2.9.2 Field Duplicates

There were no field duplicate samples associated with this SDG.

2.10 COMPOUND IDENTIFICATION

The laboratory analyzed for EFH n-alkane range C13-C22 by EPA SW-846 Method 8015B. Compound identification is verified at a Level IV validation. Review of chromatograms and retention times indicated no problems with compound identification for this SDG. No qualifications were required.

2.11 COMPOUND QUANTIFICATION AND REPORTED DETECTION LIMITS

Compound quantification was verified for this SDG by recalculating any sample detect, blank spike recoveries, and a representative number of surrogate recoveries. Reporting limits were supported by the low level standard of the initial calibration and by the laboratory MDL. Results were reported in mg/L (ppm). No qualifications were required.



Del Mar Analytical

17461 Denan Ave., Suite 100, Irvine, CA 92614 (949) 261-1022 FAX (949) 260-3297
1014 E. Cooley Dr., Suite A, Colton, CA 92324 (909) 370-4667 FAX (949) 370-1046
9484 Chesapeake Dr., Suite 805, San Diego, CA 92123 (858) 505-8396 FAX (858) 505-9689
9330 South 51st St., Suite B-120, Phoenix, AZ 85044 (480) 785-0043 FAX (480) 785-0651
2520 E. Sunset Rd. #3, Las Vegas, NV 89120 (702) 798-3620 FAX (702) 798-3621

MWH-Pasadena/Boeing
300 North Lake Avenue, Suite 1200
Pasadena, CA 91101
Attention: Bronwyn Kelly

Project ID: Alfa Outfall 012 - During Test

Report Number: IOD0229

Sampled: 04/04/05
Received: 04/05/05

DRAFT: EXTRACTABLE FUEL HYDROCARBONS (CADHS/8015 Modified)

Analyte	Method	Batch	MDL Limit	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IOD0229-01 (DRAFT: Outfall 012 - Water) - cont.					Sampled: 04/04/05				
Reporting Units: mg/l									
EFH (C13 - C22)	EPA 8015B	5D08084	0.082	0.50	0.81	0.962	04/08/05	04/09/05	view qual qual code
Surrogate: n-Octacosane (40-125%)					80 %				

**AMEC VALIDATED
LEVEL IV**

DRAFT REPORT
DRAFT REPORT
DATA SUBJECT TO CHANGE


The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from Del Mar Analytical.

CONTRACT COMPLIANCE SCREENING FORM FOR HARDCOPY DATA

AMEC Earth & Environmental
 550 South Wadsworth Boulevard
 Suite 500
 Lakewood, CO 80226

Package ID T711VO97
 Task Order 313150010
 SDG No. IOD0229
 No. of Analyses 2

Laboratory Del Mar Analytical
 Reviewer K. Shadowlight
 Analysis/Method Volatiles by 624

Date May 6, 2005
 Reviewer's Signature


ACTION ITEMS*	
1. Case Narrative	
Deficiencies	
2. Out of Scope	
Analyses	
3. Analyses Not Conducted	
4. Missing Hardcopy	
Deliverables	
5. Incorrect Hardcopy	
Deliverables	
6. Deviations from Analysis	Qualifications were assigned for %D outlier in the continuing calibration
GC/MS Tune/Inst. Perform	
Calibrations	
Blanks	
Surrogates	
Matrix Spike/Dup LCS	
Field QC	
Internal Standard Performance	
Compound Identification and	
Quantitation	
System Performance	
COMMENTS^b	
^a Subcontracted analytical laboratory is not meeting contract and/or method requirements. ^b Differences in protocol have been adopted by the laboratory but no action against the laboratory is required.	

Data Qualifier Reference Table

Qualifier	Organics	Inorganics
U	The analyte was analyzed for, but was not detected above the reported sample quantitation limit.	The material was analyzed for, but was not detected above the level of the associated value. The associated value is either the sample quantitation limit or the sample detection limit.
J	The analyte was positively identified; the associated numerical value is the approximate concentration of the analyte in the sample.	The associated value is an estimated quantity.
N	The analysis indicates the presence of an analyte for which there is presumptive evidence to make a "tentative identification."	Not applicable.
NJ	The analysis indicates the presence of an analyte that has been "tentatively identified" and the associated numerical value represents its approximate concentration.	Not applicable.
UU	The analyte was not deemed above the reported sample quantitation limit. However, the reported quantitation limit is approximate and may or may not represent the actual limit of quantitation necessary to accurately and precisely measure the analyte in the sample.	The material was analyzed for, but was not detected. The associated value is an estimate and may be inaccurate or imprecise.
R	The sample results are rejected due to serious deficiencies in the ability to analyze the sample and to meet quality control criteria. The presence or absence of the analyte cannot be verified.	The data are unusable. (Note: Analyte may or may not be present).

Qualification Code Reference Table

Qualifier	Organics	Inorganics
H	Holding times were exceeded.	Holding times were exceeded.
S	Surrogate recovery was outside QC limits.	The sequence or number of standards used for the calibration was incorrect
C	Calibration %RSD or %D were noncompliant.	Correlation coefficient is <0.995.
R	Calibration RRF was <0.05.	%R for calibration is not within control limits.
B	Presumed contamination from preparation (method) blank.	Presumed contamination from preparation (method) or calibration blank.
L	Laboratory Blank Spike/Blank Spike Duplicate %R was not within control limits.	Laboratory Control Sample %R was not within control limits.
Q	MS/MSD recovery was poor or RPD high.	MS recovery was poor.
E	Not applicable.	Duplicates showed poor agreement.
I	Internal standard performance was unsatisfactory.	ICP ICS results were unsatisfactory.
A	Not applicable.	ICP Serial Dilution %D were not within control limits.
M	Tuning (BFB or DFTPP) was noncompliant.	Not applicable.
T	Presumed contamination from trip blank.	Not applicable.
+	False positive – reported compound was not present. Not applicable.	
-	False negative – compound was present but not reported.	Not applicable.
F	Presumed contamination from FB, or ER.	Presumed contamination from FB or ER.
\$	Reported result or other information was incorrect.	Reported result or other information was incorrect.
?	TIC identity or reported retention time has been changed.	Not applicable.
D	The analysis with this flag should not be used because another more technically sound analysis is available.	The analysis with this flag should not be used because another more technically sound analysis is available.
P	Instrument performance for pesticides was poor.	Post Digestion Spike recovery was not within control limits.
DNQ	The compound was detected between the MDL and the RL and, by definition, is considered an estimated value.	The compound was detected between the MDL and the RL and, by definition, is considered an estimated value.
**	Unusual problems found with the data that have been described in Section 2.#, "Data Validation Findings." The number following the asterisk (*) will indicate the subsection where a description of the problem can be found (eg. *1 would indicate a sample was not within temperature limits).	Unusual problems found with the data that have been described in Section 2.#, "Data Validation Findings." The number following the asterisk (*) will indicate the subsection where a description of the problem can be found (eg. *1 would indicate a sample was not within temperature limits).



DATA VALIDATION REPORT

NPDES Monitoring

ANALYSIS: VOLATILES

SAMPLE DELIVERY GROUP: IOD0229

Prepared by

AMEC Denver Operations
550 South Wadsworth Boulevard, Suite 500
Lakewood, Colorado 80226